

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3085

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3730	4000	93	90 - 110	P	09/15/2025	11:31	LB137198

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3085

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.7	12.0	98	80 - 120	P	09/15/2025	11:36	LB137198

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3085

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4940	5000	99	90 - 110	P	09/15/2025	12:35	LB137198
CCV02	Lead	4710	5000	94	90 - 110	P	09/15/2025	13:42	LB137198
CCV03	Lead	4810	5000	96	90 - 110	P	09/15/2025	15:11	LB137198
CCV04	Lead	4840	5000	97	90 - 110	P	09/15/2025	16:12	LB137198
CCV05	Lead	4760	5000	95	90 - 110	P	09/15/2025	16:55	LB137198