

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3086

MATRIX: Water

METHOD: SM2540 C, SM2540 D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. The Blank Spike met requirements for all compounds.			✓
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS: The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

QA REVIEW

Date