

LAB CHRONICLE

OrderID:	Q3086	OrderDate:	9/12/2025 10:59:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3086-01	RW8-SP100-2025091 1	Water	SVOC-SIMGroup1	8270-Modified	09/11/25	09/15/25	09/17/25	09/12/25
Q3086-02	RW8-SP303-2025091 1	Water	SVOC-SIMGroup1	8270-Modified	09/11/25	09/15/25	09/17/25	09/12/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q3086
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW8-SP100-20250911							
Q3086-01	RW8-SP100-20250911	WATER	1,4-Dioxane	0.070	J	0.07	0.2	0.2 ug/L
Total Svoc :				0.07				
Total Concentration:				0.07				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP100-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037745.D	1	09/15/25 10:15	09/17/25 14:35	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.070	J	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		89%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		101%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.55	*	58 - 132		138%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6570	7.84				
1146-65-2	Naphthalene-d8	18000	10.637				
15067-26-2	Acenaphthene-d10	8330	14.495				
1517-22-2	Phenanthrene-d10	15400	17.235				
1719-03-5	Chrysene-d12	14200	21.411				
1520-96-3	Perylene-d12	14200	23.75				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP303-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037746.D	1	09/15/25 10:15	09/17/25 15:11	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		91%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.41		53 - 106		103%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		130%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6530	7.84				
1146-65-2	Naphthalene-d8	17800	10.637				
15067-26-2	Acenaphthene-d10	8070	14.495				
1517-22-2	Phenanthrene-d10	13900	17.235				
1719-03-5	Chrysene-d12	10600	21.411				
1520-96-3	Perylene-d12	10300	23.75				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: **Q3086**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8270-Modified**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169692BL	PB169692BL	2-Methylnaphthalene-d10	0.4	0.31	79		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.34	84		58	132
PB169692BS	PB169692BS	2-Methylnaphthalene-d10	0.4	0.30	76		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.31	78		58	132
Q3086-01	RW8-SP100-20250911	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.40	101		53	106
		Terphenyl-d14	0.4	0.55	138	*	58	132
Q3086-02	RW8-SP303-20250911	2-Methylnaphthalene-d10	0.4	0.36	91		30	150
		Fluoranthene-d10	0.4	0.35	86		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.41	103		53	106
		Terphenyl-d14	0.4	0.52	130		58	132
Q3095-06MS	MW203D2-20250909MS	2-Methylnaphthalene-d10	0.4	0.29	72		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.30	75		55	111
		2-Fluorobiphenyl	0.4	0.32	79		53	106
		Terphenyl-d14	0.4	0.40	99		58	132
Q3095-07MSD	MW203D2-20250909MSD	2-Methylnaphthalene-d10	0.4	0.30	74		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.30	75		55	111
		2-Fluorobiphenyl	0.4	0.32	79		53	106
		Terphenyl-d14	0.4	0.40	99		58	132



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.:

Q3086

Analytical Method:

SW8270-Modified

Client:

Tetra Tech NUS, Inc.

DataFile:

BN037743.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3095-06MS	Client Sample ID:	MW203D2-20250909MS								
1,4-Dioxane	0.4	0.85	0.97	ug/L	30	*			70	130	

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.:	Q3086	Analytical Method:	SW8270-Modified
Client:	Tetra Tech NUS, Inc.	DataFile:	BN037744.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3095-07MSD	Client Sample ID:	MW203D2-20250909MSD								
1,4-Dioxane	0.4	0.85	0.96	ug/L	27	*	11		70	130	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3086

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037791.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169692BS	1,4-Dioxane	0.4	0.30	ug/L	75				70	130	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169692BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3086

Lab File ID: BN037796.D

Lab Sample ID: PB169692BL

Instrument ID: BNA_N

Date Extracted: 09/15/2025

Matrix: (soil/water) Water

Date Analyzed: 09/19/2025

Level: (low/med) LOW

Time Analyzed: 03:24

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
RW8-SP100-20250911	Q3086-01	BN037745.D	09/17/2025
RW8-SP303-20250911	Q3086-02	BN037746.D	09/17/2025
MW203D2-20250909MS	Q3095-06MS	BN037743.D	09/17/2025
MW203D2-20250909MSD	Q3095-07MSD	BN037744.D	09/17/2025
PB169692BS	PB169692BS	BN037791.D	09/18/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037689.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3086
DFTPP Injection Date: 09/11/2025
DFTPP Injection Time: 07:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	79.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.5 (20.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037691.D	09/11/2025	09:56
SSTDICC0.2	SSTDICC0.2	BN037692.D	09/11/2025	10:32
SSTDICCC0.4	SSTDICCC0.4	BN037693.D	09/11/2025	11:08
SSTDICC0.8	SSTDICC0.8	BN037694.D	09/11/2025	11:44
SSTDICC1.6	SSTDICC1.6	BN037695.D	09/11/2025	12:20
SSTDICC3.2	SSTDICC3.2	BN037696.D	09/11/2025	12:56
SSTDICC5.0	SSTDICC5.0	BN037697.D	09/11/2025	13:33

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037739.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3086
DFTPP Injection Date: 09/17/2025
DFTPP Injection Time: 10:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.3) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	82.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (21.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037740.D	09/17/2025	11:25
MW203D2-20250909MS	Q3095-06MS	BN037743.D	09/17/2025	13:14
MW203D2-20250909MSD	Q3095-07MSD	BN037744.D	09/17/2025	13:59
RW8-SP100-20250911	Q3086-01	BN037745.D	09/17/2025	14:35
RW8-SP303-20250911	Q3086-02	BN037746.D	09/17/2025	15:11
SSTDCCC0.4EC	SSTDCCC0.4	BN037757.D	09/17/2025	21:50

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037775.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3086
DFTPP Injection Date: 09/18/2025
DFTPP Injection Time: 13:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	85.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.2 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037776.D	09/18/2025	14:32
PB169692BS	PB169692BS	BN037791.D	09/18/2025	23:40
SSTDCCC0.4EC	SSTDCCC0.4	BN037793.D	09/19/2025	00:52

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037794.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3086
DFTPP Injection Date: 09/19/2025
DFTPP Injection Time: 02:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.4) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	86.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.9 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037795.D	09/19/2025	02:47
PB169692BL	PB169692BL	BN037796.D	09/19/2025	03:24
SSTDCCC0.4EC	SSTDCCC0.4	BN037811.D	09/19/2025	13:03

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/17/2025

Lab File ID: BN037740.D

Time Analyzed: 11:25

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6660	7.84	18396	10.65	8930	14.50
UPPER LIMIT	13320	8.34	36792	11.148	17860	14.995
LOWER LIMIT	3330	7.34	9198	10.148	4465	13.995
EPA SAMPLE NO.						
01 RW8-SP100-20250911	6573	7.84	18040	10.64	8326	14.50
02 RW8-SP303-20250911	6528	7.84	17791	10.64	8068	14.50
03 MW203D2-20250909MS	6853	7.84	18950	10.64	8815	14.49
04 MW203D2-20250909MSD	7569	7.85	20880	10.65	9823	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID: SSTDCCC0.4

Date Analyzed: 09/17/2025

Lab File ID: BN037740.D

Time Analyzed: 11:25

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	15814	17.235	13841	21.42	14132	23.753
UPPER LIMIT	31628	17.735	27682	21.92	28264	24.253
LOWER LIMIT	7907	16.735	6920.5	20.92	7066	23.253
EPA SAMPLE NO.						
01 RW8-SP100-20250911	15420	17.24	14189	21.41	14177	23.75
02 RW8-SP303-20250911	13903	17.24	10625	21.41	10300	23.75
03 MW203D2-20250909MS	15289	17.24	13059	21.42	12757	23.75
04 MW203D2-20250909MSD	16954	17.24	14253	21.42	13967	23.76

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/18/2025

Lab File ID: BN037776.D

Time Analyzed: 14:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7127	7.833	18984	10.64	8698	14.48
UPPER LIMIT	14254	8.333	37968	11.137	17396	14.984
LOWER LIMIT	3563.5	7.333	9492	10.137	4349	13.984
EPA SAMPLE NO.						
01 PB169692BS	6983	7.83	18103	10.64	7729	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID: SSTDCCC0.4

Date Analyzed: 09/18/2025

Lab File ID: BN037776.D

Time Analyzed: 14:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14526	17.235	10438	21.411	10600	23.744
UPPER LIMIT	29052	17.735	20876	21.911	21200	24.244
LOWER LIMIT	7263	16.735	5219	20.911	5300	23.244
EPA SAMPLE NO.						
PB169692BS	12277	17.24	10644	21.41	10808	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037795.D

Time Analyzed: 02:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7724	7.833	19361	10.64	8621	14.48
UPPER LIMIT	15448	8.333	38722	11.137	17242	14.984
LOWER LIMIT	3862	7.333	9680.5	10.137	4310.5	13.984
EPA SAMPLE NO.						
01 PB169692BL	7701	7.83	18665	10.64	7806	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID: SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037795.D

Time Analyzed: 02:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14380	17.235	11169	21.411	11840	23.742
UPPER LIMIT	28760	17.735	22338	21.911	23680	24.242
LOWER LIMIT	7190	16.735	5584.5	20.911	5920	23.242
EPA SAMPLE NO.						
01 PB169692BL	12511	17.24	10022	21.41	10278	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB169692BL		SDG No.:	Q3086	
Lab Sample ID:	PB169692BL		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037796.D	1	09/15/25 10:15	09/19/25 03:24	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		75%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		92%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.34		58 - 132		84%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	7700	7.833				
1146-65-2	Naphthalene-d8	18700	10.637				
15067-26-2	Acenaphthene-d10	7810	14.484				
1517-22-2	Phenanthrene-d10	12500	17.236				
1719-03-5	Chrysene-d12	10000	21.411				
1520-96-3	Perylene-d12	10300	23.739				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169692BS		SDG No.:	Q3086
Lab Sample ID:	PB169692BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037791.D	1	09/15/25 10:15	09/18/25 23:40	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.30		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		75%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		88%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.31		58 - 132		78%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6980	7.833				
1146-65-2	Naphthalene-d8	18100	10.637				
15067-26-2	Acenaphthene-d10	7730	14.484				
1517-22-2	Phenanthrene-d10	12300	17.235				
1719-03-5	Chrysene-d12	10600	21.411				
1520-96-3	Perylene-d12	10800	23.739				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/09/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	MW203D2-20250909MS	SDG No.:	Q3086
Lab Sample ID:	Q3095-06MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037743.D	1	09/15/25 10:15	09/17/25 13:14	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.97		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.29		30 - 150		72%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		75%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.32		53 - 106		79%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		99%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6850	7.84				
1146-65-2	Naphthalene-d8	19000	10.637				
15067-26-2	Acenaphthene-d10	8820	14.494				
1517-22-2	Phenanthrene-d10	15300	17.235				
1719-03-5	Chrysene-d12	13100	21.42				
1520-96-3	Perylene-d12	12800	23.75				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/09/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	MW203D2-20250909MSD	SDG No.:	Q3086
Lab Sample ID:	Q3095-07MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037744.D	1	09/15/25 10:15	09/17/25 13:59	PB169692

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.96		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		74%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		75%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.32		53 - 106		79%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		99%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	7570	7.847				
1146-65-2	Naphthalene-d8	20900	10.648				
15067-26-2	Acenaphthene-d10	9820	14.494				
1517-22-2	Phenanthrene-d10	17000	17.235				
1719-03-5	Chrysene-d12	14300	21.42				
1520-96-3	Perylene-d12	14000	23.756				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN091125.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 11 14:20:01 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037691.D 0.2 =BN037692.D 0.4 =BN037693.D 0.8 =BN037694.D 1.6 =BN037695.D 3.2 =BN037696.D 5 =BN037697.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.429	0.422	0.411	0.432	0.418	0.388	0.417	3.80	
3)	n-Nitrosodimet...	0.530	0.564	0.515	0.556	0.550	0.551	0.544	3.34	
4) S	2-Fluorophenol	0.929	0.913	0.892	0.840	0.912	0.937	1.004	0.918	5.37
5) S	Phenol-d6	1.191	1.193	1.156	1.088	1.193	1.225	1.289	1.191	5.15
6)	bis(2-Chloroet...	1.076	1.075	1.070	1.019	1.078	1.066	1.063	1.064	1.94
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.287	0.279	0.285	0.272	0.296	0.303	0.325	0.293	6.05
9)	Naphthalene	1.091	1.072	1.074	1.036	1.117	1.115	1.149	1.093	3.38
10)	Hexachlorobuta...	0.199	0.196	0.197	0.189	0.203	0.199	0.202	0.198	2.29
11) SURR	2-Methylnaphth...	0.497	0.497	0.498	0.478	0.522	0.550	0.668	0.530	12.31
12)	2-Methylnaphth...	0.643	0.651	0.659	0.633	0.697	0.712	0.740	0.676	5.91
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.088	0.098	0.101	0.096	0.113	0.131	0.104	14.74	
15) S	2-Fluorobiphenyl	1.663	1.666	1.649	1.586	1.689	1.674	1.785	1.673	3.54
16)	Acenaphthylene	1.754	1.756	1.783	1.727	1.885	1.916	2.020	1.834	5.90
17)	Acenaphthene	1.165	1.199	1.215	1.187	1.272	1.296	1.348	1.240	5.37
18)	Fluorene	1.404	1.402	1.441	1.419	1.545	1.633	1.671	1.502	7.59
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.032	0.033	0.033	0.040	0.046	0.037	16.34		
21)	4-Bromophenyl-...	0.250	0.252	0.251	0.248	0.272	0.269	0.284	0.261	5.43
22)	Hexachlorobenzene	0.301	0.308	0.294	0.280	0.305	0.302	0.310	0.300	3.39
23)	Atrazine	0.145	0.151	0.152	0.146	0.169	0.190	0.209	0.166	15.00
24)	Pentachlorophenol	0.074	0.074	0.073	0.086	0.100	0.115	0.087	19.46	
25)	Phenanthrene	1.228	1.215	1.223	1.173	1.285	1.321	1.371	1.259	5.47
26)	Anthracene	1.038	1.029	1.070	1.030	1.140	1.221	1.281	1.116	9.09
27) SURR	Fluoranthene-d10	0.912	0.887	0.935	0.883	0.954	1.108	1.350	1.004	16.98
28)	Fluoranthene	1.257	1.224	1.310	1.269	1.393	1.578	1.637	1.381	11.91
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.578	1.618	1.557	1.473	1.603	1.542	1.590	1.566	3.09
31) S	Terphenyl-d14	0.806	0.833	0.801	0.754	0.812	0.784	0.892	0.812	5.30
32)	Benzo(a)anthra...	1.386	1.345	1.347	1.297	1.410	1.459	1.522	1.395	5.46
33)	Chrysene	1.523	1.491	1.496	1.402	1.510	1.506	1.543	1.496	3.01
34)	Bis(2-ethylhex...	0.440	0.376	0.345	0.380	0.438	0.504	0.414	13.96	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN091125.M

36)	Indeno(1,2,3-c...	1.517	1.491	1.543	1.532	1.791	1.865	2.025	1.681	12.58
37)	Benzo(b)fluora...	1.463	1.475	1.554	1.444	1.636	1.674	1.778	1.575	7.99
38)	Benzo(k)fluora...	1.516	1.543	1.524	1.501	1.705	1.696	1.756	1.606	6.73
39) C	Benzo(a)pyrene	1.155	1.172	1.201	1.151	1.314	1.367	1.467	1.261	9.78
40)	Dibenzo(a,h)an...	1.157	1.175	1.210	1.232	1.463	1.519	1.636	1.342	14.39
41)	Benzo(g,h,i)pe...	1.413	1.402	1.451	1.375	1.568	1.607	1.723	1.506	8.59

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3086
 Instrument ID: BNA_N Calibration Date/Time: 09/17/2025 11:25
 Lab File ID: BN037740.D Init. Calib. Date(s): 09/11/2025 09/11/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 09:56 13:33
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.517		-2.5	20.0
Fluoranthene-d10	1.004	0.936		-6.8	20.0
2-Fluorophenol	0.918	0.926		0.9	20.0
Phenol-d6	1.191	1.176		-1.3	20.0
Nitrobenzene-d5	0.293	0.284		-3.1	20.0
2-Fluorobiphenyl	1.673	1.620		-3.2	20.0
2,4,6-Tribromophenol	0.104	0.123		18.3	20.0
Terphenyl-d14	0.812	0.757		-6.8	20.0
1,4-Dioxane	0.417	0.449		7.7	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3086
Instrument ID: BNA_N Calibration Date/Time: 09/17/2025 21:50
Lab File ID: BN037757.D Init. Calib. Date(s): 09/11/2025 09/11/2025
EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:56 13:33
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.503		-5.1	50.0
Fluoranthene-d10	1.004	0.936		-6.8	50.0
2-Fluorophenol	0.918	0.971		5.8	50.0
Phenol-d6	1.191	1.159		-2.7	50.0
Nitrobenzene-d5	0.293	0.283		-3.4	50.0
2-Fluorobiphenyl	1.673	1.646		-1.6	50.0
2,4,6-Tribromophenol	0.104	0.122		17.3	50.0
Terphenyl-d14	0.812	0.761		-6.3	50.0
1,4-Dioxane	0.417	0.432		3.6	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3086
 Instrument ID: BNA_N Calibration Date/Time: 09/18/2025 14:32
 Lab File ID: BN037776.D Init. Calib. Date(s): 09/11/2025 09/11/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 09:56 13:33
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.497		-6.2	20.0
Fluoranthene-d10	1.004	0.839		-16.4	20.0
2-Fluorophenol	0.918	0.890		-3.0	20.0
Phenol-d6	1.191	1.130		-5.1	20.0
Nitrobenzene-d5	0.293	0.298		1.7	20.0
2-Fluorobiphenyl	1.673	1.679		0.4	20.0
2,4,6-Tribromophenol	0.104	0.114		9.6	20.0
Terphenyl-d14	0.812	0.801		-1.4	20.0
1,4-Dioxane	0.417	0.428		2.6	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3086</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>09/19/2025 00:52</u>
Lab File ID:	<u>BN037793.D</u>	Init. Calib. Date(s):	<u>09/11/2025 09/11/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>09:56 13:33</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.496		-6.4	50.0
Fluoranthene-d10	1.004	0.954		-5.0	50.0
2-Fluorophenol	0.918	0.923		0.5	50.0
Phenol-d6	1.191	1.148		-3.6	50.0
Nitrobenzene-d5	0.293	0.295		0.7	50.0
2-Fluorobiphenyl	1.673	1.632		-2.5	50.0
2,4,6-Tribromophenol	0.104	0.127		22.1	50.0
Terphenyl-d14	0.812	0.785		-3.3	50.0
1,4-Dioxane	0.417	0.436		4.6	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3086</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>09/19/2025 02:47</u>
Lab File ID:	<u>BN037795.D</u>	Init. Calib. Date(s):	<u>09/11/2025 09/11/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>09:56 13:33</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.490		-7.5	20.0
Fluoranthene-d10	1.004	0.858		-14.5	20.0
2-Fluorophenol	0.918	0.830		-9.6	20.0
Phenol-d6	1.191	1.152		-3.3	20.0
Nitrobenzene-d5	0.293	0.308		5.1	20.0
2-Fluorobiphenyl	1.673	1.696		1.4	20.0
2,4,6-Tribromophenol	0.104	0.120		15.4	20.0
Terphenyl-d14	0.812	0.771		-5.0	20.0
1,4-Dioxane	0.417	0.391		-6.2	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3086
 Instrument ID: BNA_N Calibration Date/Time: 09/19/2025 13:03
 Lab File ID: BN037811.D Init. Calib. Date(s): 09/11/2025 09/11/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:56 13:33
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.497		-6.2	50.0
Fluoranthene-d10	1.004	0.957		-4.7	50.0
2-Fluorophenol	0.918	1.044		13.7	50.0
Phenol-d6	1.191	1.176		-1.3	50.0
Nitrobenzene-d5	0.293	0.292		-0.3	50.0
2-Fluorobiphenyl	1.673	1.615		-3.5	50.0
2,4,6-Tribromophenol	0.104	0.131		26.0	50.0
Terphenyl-d14	0.812	0.765		-5.8	50.0
1,4-Dioxane	0.417	0.424		1.7	50.0

All other compounds must meet a minimum RRF of 0.010.

LAB CHRONICLE

OrderID:	Q3086	OrderDate:	9/12/2025 10:59:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3086-01	RW8-SP100-2025091 1	Water	Metals Group4	6010D	09/11/25	09/15/25	09/16/25	09/12/25
Q3086-02	RW8-SP303-2025091 1	Water	Metals Group4	6010D	09/11/25	09/15/25	09/16/25	09/12/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q3086	Order ID:	Q3086
Client:	Tetra Tech NUS, Inc.	Project ID:	NWIRP Bethpage 112G08005-WE13

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW8-SP100-20250911								
Q3086-01	RW8-SP100-20250911	Water	Iron	662		11.7	40.0	50.0	ug/L
Client ID :	RW8-SP303-20250911								
Q3086-02	RW8-SP303-20250911	Water	Iron	30.3	J	11.7	40.0	50.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP100-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	662		1	11.7	40.0	50.0	ug/L	09/15/25 10:20	09/16/25 12:36	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP303-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	30.3	J	1	11.7	40.0	50.0	ug/L	09/15/25 10:20	09/16/25 12:40	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	90 - 110	P	09/16/2025	11:19	LB137209

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	97.4	100	97	80 - 120	P	09/16/2025	11:29	LB137209

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5020	5000	100	90 - 110	P	09/16/2025	11:59	LB137209
CCV02	Iron	4880	5000	98	90 - 110	P	09/16/2025	13:39	LB137209
CCV03	Iron	4940	5000	99	90 - 110	P	09/16/2025	14:40	LB137209
CCV04	Iron	4850	5000	97	90 - 110	P	09/16/2025	15:32	LB137209
CCV05	Iron	5020	5000	100	90 - 110	P	09/16/2025	16:27	LB137209
CCV06	Iron	4950	5000	99	90 - 110	P	09/16/2025	16:48	LB137209



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	98.2	100	98	65 - 135	P	09/16/2025	11:38	LB137209



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	11:33	LB137209

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	12:05	LB137209
CCB02	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	13:43	LB137209
CCB03	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	14:47	LB137209
CCB04	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	15:36	LB137209
CCB05	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	16:31	LB137209
CCB06	Iron	23.4	+/-50	U	80.0	100	P	09/16/2025	16:53	LB137209

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169686BL		WATER		Batch Number:		PB169686		Prep Date:	09/15/2025	
	Iron	11.7	<25	U	40.0	50.0	P	09/16/2025	12:23	LB137209

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	101000	101000	100	85600	116500	09/16/2025	11:42	LB137209
ICSAB01	Iron	99000	99300	100	84400	114500	09/16/2025	11:46	LB137209



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client:	<u>Tetra Tech NUS, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q3086</u>
contract:	<u>TETR06</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3086-02</u>	client id:	<u>RW8-SP303-20250911MS</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3086-02MS	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1380		30.3	J	1500	90		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Tetra Tech NUS, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q3086</u>
contract:	<u>TETR06</u>			lab code:	<u>ACE</u>
matrix:	<u>Water</u>	sample id:	<u>Q3086-02</u>	client id:	<u>RW8-SP303-20250911MSD</u>
Percent Solids for Sample:	NA	Spiked ID:	Q3086-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1370		30.3	J	1500	90		P

Metals
- 5b -

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Matrix: _____

Level: LOW

Client ID: _____

Sample ID: _____

Spiked ID: _____

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

A
B
C
D
E
F
G
H

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3086</u>
Contract:	<u>TETR06</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3086-02</u>	Client ID:	<u>RW8-SP303-20250911DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3086-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	30.3	J	32.3	J	6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3086</u>
Contract:	<u>TETR06</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3086-02MS</u>	Client ID:	<u>RW8-SP303-20250911MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3086-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	1380		1370		1		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169686BS Iron	ug/L	1500	1470		98	87 - 115	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

RW8-SP303-20250911L

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **Lb No.:** lb137209 **Lab Sample ID :** Q3086-02L **SDG No.:** Q3086
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
Iron	30.3	J	250	U	100.0		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Method:

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169686							
PB169686BL	PB169686BL	MB	WATER	09/15/2025	50.0	25.0	
PB169686BS	PB169686BS	LCS	WATER	09/15/2025	50.0	25.0	
Q3086-01	RW8-SP100-20250911	SAM	WATER	09/15/2025	50.0	25.0	
Q3086-02	RW8-SP303-20250911	SAM	WATER	09/15/2025	50.0	25.0	
Q3086-02DUP	RW8-SP303-20250911DUP	DUP	WATER	09/15/2025	50.0	25.0	
Q3086-02MS	RW8-SP303-20250911MS	MS	WATER	09/15/2025	50.0	25.0	
Q3086-02MSD	RW8-SP303-20250911MSD	MSD	WATER	09/15/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Tetra Tech NUS, Inc.

Contract: TETR06

Lab code: ACE

Sdg no.: Q3086

Instrument id number: _____

Method: _____

Run number: LB137209

Start date: 09/16/2025

End date: 09/16/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1053	Fe
S1	S1	1	1058	Fe
S2	S2	1	1102	Fe
S3	S3	1	1106	Fe
S4	S4	1	1110	Fe
S5	S5	1	1115	Fe
ICV01	ICV01	1	1119	Fe
LLICV01	LLICV01	1	1129	Fe
ICB01	ICB01	1	1133	Fe
CRI01	CRI01	1	1138	Fe
ICSA01	ICSA01	1	1142	Fe
ICSAB01	ICSAB01	1	1146	Fe
CCV01	CCV01	1	1159	Fe
CCB01	CCB01	1	1205	Fe
PB169686BL	PB169686BL	1	1223	Fe
PB169686BS	PB169686BS	1	1227	Fe
Q3086-01	RW8-SP100-20250911	1	1236	Fe
Q3086-02	RW8-SP303-20250911	1	1240	Fe
Q3086-02DUP	RW8-SP303-20250911DUP	1	1250	Fe
Q3086-02L	RW8-SP303-20250911L	5	1254	Fe
Q3086-02MS	RW8-SP303-20250911MS	1	1305	Fe
CCV02	CCV02	1	1339	Fe
CCB02	CCB02	1	1343	Fe
Q3086-02MSD	RW8-SP303-20250911MSD	1	1347	Fe
CCV03	CCV03	1	1440	Fe
CCB03	CCB03	1	1447	Fe
CCV04	CCV04	1	1532	Fe
CCB04	CCB04	1	1536	Fe
CCV05	CCV05	1	1627	Fe
CCB05	CCB05	1	1631	Fe
CCV06	CCV06	1	1648	Fe
CCB06	CCB06	1	1653	Fe

LAB CHRONICLE

OrderID:	Q3086	OrderDate:	9/12/2025 10:59:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3086-01	RW8-SP100-2025091 1	WATER			09/11/25 13:05			09/12/25
			TDS	SM2540 C			09/12/25 16:30	
			TSS	SM2540 D			09/15/25 11:00	
Q3086-02	RW8-SP303-2025091 1	WATER			09/11/25 13:13			09/12/25
			TDS	SM2540 C			09/12/25 16:30	
			TSS	SM2540 D			09/15/25 11:00	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25 13:05
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP100-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	1.00	J	1	1.00	10.0	10.0	mg/L		09/12/25 16:30	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		09/15/25 11:00	SM 2540 D-20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/11/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/12/25
Client Sample ID:	RW8-SP303-20250911	SDG No.:	Q3086
Lab Sample ID:	Q3086-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	2.00	J	1	1.00	10.0	10.0	mg/L		09/12/25 16:30	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		09/15/25 11:00	SM 2540 D-20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Preparation Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Project: NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137186BL							
TSS	mg/L	1	2.0000	J	1	4	09/15/2025
Sample ID: LB137187BL							
TDS	mg/L	1	5.0000	J	1.0	10	09/12/2025

Duplicate Sample Summary

Client: Tetra Tech NUS, Inc. Project: NWIRP Bethpage 112G08005-WE13 Client ID: COMPDUP	SDG No.: Q3086 Sample ID: Q3067-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	241		238		1	1.25		09/15/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q3086
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q3086-02
Client ID:	RW8-SP303-20250911DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	2.00	J	2.00	J	1	0		09/12/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137186

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137186BS							
TSS	mg/L	550	531		96	1	90-110	09/15/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137187

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137187BS							
TDS	mg/L	100	93.0		93	1	90-110	09/12/2025