

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	90 - 110	P	09/16/2025	11:19	LB137209

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	97.4	100	97	80 - 120	P	09/16/2025	11:29	LB137209

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3086

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5020	5000	100	90 - 110	P	09/16/2025	11:59	LB137209
CCV02	Iron	4880	5000	98	90 - 110	P	09/16/2025	13:39	LB137209
CCV03	Iron	4940	5000	99	90 - 110	P	09/16/2025	14:40	LB137209
CCV04	Iron	4850	5000	97	90 - 110	P	09/16/2025	15:32	LB137209
CCV05	Iron	5020	5000	100	90 - 110	P	09/16/2025	16:27	LB137209
CCV06	Iron	4950	5000	99	90 - 110	P	09/16/2025	16:48	LB137209