

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3086</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>09/17/2025 11:25</u>
Lab File ID:	<u>BN037740.D</u>	Init. Calib. Date(s):	<u>09/11/2025 09/11/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>09:56 13:33</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.530	0.517		-2.5	20.0
Fluoranthene-d10	1.004	0.936		-6.8	20.0
2-Fluorophenol	0.918	0.926		0.9	20.0
Phenol-d6	1.191	1.176		-1.3	20.0
Nitrobenzene-d5	0.293	0.284		-3.1	20.0
2-Fluorobiphenyl	1.673	1.620		-3.2	20.0
2,4,6-Tribromophenol	0.104	0.123		18.3	20.0
Terphenyl-d14	0.812	0.757		-6.8	20.0
1,4-Dioxane	0.417	0.449		7.7	20.0

All other compounds must meet a minimum RRF of 0.010.