

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091725\
 Data File : BN037757.D
 Acq On : 17 Sep 2025 21:50
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 18 05:02:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	133	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	136	0.00
3	n-Nitrosodimethylamine	0.400	0.384	4.0	123	0.00
4 S	2-Fluorophenol	0.400	0.423	-5.7	145	0.00
5 S	Phenol-d6	0.400	0.389	2.8	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	134	-0.01
8 S	Nitrobenzene-d5	0.400	0.386	3.5	134	-0.01
9	Naphthalene	0.400	0.390	2.5	133	0.00
10	Hexachlorobutadiene	0.400	0.390	2.5	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.380	5.0	136	-0.01
12	2-Methylnaphthalene	0.400	0.385	3.8	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	133	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.467	-16.8	161	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	133	0.00
16	Acenaphthylene	0.400	0.378	5.5	129	0.00
17	Acenaphthene	0.400	0.382	4.5	130	-0.01
18	Fluorene	0.400	0.375	6.3	130	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	131	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.336	16.0	123	0.00
21	4-Bromophenyl-phenylether	0.400	0.389	2.8	132	-0.01
22	Hexachlorobenzene	0.400	0.406	-1.5	135	0.00
23	Atrazine	0.400	0.375	6.3	134	0.00
24	Pentachlorophenol	0.400	0.422	-5.5	161	0.00
25	Phenanthrene	0.400	0.389	2.8	131	-0.01
26	Anthracene	0.400	0.377	5.8	128	0.00
27 SURR	Fluoranthene-d10	0.400	0.373	6.8	131	0.00
28	Fluoranthene	0.400	0.379	5.3	130	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	139	0.00
30	Pyrene	0.400	0.378	5.5	132	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	132	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	134	0.00
33	Chrysene	0.400	0.395	1.3	137	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.399	0.3	152	0.00
35 I	Perylene-d12	0.400	0.400	0.0	150	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.420	-5.0	172	-0.01
37	Benzo(b)fluoranthene	0.400	0.378	5.5	144	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	155	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	147	0.00
40	Dibenzo(a,h)anthracene	0.400	0.425	-6.2	177	-0.01
41	Benzo(g,h,i)perylene	0.400	0.395	1.3	154	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4