

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091825\
 Data File : BN037811.D
 Acq On : 19 Sep 2025 13:03
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 13:27:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	117	0.00
3	n-Nitrosodimethylamine	0.400	0.415	-3.7	117	0.00
4 S	2-Fluorophenol	0.400	0.455	-13.7	136	0.00
5 S	Phenol-d6	0.400	0.395	1.3	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	119	-0.01
8 S	Nitrobenzene-d5	0.400	0.399	0.3	122	-0.01
9	Naphthalene	0.400	0.391	2.3	118	-0.01
10	Hexachlorobutadiene	0.400	0.387	3.3	115	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.375	6.3	118	-0.02
12	2-Methylnaphthalene	0.400	0.384	4.0	117	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.501	-25.2#	154	-0.01
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	116	-0.01
16	Acenaphthylene	0.400	0.370	7.5	113	0.00
17	Acenaphthene	0.400	0.379	5.3	115	-0.01
18	Fluorene	0.400	0.382	4.5	118	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.332	17.0	112	-0.01
21	4-Bromophenyl-phenylether	0.400	0.369	7.8	115	-0.01
22	Hexachlorobenzene	0.400	0.401	-0.3	123	-0.01
23	Atrazine	0.400	0.375	6.3	123	-0.01
24	Pentachlorophenol	0.400	0.471	-17.7	165	0.00
25	Phenanthrene	0.400	0.391	2.3	121	-0.01
26	Anthracene	0.400	0.378	5.5	118	-0.01
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	123	0.00
28	Fluoranthene	0.400	0.391	2.3	124	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	132	0.00
30	Pyrene	0.400	0.377	5.8	125	0.00
31 S	Terphenyl-d14	0.400	0.377	5.8	126	-0.01
32	Benzo(a)anthracene	0.400	0.379	5.3	129	0.00
33	Chrysene	0.400	0.396	1.0	131	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.446	-11.5	162	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	144	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	175	-0.02
37	Benzo(b)fluoranthene	0.400	0.389	2.8	142	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	153	0.00
39 C	Benzo(a)pyrene	0.400	0.388	3.0	147	0.00
40	Dibenzo(a,h)anthracene	0.400	0.437	-9.2	175	-0.02
41	Benzo(g,h,i)perylene	0.400	0.391	2.3	146	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4