

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/17/2025

Lab File ID: BN037740.D

Time Analyzed: 11:25

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6660	7.84	18396	10.65	8930	14.50
UPPER LIMIT	13320	8.34	36792	11.148	17860	14.995
LOWER LIMIT	3330	7.34	9198	10.148	4465	13.995
EPA SAMPLE NO.						
01 RW8-SP100-20250911	6573	7.84	18040	10.64	8326	14.50
02 RW8-SP303-20250911	6528	7.84	17791	10.64	8068	14.50
03 MW203D2-20250909MS	6853	7.84	18950	10.64	8815	14.49
04 MW203D2-20250909MSD	7569	7.85	20880	10.65	9823	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.