

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/18/2025

Lab File ID: BN037776.D

Time Analyzed: 14:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7127	7.833	18984	10.64	8698	14.48
UPPER LIMIT	14254	8.333	37968	11.137	17396	14.984
LOWER LIMIT	3563.5	7.333	9492	10.137	4349	13.984
EPA SAMPLE NO.						
01 PB169692BS	6983	7.83	18103	10.64	7729	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.