

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID : SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037795.D

Time Analyzed: 02:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7724	7.833	19361	10.64	8621	14.48
UPPER LIMIT	15448	8.333	38722	11.137	17242	14.984
LOWER LIMIT	3862	7.333	9680.5	10.137	4310.5	13.984
EPA SAMPLE NO.						
01 PB169692BL	7701	7.83	18665	10.64	7806	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.