

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3086

Client ID: SSTDCCC0.4

Date Analyzed: 09/19/2025

Lab File ID: BN037795.D

Time Analyzed: 02:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14380	17.235	11169	21.411	11840	23.742
UPPER LIMIT	28760	17.735	22338	21.911	23680	24.242
LOWER LIMIT	7190	16.735	5584.5	20.911	5920	23.242
EPA SAMPLE NO.						
PB169692BL	12511	17.24	10022	21.41	10278	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.