

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091625

Review By	John Carlone	Review On	9/17/2025 8:23:11 AM
Supervise By	Mahesh Dadoda	Supervise On	9/18/2025 3:21:51 PM
SubDirectory	VN091625	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135496
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135497,VP135498

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087846.D	16 Sep 2025 11:25		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087847.D	16 Sep 2025 13:55	pH#Lot#V12668	JC\MD	Ok,M
3	VN0916MBL01	VN0916MBL01	VN087848.D	16 Sep 2025 14:43		JC\MD	Ok
4	VN0916WBL01	VN0916WBL01	VN087849.D	16 Sep 2025 15:04		JC\MD	Ok
5	VN0916WBS01	VN0916WBS01	VN087850.D	16 Sep 2025 15:25	BS Failed Low For Com.#80,81	JC\MD	Ok,M
6	VN0916WBSD01	VN0916WBSD01	VN087851.D	16 Sep 2025 15:46	BSD Failed Low For Comp.#23	JC\MD	Ok,M
7	Q3103-01	TW-WTS-14	VN087852.D	16 Sep 2025 16:08	vial A pH<2	JC\MD	Ok
8	Q3108-01	MW2	VN087853.D	16 Sep 2025 16:29	vial A pH<2	JC\MD	Ok,M
9	Q3109-01	MW1D	VN087854.D	16 Sep 2025 16:50	vial A pH<2	JC\MD	Ok
10	Q3109-02	MW1	VN087855.D	16 Sep 2025 17:11	vial A pH<2 Need 20X	JC\MD	Dilution
11	VN0916MBS01	VN0916MBS01	VN087856.D	16 Sep 2025 17:32		JC\MD	Ok,M
12	VN0916MBSD01	VN0916MBSD01	VN087857.D	16 Sep 2025 17:54		JC\MD	Ok,M
13	Q3077-01	LAW-25-0137	VN087858.D	16 Sep 2025 18:15	cloth material	JC\MD	Ok
14	Q3021-01	821-B	VN087859.D	16 Sep 2025 18:36	sample is oil	JC\MD	Ok
15	Q3021-03	821-D	VN087860.D	16 Sep 2025 18:58	sample is oil	JC\MD	Ok
16	Q3021-05	821-I	VN087861.D	16 Sep 2025 19:19	sample is oil	JC\MD	Ok
17	IBLK	IBLK	VN087862.D	16 Sep 2025 19:40		JC\MD	Ok
18	Q3087-01	STORAGE-BLANK-SAI	VN087863.D	16 Sep 2025 20:01	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091625

Review By	John Carlone	Review On	9/17/2025 8:23:11 AM
Supervise By	Maresh Dadoda	Supervise On	9/18/2025 3:21:51 PM
SubDirectory	VN091625	HP Acquire Method	HP Processing Method 82N082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135496		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135497,VP135498		

19	Q3087-02	STORAGE-BLANK-WA	VN087864.D	16 Sep 2025 20:22	vial A pH<2	JC\MD	Ok
20	Q3087-03	STORAGE-BLANK-WA	VN087865.D	16 Sep 2025 20:43	vial A pH<2	JC\MD	Ok
21	Q3087-04	STORAGE-BLANK-SAI	VN087866.D	16 Sep 2025 21:05	vial A pH<2	JC\MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VN087867.D	16 Sep 2025 21:26	Low Spike	JC\MD	Not Ok

M : Manual Integration