

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3087
 Lab File ID: VN087847.D Date Analyzed: 09/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:55
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	226483	13.77				
UPPER LIMIT	452966	14.27				
LOWER LIMIT	113242	13.27				
EPA SAMPLE NO.						
STORAGE-BLANK-SAMPLE-MANAGEMENT	261215	13.77				
STORAGE-BLANK-WATER-REF-5	209847	13.77				
STORAGE-BLANK-WATER-REF-4	206680	13.77				
STORAGE-BLANK-SAMPLE-MANAGEMENT	206726	13.77				
VN0916WBL01	180365	13.77				
VN0916WBS01	210239	13.77				
VN0916WBSD01	209579	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.