

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3087

Lab File ID: VN087846.D

BFB Injection Date: 09/16/2025

Instrument ID: MSVOA_N

BFB Injection Time: 11:25

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.6
75	30.0 - 60.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	5.7
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.2
175	5.0 - 9.0% of mass 174	4.5 (6.2) 1
176	95.0 - 101.0% of mass 174	70.1 (97.1) 1
177	5.0 - 9.0% of mass 176	4.5 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087847.D	09/16/2025	13:55
VN0916WBL01	VN0916WBL01	VN087849.D	09/16/2025	15:04
VN0916WBS01	VN0916WBS01	VN087850.D	09/16/2025	15:25
VN0916WBSD01	VN0916WBSD01	VN087851.D	09/16/2025	15:46
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3087-01	VN087863.D	09/16/2025	20:01
STORAGE-BLANK-WATER-REF-5	Q3087-02	VN087864.D	09/16/2025	20:22
STORAGE-BLANK-WATER-REF-4	Q3087-03	VN087865.D	09/16/2025	20:43
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3087-04	VN087866.D	09/16/2025	21:05