



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB30	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	10:17	LB137199

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB91	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	10:22	LB137199
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	10:50	LB137199
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	11:20	LB137199
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	09/16/2025	11:50	LB137199

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	11:43	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	11:43	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	11:43	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	11:43	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	11:43	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	11:43	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	11:43	LB137198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	12:39	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	12:39	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	12:39	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	12:39	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	12:39	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	12:39	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	12:39	LB137198
CCB02	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	13:47	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	13:47	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	13:47	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	13:47	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	13:47	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	13:47	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	13:47	LB137198
CCB03	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	15:15	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	15:15	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	15:15	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	15:15	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	15:15	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	15:15	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	15:15	LB137198
CCB04	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	16:16	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	16:16	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	16:16	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	16:16	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	16:16	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	16:16	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	16:16	LB137198
CCB05	Arsenic	5.12	+/-10	U	20.0	P	09/15/2025	16:59	LB137198
	Barium	14.6	+/-50	U	100	P	09/15/2025	16:59	LB137198
	Cadmium	0.50	+/-3	U	6.00	P	09/15/2025	16:59	LB137198
	Chromium	2.12	+/-5	U	10.0	P	09/15/2025	16:59	LB137198
	Lead	2.30	+/-6	U	12.0	P	09/15/2025	16:59	LB137198
	Selenium	9.64	+/-10	U	20.0	P	09/15/2025	16:59	LB137198
	Silver	1.62	+/-5	U	10.0	P	09/15/2025	16:59	LB137198