

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	3.35			-20	20	09/15/2025	12:12	LB137198
	Barium	3.67	6.0	61	-94	106	09/15/2025	12:12	LB137198
	Cadmium	1.26	1.0	126	-5	7	09/15/2025	12:12	LB137198
	Chromium	55.6	52.0	107	42	62	09/15/2025	12:12	LB137198
	Lead	-0.061			-12	12	09/15/2025	12:12	LB137198
	Selenium	2.20			-20	20	09/15/2025	12:12	LB137198
	Silver	1.40			-10	10	09/15/2025	12:12	LB137198
ICSAB01	Arsenic	99.1	104	95	88.4	120	09/15/2025	12:16	LB137198
	Barium	473	537	88	437	637	09/15/2025	12:16	LB137198
	Cadmium	963	972	99	826	1120	09/15/2025	12:16	LB137198
	Chromium	537	542	99	460	624	09/15/2025	12:16	LB137198
	Lead	45.1	49.0	92	37	61	09/15/2025	12:16	LB137198
	Selenium	49.1	46.0	107	26	66	09/15/2025	12:16	LB137198
	Silver	204	201	102	170	232	09/15/2025	12:16	LB137198