

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169685BS							
Arsenic	ug/L	4000	3830		96	80 - 120	P
Barium	ug/L	1000	909		91	80 - 120	P
Cadmium	ug/L	1000	931		93	80 - 120	P
Chromium	ug/L	2000	1940		97	80 - 120	P
Lead	ug/L	5000	4640		93	80 - 120	P
Selenium	ug/L	10000	9750		98	80 - 120	P
Silver	ug/L	380	365		96	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Contract: ATLA10

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169703BS Mercury	ug/L	4.0	3.89		97	80 - 120	CV