

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Atlantic Recovery Services, Inc. **Level:** LOW **SDG No.:** Q3092
Contract: ATLA10 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3092-04 **Client ID:** 279182DUP
Percent Solids for Sample: NA **Duplicate ID** Q3092-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1200		1140		5		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	19.3	J	16.2	J	17		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	13.1	J	13.9	J	6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Atlantic Recovery Services, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3092</u>
Contract:	<u>ATLA10</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3092-04MS</u>	Client ID:	<u>279182MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3092-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3640		3960		8		P
Barium	ug/L	20	1840		2180		17		P
Cadmium	ug/L	20	874		952		9		P
Chromium	ug/L	20	1740		1910		9		P
Lead	ug/L	20	4150		4510		8		P
Mercury	ug/L	20	40.0		39.0		3		CV
Selenium	ug/L	20	9180		10000		9		P
Silver	ug/L	20	338		378		11		P