

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 21:53

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.29	7.29	7.19	7.39	0.01
2,4-D	8.42	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.29	9.29	9.19	9.39	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 21:53

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.75	7.74	7.64	7.84	0.00
2,4-D	9.00	9.00	8.90	9.10	0.00
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00