

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031782.D **Time Analyzed:** 10:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.284	9.185	9.385	755.340	712.500	6.0
2,4-D	8.413	8.314	8.514	736.820	705.000	4.5
2,4-DCAA	7.284	7.185	7.385	782.620	750.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031782.D **Time Analyzed:** 10:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.892	9.793	9.993	707.640	712.500	-0.7
2,4-D	9.000	8.900	9.100	682.140	705.000	-3.2
2,4-DCAA	7.743	7.643	7.843	725.050	750.000	-3.3