

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031806.D **Time Analyzed:** 02:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.284	9.185	9.385	774.030	712.500	8.6
2,4-D	8.411	8.314	8.514	789.270	705.000	12.0
2,4-DCAA	7.283	7.185	7.385	724.500	750.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031806.D **Time Analyzed:** 02:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.892	9.793	9.993	734.360	712.500	3.1
2,4-D	8.999	8.900	9.100	734.670	705.000	4.2
2,4-DCAA	7.743	7.643	7.843	783.230	750.000	4.4