

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 11:33

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.00       | 9.00      | 8.90                   | 9.10 | 0.00       |
| Tetrachloro-m-xylene | 3.53       | 3.53      | 3.43                   | 3.63 | 0.00       |
| gamma-BHC (Lindane)  | 4.30       | 4.30      | 4.20                   | 4.40 | 0.00       |
| Heptachlor           | 4.89       | 4.89      | 4.79                   | 4.99 | 0.00       |
| Heptachlor epoxide   | 5.65       | 5.65      | 5.55                   | 5.75 | 0.00       |
| Endrin               | 6.53       | 6.53      | 6.43                   | 6.63 | 0.00       |
| Methoxychlor         | 7.45       | 7.45      | 7.35                   | 7.55 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 11:33

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Decachlorobiphenyl   | 7.98       | 7.98      | 7.88      | 8.08 | 0.00       |
| Tetrachloro-m-xylene | 2.82       | 2.82      | 2.72      | 2.92 | 0.00       |
| gamma-BHC (Lindane)  | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Heptachlor           | 4.01       | 4.01      | 3.91      | 4.11 | 0.00       |
| Heptachlor epoxide   | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Endrin               | 5.70       | 5.70      | 5.60      | 5.80 | 0.00       |
| Methoxychlor         | 6.67       | 6.67      | 6.57      | 6.77 | 0.00       |