

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.001	8.902	9.102	48.730	50.000	-2.5
Endrin	6.531	6.432	6.632	46.480	50.000	-7.0
gamma-BHC (Lindane)	4.302	4.202	4.402	51.410	50.000	2.8
Heptachlor	4.892	4.793	4.993	49.840	50.000	-0.3
Heptachlor epoxide	5.651	5.552	5.752	50.370	50.000	0.7
Methoxychlor	7.452	7.352	7.552	44.040	50.000	-11.9
Tetrachloro-m-xylene	3.527	3.428	3.628	51.250	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.984	7.884	8.084	40.500	50.000	-19.0
Endrin	5.702	5.603	5.803	44.550	50.000	-10.9
gamma-BHC (Lindane)	3.661	3.561	3.761	49.630	50.000	-0.7
Heptachlor	4.009	3.909	4.109	48.080	50.000	-3.8
Heptachlor epoxide	4.792	4.693	4.893	48.930	50.000	-2.1
Methoxychlor	6.674	6.574	6.774	41.040	50.000	-17.9
Tetrachloro-m-xylene	2.824	2.724	2.924	49.030	50.000	-1.9