

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.902	9.102	53.620	50.000	7.2
Endrin	6.531	6.432	6.632	50.770	50.000	1.5
gamma-BHC (Lindane)	4.302	4.202	4.402	55.280	50.000	10.6
Heptachlor	4.893	4.793	4.993	53.040	50.000	6.1
Heptachlor epoxide	5.651	5.552	5.752	54.390	50.000	8.8
Methoxychlor	7.451	7.352	7.552	48.210	50.000	-3.6
Tetrachloro-m-xylene	3.527	3.428	3.628	54.690	50.000	9.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.470	50.000	-3.1
Endrin	5.702	5.603	5.803	50.690	50.000	1.4
gamma-BHC (Lindane)	3.661	3.561	3.761	54.800	50.000	9.6
Heptachlor	4.009	3.909	4.109	52.810	50.000	5.6
Heptachlor epoxide	4.792	4.693	4.893	53.620	50.000	7.2
Methoxychlor	6.673	6.574	6.774	46.500	50.000	-7.0
Tetrachloro-m-xylene	2.824	2.724	2.924	53.880	50.000	7.8