

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097267.D	18 Sep 2025 16:11	AR\AJ	Ok
2	I.BLK	PL097268.D	18 Sep 2025 16:25	AR\AJ	Ok
3	PEM	PL097269.D	18 Sep 2025 16:39	AR\AJ	Ok,M
4	RESCHK	PL097270.D	18 Sep 2025 16:52	AR\AJ	Ok
5	PSTDICC100	PL097271.D	18 Sep 2025 17:06	AR\AJ	Ok
6	PSTDICC075	PL097272.D	18 Sep 2025 17:19	AR\AJ	Ok
7	PSTDICC050	PL097273.D	18 Sep 2025 17:33	AR\AJ	Ok
8	PSTDICC025	PL097274.D	18 Sep 2025 17:47	AR\AJ	Ok
9	PSTDICC005	PL097275.D	18 Sep 2025 18:00	AR\AJ	Ok,M
10	PCHLORICC1000	PL097276.D	18 Sep 2025 18:14	AR\AJ	Ok
11	PCHLORICC750	PL097277.D	18 Sep 2025 18:28	AR\AJ	Ok
12	PCHLORICC500	PL097278.D	18 Sep 2025 18:41	AR\AJ	Ok
13	PCHLORICC250	PL097279.D	18 Sep 2025 18:55	AR\AJ	Ok
14	PCHLORICC050	PL097280.D	18 Sep 2025 19:08	AR\AJ	Ok
15	PTOXICC1000	PL097281.D	18 Sep 2025 19:22	AR\AJ	Ok
16	PTOXICC750	PL097282.D	18 Sep 2025 19:35	AR\AJ	Ok,M
17	PTOXICC500	PL097283.D	18 Sep 2025 19:49	AR\AJ	Ok
18	PTOXICC250	PL097284.D	18 Sep 2025 20:03	AR\AJ	Ok
19	PTOXICC100	PL097285.D	18 Sep 2025 20:16	AR\AJ	Ok,M
20	PSTDICV050	PL097286.D	19 Sep 2025 08:39	AR\AJ	Ok,M
21	PCHLORICV500	PL097287.D	19 Sep 2025 08:52	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL097288.D	19 Sep 2025 09:25	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097314.D	22 Sep 2025 10:12	AR\AJ	Ok
2	I.BLK	PL097315.D	22 Sep 2025 11:06	AR\AJ	Ok
3	PEM	PL097316.D	22 Sep 2025 11:20	AR\AJ	Ok,M
4	PSTDCCC050	PL097317.D	22 Sep 2025 11:33	AR\AJ	Ok,M
5	PB169749BS	PL097318.D	22 Sep 2025 12:31	AR\AJ	Ok,M
6	PB169772BL	PL097319.D	22 Sep 2025 13:30	AR\AJ	Ok
7	PB169772BS	PL097320.D	22 Sep 2025 13:44	AR\AJ	Ok,M
8	Q3152-01	PL097321.D	22 Sep 2025 14:05	AR\AJ	Ok,M
9	Q3135-01	PL097322.D	22 Sep 2025 14:19	AR\AJ	Ok,M
10	Q3151-02	PL097323.D	22 Sep 2025 14:32	AR\AJ	Ok
11	PB169766BL	PL097324.D	22 Sep 2025 14:46	AR\AJ	Ok
12	Q3153-04	PL097325.D	22 Sep 2025 17:30	AR\AJ	Ok
13	Q3153-08	PL097326.D	22 Sep 2025 17:44	AR\AJ	Ok
14	I.BLK	PL097327.D	22 Sep 2025 17:58	AR\AJ	Ok
15	PSTDCCC050	PL097328.D	22 Sep 2025 18:11	AR\AJ	Ok,M
16	PB169766BS	PL097329.D	22 Sep 2025 18:25	AR\AJ	Not Ok
17	PB169766BSD	PL097330.D	22 Sep 2025 18:39	AR\AJ	Not Ok
18	PB169782BL	PL097331.D	22 Sep 2025 18:52	AR\AJ	Ok
19	PB169782BS	PL097332.D	22 Sep 2025 19:06	AR\AJ	Not Ok
20	Q3065-04	PL097333.D	22 Sep 2025 19:20	AR\AJ	Ok
21	Q3092-02	PL097334.D	22 Sep 2025 19:33	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3092-02MS	PL097335.D	22 Sep 2025 19:47	AR\AJ	Ok,M
23	Q3092-02MSD	PL097336.D	22 Sep 2025 20:00	AR\AJ	Ok,M
24	PB169677TB	PL097337.D	22 Sep 2025 20:14	AR\AJ	Ok
25	PB169754TB	PL097338.D	22 Sep 2025 20:28	AR\AJ	Ok
26	I.BLK	PL097339.D	22 Sep 2025 20:41	AR\AJ	Ok
27	PEM	PL097340.D	22 Sep 2025 20:55	AR\AJ	Ok,M
28	PSTDCCC050	PL097341.D	22 Sep 2025 21:08	AR\AJ	Ok,M
29	Q3092-04	PL097342.D	22 Sep 2025 21:49	AR\AJ	Ok,M
30	Q3157-02	PL097343.D	22 Sep 2025 22:03	AR\AJ	Ok
31	Q3145-01	PL097344.D	22 Sep 2025 22:16	AR\AJ	Ok,M
32	Q3145-03	PL097345.D	22 Sep 2025 22:30	AR\AJ	Ok,M
33	Q3153-01	PL097346.D	22 Sep 2025 22:44	AR\AJ	Ok,M
34	Q3153-01MS	PL097347.D	22 Sep 2025 22:57	AR\AJ	Ok,M
35	Q3153-01MSD	PL097348.D	22 Sep 2025 23:11	AR\AJ	Ok,M
36	Q3153-05	PL097349.D	22 Sep 2025 23:24	AR\AJ	Ok,M
37	Q3155-01	PL097350.D	22 Sep 2025 23:38	AR\AJ	Ok,M
38	Q3155-02	PL097351.D	22 Sep 2025 23:52	AR\AJ	Ok,M
39	Q3155-03	PL097352.D	23 Sep 2025 00:05	AR\AJ	Ok,M
40	Q3157-01	PL097353.D	23 Sep 2025 00:19	AR\AJ	Ok,M
41	I.BLK	PL097354.D	23 Sep 2025 00:32	AR\AJ	Ok
42	PSTDCCC050	PL097355.D	23 Sep 2025 00:46	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092325

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	PL092325	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097356.D	23 Sep 2025 09:02	AR\AJ	Ok
2	I.BLK	PL097357.D	23 Sep 2025 09:16	AR\AJ	Ok
3	PEM	PL097358.D	23 Sep 2025 09:29	AR\AJ	Ok,NR
4	PSTDCCC050	PL097359.D	23 Sep 2025 10:03	AR\AJ	Ok,NR
5	PB169766BS	PL097360.D	23 Sep 2025 11:14	AR\AJ	Ok,NR
6	PB169766BSD	PL097361.D	23 Sep 2025 11:34	AR\AJ	Ok,NR
7	PB169782BS	PL097362.D	23 Sep 2025 11:47	AR\AJ	Ok,NR
8	I.BLK	PL097363.D	23 Sep 2025 12:09	AR\AJ	Ok
9	PSTDCCC050	PL097364.D	23 Sep 2025 12:52	AR\AJ	Ok,NR

M : Manual Integration