

Cover Page

Order ID : Q3092

Project ID : PGW Phila. Gas Works

Client : Atlantic Recovery Services, Inc.

Lab Sample Number

Q3092-01
Q3092-02
Q3092-03
Q3092-04

Client Sample Number

291376
291376
279182
279182

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 9/24/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Atlantic Recovery Services, Inc.

Project Name: PGW Phila. Gas Works

Project # N/A

Order ID # Q3092

Test Name: TCLP Herbicide, TCLP Pesticide

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 09/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:

. This data package contains results for TCLP Herbicide(8151A), TCLP Pesticide(8081B).

C. Analytical Techniques:

TCLP Herbicide : The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 um df,: Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324. The analysis of TCLP Herbicides was based on method 8151A and extraction was done based on method 3510 and TCLP extraction method was 1311.

TCLP Pesticide : The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df,: Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of TCLP Pesticides was based on method 8081B and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3092

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRADIP PRAJAPATI

Date: 09/24/2025

LAB CHRONICLE

OrderID: Q3092
Client: Atlantic Recovery Services, Inc.
Contact: Kenny Pyle

OrderDate: 9/12/2025 12:02:28 PM
Project: PGW Phila. Gas Works
Location: D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3092-02	291376	TCLP			09/12/25			09/12/25
			TCLP Herbicide	8151A		09/22/25	09/22/25	
			TCLP Pesticide	8081B		09/22/25	09/22/25	
Q3092-04	279182	TCLP			09/12/25			09/12/25
			TCLP Herbicide	8151A		09/22/25	09/22/25	
			TCLP Pesticide	8081B		09/22/25	09/22/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3092

Order ID: Q3092

Client: Atlantic Recovery Services, Inc.

Project ID: PGW Phila. Gas Works

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q3092**

Client: **Atlantic Recovery Services, Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097268.D	PIBLK-PL097268.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	15.6	78		61	148
		Decachlorobiphen	2	20	17.1	86		57	171
		Tetrachloro-m-xyl	2	20	16.0	80		61	148
I.BLK-PL097327.D	PIBLK-PL097327.D	Decachlorobiphen	1	20	20.2	101		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	17.3	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
PB169782BL	PB169782BL	Decachlorobiphen	1	20	17.5	88		57	171
		Tetrachloro-m-xyl	1	20	16.6	83		61	148
		Decachlorobiphen	2	20	15.5	77		57	171
		Tetrachloro-m-xyl	2	20	16.9	85		61	148
Q3092-02	291376	Decachlorobiphen	1	20	16.6	83		57	171
		Tetrachloro-m-xyl	1	20	16.1	80		61	148
		Decachlorobiphen	2	20	15.0	75		57	171
		Tetrachloro-m-xyl	2	20	16.8	84		61	148
Q3092-02MS	291376MS	Decachlorobiphen	1	20	16.9	84		57	171
		Tetrachloro-m-xyl	1	20	16.4	82		61	148
		Decachlorobiphen	2	20	15.3	77		57	171
		Tetrachloro-m-xyl	2	20	17.4	87		61	148
Q3092-02MSD	291376MSD	Decachlorobiphen	1	20	16.9	85		57	171
		Tetrachloro-m-xyl	1	20	16.5	82		61	148
		Decachlorobiphen	2	20	15.1	76		57	171
		Tetrachloro-m-xyl	2	20	17.1	86		61	148
PB169677TB	PB169677TB	Decachlorobiphen	1	20	17.6	88		57	171
		Tetrachloro-m-xyl	1	20	16.4	82		61	148
		Decachlorobiphen	2	20	16.2	81		57	171
		Tetrachloro-m-xyl	2	20	16.8	84		61	148
I.BLK-PL097339.D	PIBLK-PL097339.D	Decachlorobiphen	1	20	21.0	105		57	171
		Tetrachloro-m-xyl	1	20	18.3	91		61	148
		Decachlorobiphen	2	20	19.2	96		57	171
		Tetrachloro-m-xyl	2	20	18.9	95		61	148
Q3092-04	279182	Decachlorobiphen	1	20	16.2	81		57	171
		Tetrachloro-m-xyl	1	20	15.8	79		61	148
		Decachlorobiphen	2	20	15.3	77		57	171
		Tetrachloro-m-xyl	2	20	16.6	83		61	148
I.BLK-PL097354.D	PIBLK-PL097354.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	18.1	91		61	148
		Decachlorobiphen	2	20	17.4	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
I.BLK-PL097357.D	PIBLK-PL097357.D	Decachlorobiphen	1	20	22.2	111		57	171

Surrogate Summary

SDG No.: Q3092

Client: Atlantic Recovery Services, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097357.D	PIBLK-PL097357.D	Tetrachloro-m-xyl	1	20	19.7	98		61	148
		Decachlorobiphen	2	20	20.1	100		57	171
PB169782BS	PB169782BS	Tetrachloro-m-xyl	2	20	19.3	97		61	148
		Decachlorobiphen	1	20	18.1	91		57	171
		Tetrachloro-m-xyl	1	20	17.0	85		61	148
		Decachlorobiphen	2	20	17.4	87		57	171
		Tetrachloro-m-xyl	2	20	17.4	87		61	148
		Decachlorobiphen	1	20	21.4	107		57	171
I.BLK-PL097363.D	PIBLK-PL097363.D	Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	20.4	102		57	171
		Tetrachloro-m-xyl	2	20	19.6	98		61	148
		Decachlorobiphen	1	20	21.4	107		57	171

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3092</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Atlantic Recovery Services, Inc.</u>	DataFile :	<u>PL097335.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 1)										
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.40	ug/L	88				76	144
	Methoxychlor	5	0	4.20	ug/L	84				70	142
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 2)										
	gamma-BHC (Lindane)	5	0	4.70	ug/L	94				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.50	ug/L	90				76	144
	Methoxychlor	5	0	4.00	ug/L	80				70	142

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3092</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Atlantic Recovery Services, Inc.</u>	DataFile :	<u>PL097336.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		0		60	152	20
	Heptachlor	5	0	4.50	ug/L	90		0		56	147	20
	Heptachlor epoxide	5	0	4.70	ug/L	94		0		77	143	20
	Endrin	5	0	4.50	ug/L	90		2		76	144	20
	Methoxychlor	5	0	4.20	ug/L	84		0		70	142	20
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		2		60	152	20
	Heptachlor	5	0	4.40	ug/L	88		2		56	147	20
	Heptachlor epoxide	5	0	4.60	ug/L	92		2		77	143	20
	Endrin	5	0	4.50	ug/L	90		0		76	144	20
	Methoxychlor	5	0	4.00	ug/L	80		0		70	142	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3092</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Atlantic Recovery Services, Inc.</u>	Datafile :	<u>PL097362.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169782BS (Column 1)	gamma-BHC (Lindane)	0.5	0.51	ug/L	102				82	129		
	Heptachlor	0.5	0.48	ug/L	97				79	127		
	Heptachlor epoxide	0.5	0.52	ug/L	104				81	124		
	Endrin	0.5	0.45	ug/L	91				81	128		
	Methoxychlor	0.5	0.42	ug/L	84				78	108		
PB169782BS (Column 2)	gamma-BHC (Lindane)	0.5	0.52	ug/L	104				82	129		
	Heptachlor	0.5	0.49	ug/L	97				79	127		
	Heptachlor epoxide	0.5	0.51	ug/L	102				81	124		
	Endrin	0.5	0.45	ug/L	91				81	128		
	Methoxychlor	0.5	0.41	ug/L	82				78	108		

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB169782BL

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Lab Sample ID: PB169782BL

Lab File ID: PL097331.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 09/22/2025

Date Analyzed (1): 09/22/2025

Date Analyzed (2): 09/22/2025

Time Analyzed (1): 18:52

Time Analyzed (2): 18:52

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
291376	Q3092-02	PL097334.D	09/22/2025	09/22/2025
291376MS	Q3092-02MS	PL097335.D	09/22/2025	09/22/2025
291376MSD	Q3092-02MSD	PL097336.D	09/22/2025	09/22/2025
PB169677TB	PB169677TB	PL097337.D	09/22/2025	09/22/2025
279182	Q3092-04	PL097342.D	09/22/2025	09/22/2025
PB169782BS	PB169782BS	PL097362.D	09/23/2025	09/23/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	
Project:	PGW Phila. Gas Works	Date Received:	09/22/25
Client Sample ID:	PB169677TB	SDG No.:	Q3092
Lab Sample ID:	PB169677TB	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	100 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097337.D	1	09/22/25 10:05	09/22/25 20:14	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.6		57 - 171	88%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.8		61 - 148	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:14
Operator : AR\AJ
Sample : PB169677TB
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169677TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	74085152	110.7E6	16.365	16.828
28)	SA Decachlor...	8.999	7.983	57745043	91948485	17.647	16.155

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:14
Operator : AR\AJ
Sample : PB169677TB
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
PB169677TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Signal: PL097337.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 74085152
Conc: 16.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169677TB


Signal: PL097337.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 110656413
Conc: 16.83 ng/ml


Signal: PL097337.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 57745043
Conc: 17.65 ng/ml

Signal: PL097337.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 91948485
Conc: 16.16 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/12/25
Project:	PGW Phila. Gas Works	Date Received:	09/12/25
Client Sample ID:	291376	SDG No.:	Q3092
Lab Sample ID:	Q3092-02	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097334.D	1	09/22/25 10:05	09/22/25 19:33	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.6		57 - 171	83%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.8		61 - 148	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 19:33
 Operator : AR\AJ
 Sample : Q3092-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 291376

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.526	2.824	72730683	110.6E6	16.065m	16.823
28) SA Decachlor...	8.999	7.984	54384059	85309141	16.620	14.989

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 19:33
Operator : AR\AJ
Sample : Q3092-02
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
291376


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.526 min
Delta R.T.: 0.000 min
Response: 72730683
Conc: 16.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
291376


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 110620731
Conc: 16.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 54384059
Conc: 16.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 85309141
Conc: 14.99 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/12/25
Project:	PGW Phila. Gas Works	Date Received:	09/12/25
Client Sample ID:	279182	SDG No.:	Q3092
Lab Sample ID:	Q3092-04	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097342.D	1	09/22/25 10:05	09/22/25 21:49	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.2		57 - 171	81%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.6		61 - 148	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 21:49
Operator : AR\AJ
Sample : Q3092-04
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
279182

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.825	71777274	109.4E6	15.855m	16.643
28)	SA Decachlor...	8.999	7.983	53114770	87236577	16.232	15.327

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 21:49
Operator : AR\AJ
Sample : Q3092-04
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
279182


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 71777274
Conc: 15.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
279182


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: 0.001 min
Response: 109436586
Conc: 16.64 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 53114770
Conc: 16.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: -0.001 min
Response: 87236577
Conc: 15.33 ng/ml



CALIBRATION SUMMARY

Calibration Times: **17:06** **18:00**

LAB FILE ID:	RT 100 =	<u>PL097271.D</u>	RT 075 =	<u>PL097272.D</u>
	RT 050 =	PL097273.D	RT 025 =	PL097274.D
			RT 005 =	PL097275.D

[illegible]

Calibration Times: **17:06** **18:00**

LAB FILE ID:	RT 100 =	<u>PL097271.D</u>	RT 075 =	<u>PL097272.D</u>
	RT 050 =	PL097273.D	RT 025 =	PL097274.D
			RT 005 =	PL097275.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Instrument ID: ECD_L

Calibration Date(s): 09/18/2025 09/18/2025

Calibration Times: 17:06 18:00

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>		
CF 050 =		<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	3151670000	3121880000	3187620000	3242730000	3657150000	3272210000	7
Endrin	4777540000	4640970000	4618310000	4356250000	4672620000	4613140000	3
gamma-BHC (Lindane)	6940420000	6699850000	6593990000	6320490000	6510990000	6613150000	3
Heptachlor	6809100000	6621120000	6608040000	6426970000	6885270000	6670100000	3
Heptachlor epoxide	5764160000	5655320000	5756800000	5829540000	5728960000	5746960000	1
Methoxychlor	2052960000	2027370000	2062620000	2055200000	2228740000	2085380000	4
Tetrachloro-m-xylene	4554070000	4459430000	4465020000	4402740000	4754560000	4527160000	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Instrument ID: ECD_L

Calibration Date(s): 09/18/2025 09/18/2025

Calibration Times: 17:06 18:00

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:								CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>	
CF 050 =								<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD					
Decachlorobiphenyl	5360550000	5307930000	5459740000	5623180000	6706450000	5691570000	10					
Endrin	7450260000	7335750000	7338920000	7289700000	8137930000	7510510000	5					
gamma-BHC (Lindane)	9632950000	9373760000	9300450000	8991300000	9274840000	9314660000	2					
Heptachlor	9299100000	9162550000	9216060000	9078210000	9769790000	9305140000	3					
Heptachlor epoxide	7825700000	7749420000	7831840000	7763530000	8487220000	7931540000	4					
Methoxychlor	3580590000	3582010000	3698660000	3768400000	4217720000	3769480000	7					
Tetrachloro-m-xylene	6546800000	6457210000	6449880000	6384190000	7039690000	6575560000	4					



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ATLA10

Lab Code: ACE SDG NO.: Q3092

Instrument ID: ECD_L Date(s) Analyzed: 09/18/2025 09/18/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.68	4.58	4.78	206851000
		2	5.20	5.10	5.30	216096000
		3	5.91	5.81	6.01	813754000
		4	5.99	5.89	6.09	992045000
		5	6.83	6.73	6.93	162879000
Toxaphene	500	1	6.20	6.10	6.30	42075500
		2	6.60	6.50	6.70	33847100
		3	7.01	6.91	7.11	141900000
		4	7.11	7.01	7.21	105792000
		5	7.88	7.78	7.98	74166200



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
Instrument ID: ECD_L **Date(s) Analyzed:** 09/18/2025 09/18/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.83	3.73	3.93	252157000
		2	4.41	4.31	4.51	284305000
		3	5.04	4.94	5.14	877635000
		4	5.11	5.01	5.21	766472000
		5	6.00	5.90	6.10	292534000
Toxaphene	500	1	5.07	4.97	5.17	42482700
		2	5.75	5.65	5.85	54918500
		3	6.03	5.93	6.13	57955900
		4	6.67	6.57	6.77	177984000
		5	7.11	7.01	7.21	109140000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097271.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:06
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.823	455.4E6	654.7E6	101.994	101.503
2)	SA Decachlor...	9.001	7.984	315.2E6	536.1E6	98.872	98.183
Target Compounds							
2)	A alpha-BHC	3.974	3.329	743.3E6	1044.5E6	105.888	104.303
3)	MA gamma-BHC...	4.302	3.660	694.0E6	963.3E6	105.254	103.575
4)	MA Heptachlor	4.893	4.008	680.9E6	929.9E6	103.043	100.901
5)	MB Aldrin	5.232	4.290	655.2E6	886.1E6	104.089	102.301
6)	B beta-BHC	4.489	3.956	269.2E6	397.0E6	100.455	100.557
7)	B delta-BHC	4.735	4.189	640.9E6	958.6E6	104.552	103.832
8)	B Heptachlo...	5.652	4.793	576.4E6	782.6E6	100.128	99.922
9)	A Endosulfan I	6.033	5.163	524.9E6	738.7E6	102.490	99.419
10)	B gamma-Chl...	5.905	5.044	579.3E6	855.4E6	103.252	99.860
11)	B alpha-Chl...	5.986	5.108	572.8E6	827.5E6	102.408	100.361
12)	B 4,4'-DDE	6.156	5.297	496.4E6	791.1E6	104.790	102.753
13)	MA Dieldrin	6.306	5.428	566.8E6	825.1E6	103.988	101.465
14)	MA Endrin	6.532	5.702	477.8E6	745.0E6	103.448	101.517
15)	B Endosulfa...	6.745	5.994	469.5E6	697.8E6	98.323	100.299
16)	A 4,4'-DDD	6.665	5.850	389.1E6	658.9E6	103.408	103.640
17)	MA 4,4'-DDT	6.979	6.102	440.3E6	719.0E6	103.710	103.115
18)	B Endrin al...	6.874	6.172	303.6E6	476.2E6	99.226	96.896
19)	B Endosulfa...	7.108	6.396	408.4E6	666.5E6	101.600	99.231
20)	A Methoxychlor	7.452	6.674	205.3E6	358.1E6	99.532	96.808
21)	B Endrin ke...	7.587	6.900	429.6E6	722.3E6	101.695	97.725
22)	Mirex	8.065	7.089	330.8E6	552.6E6	97.189	96.302



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:06
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:19
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	334.5E6	484.3E6	74.906	75.085
2)	SA Decachlor...	9.002	7.985	234.1E6	398.1E6	73.453	72.915
Target Compounds							
2)	A alpha-BHC	3.975	3.329	536.3E6	757.5E6	76.392	75.640
3)	MA gamma-BHC...	4.302	3.661	502.5E6	703.0E6	76.204	75.591
4)	MA Heptachlor	4.893	4.009	496.6E6	687.2E6	75.148	74.565
5)	MB Aldrin	5.232	4.291	477.7E6	650.5E6	75.880	75.100
6)	B beta-BHC	4.490	3.957	198.1E6	292.5E6	73.934	74.080
7)	B delta-BHC	4.735	4.190	463.0E6	697.0E6	75.527	75.496
8)	B Heptachlo...	5.652	4.793	424.1E6	581.2E6	73.678	74.211
9)	A Endosulfan I	6.034	5.163	383.9E6	550.9E6	74.971	74.138
10)	B gamma-Chl...	5.906	5.045	423.9E6	636.8E6	75.554	74.339
11)	B alpha-Chl...	5.986	5.109	419.8E6	614.2E6	75.053	74.493
12)	B 4,4'-DDE	6.156	5.297	361.5E6	581.0E6	76.313	75.459
13)	MA Dieldrin	6.306	5.428	412.9E6	609.0E6	75.745	74.885
14)	MA Endrin	6.533	5.703	348.1E6	550.2E6	75.368	74.968
15)	B Endosulfa...	6.745	5.995	347.8E6	518.4E6	72.847	74.509
16)	A 4,4'-DDD	6.665	5.850	284.1E6	483.2E6	75.512	76.000
17)	MA 4,4'-DDT	6.980	6.102	320.6E6	526.7E6	75.521	75.530
18)	B Endrin al...	6.875	6.173	224.8E6	356.8E6	73.496	72.597
19)	B Endosulfa...	7.108	6.396	299.3E6	496.8E6	74.448	73.965
20)	A Methoxychlor	7.453	6.674	152.1E6	268.7E6	73.718	72.634
21)	B Endrin ke...	7.588	6.900	314.3E6	536.5E6	74.403	72.588
22)	Mirex	8.066	7.090	246.6E6	415.4E6	72.461	72.394



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:19
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:33
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	223.3E6	322.5E6	50.000	50.000
28)	SA Decachlor...	9.002	7.984	159.4E6	273.0E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.975	3.329	351.0E6	500.7E6	50.000	50.000
3)	MA gamma-BHC...	4.302	3.661	329.7E6	465.0E6	50.000	50.000
4)	MA Heptachlor	4.893	4.009	330.4E6	460.8E6	50.000	50.000
5)	MB Aldrin	5.232	4.291	314.7E6	433.1E6	50.000	50.000
6)	B beta-BHC	4.490	3.957	134.0E6	197.4E6	50.000	50.000
7)	B delta-BHC	4.735	4.190	306.5E6	461.6E6	50.000	50.000
8)	B Heptachlo...	5.652	4.793	287.8E6	391.6E6	50.000	50.000
9)	A Endosulfan I	6.034	5.164	256.1E6	371.5E6	50.000	50.000
10)	B gamma-Chl...	5.906	5.045	280.5E6	428.3E6	50.000	50.000
11)	B alpha-Chl...	5.986	5.109	279.7E6	412.3E6	50.000	50.000
12)	B 4,4'-DDE	6.156	5.297	236.9E6	385.0E6	50.000	50.000
13)	MA Dieldrin	6.306	5.428	272.5E6	406.6E6	50.000	50.000
14)	MA Endrin	6.532	5.703	230.9E6	366.9E6	50.000	50.000
15)	B Endosulfa...	6.746	5.995	238.7E6	347.8E6	50.000	50.000
16)	A 4,4'-DDD	6.666	5.850	188.1E6	317.9E6	50.000	50.000
17)	MA 4,4'-DDT	6.980	6.102	212.3E6	348.6E6	50.000	50.000
18)	B Endrin al...	6.874	6.173	153.0E6	245.7E6	50.000	50.000
19)	B Endosulfa...	7.107	6.396	201.0E6	335.8E6	50.000	50.000
20)	A Methoxychlor	7.452	6.674	103.1E6	184.9E6	50.000	50.000
21)	B Endrin ke...	7.587	6.900	211.2E6	369.6E6	50.000	50.000
22)	Mirex	8.065	7.090	170.2E6	286.9E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:33
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:47
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:45:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	110.1E6	159.6E6	24.651	24.745
28)	SA Decachlor...	9.001	7.984	81068137	140.6E6	25.432	25.748
Target Compounds							
2)	A alpha-BHC	3.974	3.329	165.9E6	238.9E6	23.631	23.856
3)	MA gamma-BHC...	4.302	3.661	158.0E6	224.8E6	23.963	24.169
4)	MA Heptachlor	4.893	4.009	160.7E6	227.0E6	24.315	24.626
5)	MB Aldrin	5.232	4.291	151.8E6	210.1E6	24.117	24.257
6)	B beta-BHC	4.489	3.957	66839868	99596935	24.946	25.226
7)	B delta-BHC	4.735	4.190	146.9E6	220.9E6	23.960	23.922
8)	B Heptachlo...	5.652	4.793	145.7E6	194.1E6	25.316	24.782
9)	A Endosulfan I	6.034	5.164	125.0E6	186.4E6	24.417	25.081
10)	B gamma-Chl...	5.906	5.045	136.6E6	215.6E6	24.338	25.170
11)	B alpha-Chl...	5.986	5.109	137.3E6	206.1E6	24.543	24.992
12)	B 4,4'-DDE	6.156	5.297	112.8E6	185.2E6	23.816	24.048
13)	MA Dieldrin	6.306	5.428	131.4E6	197.7E6	24.115	24.315
14)	MA Endrin	6.532	5.702	108.9E6	182.2E6	23.581	24.832
15)	B Endosulfa...	6.746	5.995	118.2E6	173.5E6	24.764	24.933
16)	A 4,4'-DDD	6.665	5.850	89516591	156.6E6	23.790	24.632
17)	MA 4,4'-DDT	6.979	6.102	101.6E6	167.4E6	23.933	24.009
18)	B Endrin al...	6.875	6.173	75911729	128.9E6	24.814	26.230
19)	B Endosulfa...	7.108	6.396	98950148	168.3E6	24.614	25.062
20)	A Methoxychlor	7.453	6.675	51380116	94210076	24.910	25.471
21)	B Endrin ke...	7.587	6.900	102.9E6	184.8E6	24.367	25.000
22)	Mirex	8.065	7.090	87774467	148.3E6	25.790	25.847



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:47
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:45:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097275.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 18:00
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/19/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:45:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.824	23772791	35198466	5.324	5.457
2) SA Decachlor...	9.001	7.984	18285769	33532264	5.737	6.142
Target Compounds						
2) A alpha-BHC	3.974	3.328	33487008	48361581	4.770	4.829
3) MA gamma-BHC...	4.302	3.660	32554930	46374195	4.937	4.986
4) MA Heptachlor	4.893	4.008	34426346	48848941	5.210	5.300
5) MB Aldrin	5.232	4.290	31901901	43982840	5.068	5.078
6) B beta-BHC	4.490	3.957	14725134	22540181	5.496	5.709
7) B delta-BHC	4.733	4.189	28774806	45070449	4.694m	4.882
8) B Heptachlo...	5.650	4.793	28644801	42436094	4.976m	5.418
9) A Endosulfan I	6.033	5.163	27081618	44002825	5.288	5.922
10) B gamma-Chl...	5.905	5.043	28749631	47175597	5.124	5.507m
11) B alpha-Chl...	5.986	5.107	29943592	42753742	5.353	5.185m
12) B 4,4'-DDE	6.156	5.297	23948455	38300213	5.056	4.974
13) MA Dieldrin	6.305	5.428	27648945	42357479	5.073	5.209
14) MA Endrin	6.532	5.702	23363118	40689632	5.059	5.544
15) B Endosulfa...	6.744	5.994	28776249	38809493	6.027m	5.579
16) A 4,4'-DDD	6.665	5.850	19149707	32954369	5.089	5.183
17) MA 4,4'-DDT	6.979	6.102	20744145	33603734	4.886	4.819
18) B Endrin al...	6.874	6.171	16425107	35487549	5.369	7.221m#
19) B Endosulfa...	7.108	6.396	21803103	37554894	5.424	5.592
20) A Methoxychlor	7.452	6.674	11143676	21088586	5.403	5.702
21) B Endrin ke...	7.587	6.899	21956995	42459706	5.197	5.745m
22) Mirex	8.065	7.090	21205020	35794871	6.230	6.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 18:00
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/19/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:45:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097278.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 18:41
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:25:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:24:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.823	181.2E6	350.3E6	50.000	50.000
28)	SA Decachlor...	9.002	7.984	130.8E6	233.3E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.680	3.832	103.4E6	126.1E6	500.000	500.000
24)	Chlordane-2	5.204	4.412	108.0E6	142.2E6	500.000	500.000
25)	Chlordane-3	5.907	5.044	406.9E6	438.8E6	500.000	500.000
26)	Chlordane-4	5.991	5.108	496.0E6	383.2E6	500.000	500.000
27)	Chlordane-5	6.828	6.002	81439283	146.3E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 18:41
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:25:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:24:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 19:49
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 06:18:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 06:17:16 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.824	212.8E6	316.1E6	50.000	50.000
7) SA Decachlor...	9.001	7.985	151.2E6	274.7E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.202	5.066	21037740	21241348	500.000	500.000
3) Toxaphene-2	6.601	5.752	16923557	27459266	500.000	500.000
4) Toxaphene-3	7.014	6.033	70949766	28977953	500.000	500.000
5) Toxaphene-4	7.105	6.667	52895758	88991768	500.000	500.000
6) Toxaphene-5	7.883	7.107	37083116	54569782	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 19:49
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 06:18:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 06:17:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097286.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 08:39
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:03:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.820	233.1E6	328.4E6	51.482	49.938
28)	SA Decachlor...	9.006	7.985	165.2E6	255.3E6	50.482	44.851
Target Compounds							
2)	A alpha-BHC	3.974	3.325	368.4E6	500.7E6	52.723	50.281
3)	MA gamma-BHC...	4.302	3.657	344.6E6	461.1E6	52.110	49.498
4)	MA Heptachlor	4.893	4.005	349.0E6	445.3E6	52.326	47.851
5)	MB Aldrin	5.233	4.288	325.9E6	418.1E6	51.460	48.168
6)	B beta-BHC	4.490	3.954	136.9E6	187.9E6	50.230	46.246
7)	B delta-BHC	4.734	4.186	321.2E6	449.4E6	52.930m	48.893
8)	B Heptachlo...	5.651	4.789	306.1E6	376.2E6	53.263m	47.430m
9)	A Endosulfan I	6.034	5.161	264.5E6	345.6E6	51.051m	44.982
10)	B gamma-Chl...	5.905	5.042	288.6E6	406.9E6	51.050m	46.594
11)	B alpha-Chl...	5.986	5.107	285.1E6	401.4E6	50.192m	48.364
12)	B 4,4'-DDE	6.156	5.294	248.9E6	367.5E6	52.230m	47.821m
13)	MA Dieldrin	6.307	5.424	282.4E6	392.3E6	51.517	47.972m
14)	MA Endrin	6.532	5.701	241.9E6	351.7E6	52.445m	46.833
15)	B Endosulfa...	6.745	5.993	240.2E6	329.4E6	48.828m	46.333
16)	A 4,4'-DDD	6.666	5.848	191.1E6	297.9E6	50.687m	46.193
17)	MA 4,4'-DDT	6.982	6.101	229.3E6	328.0E6	54.236	47.389
18)	B Endrin al...	6.876	6.171	175.3E6	229.7E6	56.862	43.033
19)	B Endosulfa...	7.110	6.395	209.3E6	310.2E6	51.264	45.293
20)	A Methoxychlor	7.455	6.673	103.8E6	168.0E6	49.793	44.559
21)	B Endrin ke...	7.590	6.898	220.0E6	395.4E6	51.828	52.514m
22)	Mirex	8.068	7.089	174.4E6	269.2E6	49.109	45.026

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097286.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 08:39
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

ICVPL092025

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/22/2025

Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:03:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097287.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 08:52
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:12:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.822	199.7E6	365.8E6	44.121	55.633 #
28)	SA Decachlor...	9.001	7.983	139.6E6	237.0E6	42.668	41.636
Target Compounds							
23)	Chlordane-1	4.679	3.831	111.7E6	129.9E6	527.132	511.091
24)	Chlordane-2	5.201	4.410	111.1E6	150.3E6	500.634m	511.589
25)	Chlordane-3	5.905	5.043	424.6E6	455.7E6	520.618m	509.555
26)	Chlordane-4	5.989	5.106	515.0E6	380.7E6	514.915m	499.841m
27)	Chlordane-5	6.826	6.001	87854088	149.8E6	524.084m	502.416

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097287.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 08:52
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
ICVPL092025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/22/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 09:12:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 09:25
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:41:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 06:34:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.822	206.8E6	305.4E6	49.043	49.125
7) SA Decachlor...	9.004	7.985	146.0E6	259.8E6	48.901	47.512
Target Compounds						
2) Toxaphene-1	6.203	5.064	20584370	20730281	478.691	527.108m
3) Toxaphene-2	6.602	5.752	15850196	25914739	475.278	475.930
4) Toxaphene-3	7.016	6.033	66126700	27574610	472.119	485.585
5) Toxaphene-4	7.106	6.667	50737974	89589604	488.063	520.666
6) Toxaphene-5	7.885	7.107	34706442	51873806	481.757	482.163

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 09:25
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
ICVPL092025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/22/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 09:41:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 06:34:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 18:11

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 18:11

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.001	8.902	9.102	48.730	50.000	-2.5
Endrin	6.531	6.432	6.632	46.480	50.000	-7.0
gamma-BHC (Lindane)	4.302	4.202	4.402	51.410	50.000	2.8
Heptachlor	4.892	4.793	4.993	49.840	50.000	-0.3
Heptachlor epoxide	5.651	5.552	5.752	50.370	50.000	0.7
Methoxychlor	7.452	7.352	7.552	44.040	50.000	-11.9
Tetrachloro-m-xylene	3.527	3.428	3.628	51.250	50.000	2.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.884	8.084	40.500	50.000	-19.0
Endrin	5.702	5.603	5.803	44.550	50.000	-10.9
gamma-BHC (Lindane)	3.661	3.561	3.761	49.630	50.000	-0.7
Heptachlor	4.009	3.909	4.109	48.080	50.000	-3.8
Heptachlor epoxide	4.792	4.693	4.893	48.930	50.000	-2.1
Methoxychlor	6.674	6.574	6.774	41.040	50.000	-17.9
Tetrachloro-m-xylene	2.824	2.724	2.924	49.030	50.000	-1.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097328.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 18:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:47:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	232.0E6	322.4E6	51.248	49.032
28)	SA Decachlor...	9.001	7.984	159.5E6	230.5E6	48.729	40.496
Target Compounds							
2)	A alpha-BHC	3.974	3.329	362.2E6	499.4E6	51.831	50.157
3)	MA gamma-BHC...	4.302	3.661	340.0E6	462.3E6	51.413	49.634
4)	MA Heptachlor	4.892	4.009	332.4E6	447.4E6	49.840	48.077
5)	MB Aldrin	5.232	4.291	318.6E6	428.3E6	50.303	49.349
6)	B beta-BHC	4.489	3.957	141.1E6	196.9E6	51.762	48.472
7)	B delta-BHC	4.733	4.189	324.6E6	457.6E6	53.487m	49.783
8)	B Heptachlo...	5.651	4.792	289.5E6	388.1E6	50.367	48.925
9)	A Endosulfan I	6.033	5.163	252.6E6	363.4E6	48.750	47.298
10)	B gamma-Chl...	5.905	5.044	282.8E6	425.3E6	50.028	48.697
11)	B alpha-Chl...	5.986	5.108	277.1E6	406.0E6	48.784	48.919
12)	B 4,4'-DDE	6.156	5.297	240.2E6	372.2E6	50.406	48.428
13)	MA Dieldrin	6.305	5.427	267.8E6	395.0E6	48.858	48.307
14)	MA Endrin	6.531	5.702	214.4E6	334.6E6	46.481	44.546
15)	B Endosulfa...	6.745	5.994	230.4E6	328.0E6	46.849	46.131
16)	A 4,4'-DDD	6.664	5.849	196.3E6	311.7E6	52.066	48.328
17)	MA 4,4'-DDT	6.979	6.102	192.9E6	303.4E6	45.633	43.844
18)	B Endrin al...	6.874	6.172	149.5E6	235.9E6	48.491	44.193
19)	B Endosulfa...	7.107	6.396	196.2E6	311.1E6	48.060	45.418
20)	A Methoxychlor	7.452	6.674	91835427	154.7E6	44.038	41.036
21)	B Endrin ke...	7.587	6.899	207.6E6	333.6E6	48.909	44.299
22)	Mirex	8.065	7.089	158.3E6	248.2E6	44.594	41.509

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097328.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 18:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 21:08

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 21:08

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.902	9.102	53.620	50.000	7.2
Endrin	6.531	6.432	6.632	50.770	50.000	1.5
gamma-BHC (Lindane)	4.302	4.202	4.402	55.280	50.000	10.6
Heptachlor	4.893	4.793	4.993	53.040	50.000	6.1
Heptachlor epoxide	5.651	5.552	5.752	54.390	50.000	8.8
Methoxychlor	7.451	7.352	7.552	48.210	50.000	-3.6
Tetrachloro-m-xylene	3.527	3.428	3.628	54.690	50.000	9.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.470	50.000	-3.1
Endrin	5.702	5.603	5.803	50.690	50.000	1.4
gamma-BHC (Lindane)	3.661	3.561	3.761	54.800	50.000	9.6
Heptachlor	4.009	3.909	4.109	52.810	50.000	5.6
Heptachlor epoxide	4.792	4.693	4.893	53.620	50.000	7.2
Methoxychlor	6.673	6.574	6.774	46.500	50.000	-7.0
Tetrachloro-m-xylene	2.824	2.724	2.924	53.880	50.000	7.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 21:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:49:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	247.6E6	354.3E6	54.690	53.880
2)	SA Decachlor...	8.999	7.983	175.5E6	275.9E6	53.622	48.471
Target Compounds							
2)	A alpha-BHC	3.974	3.329	389.8E6	551.3E6	55.786	55.367
3)	MA gamma-BHC...	4.302	3.661	365.6E6	510.4E6	55.285	54.799
4)	MA Heptachlor	4.893	4.009	353.8E6	491.4E6	53.038	52.810
5)	MB Aldrin	5.232	4.291	344.0E6	472.6E6	54.317	54.449
6)	B beta-BHC	4.489	3.957	150.7E6	215.8E6	55.262	53.136
7)	B delta-BHC	4.733	4.190	345.8E6	503.1E6	56.983m	54.736
8)	B Heptachlo...	5.651	4.792	312.6E6	425.3E6	54.386	53.618
9)	A Endosulfan I	6.033	5.163	279.5E6	400.2E6	53.953	52.087
10)	B gamma-Chl...	5.905	5.044	310.6E6	467.2E6	54.934	53.497
11)	B alpha-Chl...	5.985	5.108	305.6E6	447.6E6	53.810	53.929
12)	B 4,4'-DDE	6.155	5.295	267.1E6	415.5E6	56.064	54.073m
13)	MA Dieldrin	6.305	5.428	294.7E6	441.9E6	53.759	54.044
14)	MA Endrin	6.531	5.702	234.2E6	380.7E6	50.768	50.690
15)	B Endosulfa...	6.744	5.994	252.8E6	372.7E6	51.394	52.423
16)	A 4,4'-DDD	6.664	5.849	216.4E6	358.1E6	57.382	55.533
17)	MA 4,4'-DDT	6.978	6.101	211.6E6	342.2E6	50.061	49.448
18)	B Endrin al...	6.873	6.172	166.4E6	267.4E6	53.991	50.099
19)	B Endosulfa...	7.106	6.395	216.7E6	352.3E6	53.088	51.427
20)	A Methoxychlor	7.451	6.673	100.5E6	175.3E6	48.210	46.501
21)	B Endrin ke...	7.586	6.899	232.0E6	404.6E6	54.663	53.724
22)	Mirex	8.064	7.089	175.9E6	296.9E6	49.552	49.664

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 21:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 00:46

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY**Lab Name:** Alliance**Contract:** ATLA10**Lab Code:** ACE**SDG NO.:** Q3092**Continuing Calib Date:** 09/23/2025**Initial Calibration Date(s):** 09/18/202509/18/2025**Continuing Calib Time:** 00:46**Initial Calibration Time(s):** 17:0618:00**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097355.D **Time Analyzed:** 00:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.902	9.102	51.350	50.000	2.7
Endrin	6.531	6.432	6.632	48.870	50.000	-2.3
gamma-BHC (Lindane)	4.301	4.202	4.402	54.300	50.000	8.6
Heptachlor	4.892	4.793	4.993	51.040	50.000	2.1
Heptachlor epoxide	5.651	5.552	5.752	52.630	50.000	5.3
Methoxychlor	7.451	7.352	7.552	44.570	50.000	-10.9
Tetrachloro-m-xylene	3.527	3.428	3.628	54.430	50.000	8.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097355.D **Time Analyzed:** 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	45.360	50.000	-9.3
Endrin	5.702	5.603	5.803	47.950	50.000	-4.1
gamma-BHC (Lindane)	3.661	3.561	3.761	54.250	50.000	8.5
Heptachlor	4.008	3.909	4.109	51.270	50.000	2.5
Heptachlor epoxide	4.792	4.693	4.893	52.550	50.000	5.1
Methoxychlor	6.673	6.574	6.774	41.840	50.000	-16.3
Tetrachloro-m-xylene	2.824	2.724	2.924	53.670	50.000	7.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097355.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 00:46
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:51:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	246.4E6	352.9E6	54.430	53.667
28)	SA Decachlor...	8.999	7.983	168.0E6	258.2E6	51.345	45.360
Target Compounds							
2)	A alpha-BHC	3.974	3.329	385.7E6	547.8E6	55.205	55.011
3)	MA gamma-BHC...	4.301	3.661	359.1E6	505.4E6	54.296	54.255
4)	MA Heptachlor	4.892	4.008	340.5E6	477.0E6	51.042	51.266
5)	MB Aldrin	5.232	4.290	335.1E6	465.2E6	52.912	53.598
6)	B beta-BHC	4.489	3.957	148.2E6	213.5E6	54.367	52.550
7)	B delta-BHC	4.733	4.189	338.2E6	494.7E6	55.740m	53.812
8)	B Heptachlo...	5.651	4.792	302.5E6	416.8E6	52.630	52.555
9)	A Endosulfan I	6.032	5.163	269.4E6	386.9E6	52.002	50.349
10)	B gamma-Chl...	5.905	5.044	300.6E6	453.9E6	53.175	51.965
11)	B alpha-Chl...	5.985	5.108	295.1E6	433.6E6	51.958	52.232
12)	B 4,4'-DDE	6.156	5.297	258.7E6	399.5E6	54.289	51.982
13)	MA Dieldrin	6.305	5.428	286.3E6	424.7E6	52.216	51.938
14)	MA Endrin	6.531	5.702	225.4E6	360.1E6	48.869	47.953
15)	B Endosulfa...	6.744	5.994	244.9E6	354.6E6	49.800	49.877
16)	A 4,4'-DDD	6.664	5.849	214.9E6	344.9E6	56.990	53.475
17)	MA 4,4'-DDT	6.978	6.101	195.4E6	306.4E6	46.235	44.275
18)	B Endrin al...	6.873	6.172	160.8E6	247.7E6	52.174	46.416
19)	B Endosulfa...	7.107	6.395	210.4E6	332.9E6	51.530	48.603
20)	A Methoxychlor	7.451	6.673	92934830	157.7E6	44.565	41.836
21)	B Endrin ke...	7.586	6.899	223.1E6	366.2E6	52.564	48.627
22)	Mirex	8.064	7.088	168.3E6	274.7E6	47.413	45.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097355.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 00:46
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:51:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 10:03

Initial Calibration Time(s): 17:06 18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 10:03

Initial Calibration Time(s): 17:06 18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.003	8.902	9.102	53.250	50.000	6.5
Endrin	6.533	6.432	6.632	49.330	50.000	-1.3
gamma-BHC (Lindane)	4.302	4.202	4.402	54.410	50.000	8.8
Heptachlor	4.893	4.793	4.993	50.970	50.000	1.9
Heptachlor epoxide	5.652	5.552	5.752	51.910	50.000	3.8
Methoxychlor	7.453	7.352	7.552	46.240	50.000	-7.5
Tetrachloro-m-xylene	3.527	3.428	3.628	52.820	50.000	5.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.884	8.084	49.070	50.000	-1.9
Endrin	5.702	5.603	5.803	47.520	50.000	-5.0
gamma-BHC (Lindane)	3.659	3.561	3.761	52.620	50.000	5.2
Heptachlor	4.007	3.909	4.109	50.390	50.000	0.8
Heptachlor epoxide	4.791	4.693	4.893	51.290	50.000	2.6
Methoxychlor	6.674	6.574	6.774	43.870	50.000	-12.3
Tetrachloro-m-xylene	2.822	2.724	2.924	52.390	50.000	4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097359.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 10:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.822	239.1E6	344.5E6	52.825	52.394
2)	SA Decachlor...	9.003	7.984	174.3E6	279.3E6	53.255	49.072
Target Compounds							
2)	A alpha-BHC	3.974	3.328	376.7E6	531.4E6	53.909	53.364
3)	MA gamma-BHC...	4.302	3.659	359.9E6	490.1E6	54.415	52.618
4)	MA Heptachlor	4.893	4.007	340.0E6	468.9E6	50.967	50.387
5)	MB Aldrin	5.232	4.289	331.7E6	453.2E6	52.377	52.212
6)	B beta-BHC	4.489	3.956	144.1E6	205.3E6	52.852	50.545
7)	B delta-BHC	4.734	4.188	329.9E6	478.4E6	54.360m	52.039
8)	B Heptachlo...	5.652	4.791	298.4E6	406.8E6	51.915	51.294
9)	A Endosulfan I	6.034	5.162	269.5E6	373.9E6	52.005	48.664
10)	B gamma-Chl...	5.906	5.043	297.1E6	428.9E6	52.544	49.103
11)	B alpha-Chl...	5.986	5.108	294.1E6	418.6E6	51.782	50.427
12)	B 4,4'-DDE	6.157	5.296	254.4E6	397.9E6	53.395	51.775
13)	MA Dieldrin	6.306	5.427	286.3E6	424.3E6	52.223	51.894
14)	MA Endrin	6.533	5.702	227.6E6	356.9E6	49.327	47.520
15)	B Endosulfa...	6.746	5.994	248.8E6	357.6E6	50.576	50.307
16)	A 4,4'-DDD	6.666	5.849	222.0E6	345.0E6	58.889	53.501
17)	MA 4,4'-DDT	6.981	6.102	201.6E6	324.2E6	47.680	46.843
18)	B Endrin al...	6.875	6.172	163.7E6	261.0E6	53.109	48.893
19)	B Endosulfa...	7.109	6.396	212.2E6	338.0E6	51.969	49.339
20)	A Methoxychlor	7.453	6.674	96417789	165.4E6	46.235	43.867
21)	B Endrin ke...	7.589	6.900	229.8E6	386.0E6	54.142	51.254
22)	Mirex	8.066	7.089	172.1E6	289.3E6	48.470	48.388



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097359.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:03
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 12:52

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 12:52

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.004	8.902	9.102	52.350	50.000	4.7
Endrin	6.533	6.432	6.632	49.760	50.000	-0.5
gamma-BHC (Lindane)	4.301	4.202	4.402	53.420	50.000	6.8
Heptachlor	4.892	4.793	4.993	50.770	50.000	1.5
Heptachlor epoxide	5.652	5.552	5.752	51.580	50.000	3.2
Methoxychlor	7.454	7.352	7.552	45.610	50.000	-8.8
Tetrachloro-m-xylene	3.526	3.428	3.628	53.100	50.000	6.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATLA10
Lab Code: ACE **SDG NO.:** Q3092
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.450	50.000	-3.1
Endrin	5.700	5.603	5.803	47.320	50.000	-5.4
gamma-BHC (Lindane)	3.657	3.561	3.761	52.520	50.000	5.0
Heptachlor	4.005	3.909	4.109	49.830	50.000	-0.3
Heptachlor epoxide	4.790	4.693	4.893	51.900	50.000	3.8
Methoxychlor	6.673	6.574	6.774	43.640	50.000	-12.7
Tetrachloro-m-xylene	2.820	2.724	2.924	52.450	50.000	4.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 12:52
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.820	240.4E6	344.9E6	53.103	52.451
2)	SA Decachlor...	9.004	7.983	171.3E6	275.8E6	52.353	48.452
Target Compounds							
2)	A alpha-BHC	3.973	3.325	378.2E6	528.3E6	54.130	53.056
3)	MA gamma-BHC...	4.301	3.657	353.3E6	489.2E6	53.419	52.520
4)	MA Heptachlor	4.892	4.005	338.7E6	463.7E6	50.774	49.832
5)	MB Aldrin	5.232	4.288	330.7E6	449.2E6	52.206	51.754
6)	B beta-BHC	4.489	3.952	143.9E6	206.1E6	52.768	50.746m
7)	B delta-BHC	4.733	4.185	339.8E6	476.3E6	55.993m	51.817m
8)	B Heptachlo...	5.652	4.790	296.4E6	411.7E6	51.577	51.904
9)	A Endosulfan I	6.034	5.161	272.5E6	376.0E6	52.595	48.932
10)	B gamma-Chl...	5.906	5.042	298.6E6	431.1E6	52.816	49.364
11)	B alpha-Chl...	5.986	5.106	296.6E6	421.0E6	52.217	50.714
12)	B 4,4'-DDE	6.157	5.294	259.9E6	396.2E6	54.543	51.557
13)	MA Dieldrin	6.306	5.425	290.1E6	425.8E6	52.921	52.079
14)	MA Endrin	6.533	5.700	229.6E6	355.4E6	49.764	47.322
15)	B Endosulfa...	6.747	5.992	240.8E6	364.6E6	48.958	51.282
16)	A 4,4'-DDD	6.666	5.848	224.3E6	345.5E6	59.478	53.572
17)	MA 4,4'-DDT	6.981	6.100	199.3E6	318.0E6	47.139	45.954
18)	B Endrin al...	6.875	6.171	162.8E6	255.1E6	52.807	47.798
19)	B Endosulfa...	7.109	6.394	215.0E6	343.7E6	52.665	50.175
20)	A Methoxychlor	7.454	6.673	95109290	164.5E6	45.608	43.639
21)	B Endrin ke...	7.589	6.899	233.9E6	396.5E6	55.108	52.654
22)	Mirex	8.067	7.088	174.8E6	295.6E6	49.227	49.442



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:52
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	15.660	20.000	-21.7
Tetrachloro-m-xylene	3.527	3.480	3.580	14.770	20.000	-26.2
alpha-BHC	3.974	3.920	4.020	6.930	10.000	-30.7
beta-BHC	4.489	4.440	4.540	7.340	10.000	-26.6
gamma-BHC (Lindane)	4.302	4.250	4.350	7.210	10.000	-27.9
Endrin	6.532	6.460	6.600	37.780	50.000	-24.4
4,4'-DDT	6.979	6.910	7.050	73.290	100.000	-26.7
Methoxychlor	7.452	7.380	7.520	180.320	250.000	-27.9

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.880	8.080	15.470	20.000	-22.7
Tetrachloro-m-xylene	2.824	2.770	2.870	14.550	20.000	-27.3
alpha-BHC	3.329	3.280	3.380	6.890	10.000	-31.1
beta-BHC	3.957	3.910	4.010	7.490	10.000	-25.1
gamma-BHC (Lindane)	3.660	3.610	3.710	7.000	10.000	-30.0
Endrin	5.702	5.630	5.770	36.670	50.000	-26.7
4,4'-DDT	6.102	6.030	6.170	76.460	100.000	-23.5
Methoxychlor	6.674	6.600	6.740	166.760	250.000	-33.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097269.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 16:39
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/19/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:42:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	66867312	95692858	14.770	14.553
28)	SA Decachlor...	9.000	7.984	51238202	88066168	15.659	15.473
Target Compounds							
2)	A alpha-BHC	3.974	3.329	48421315	68644926	6.930	6.894
3)	MA gamma-BHC...	4.302	3.660	47659315	65230882	7.207	7.003
6)	B beta-BHC	4.489	3.957	20012852	30421608	7.341	7.489
14)	MA Endrin	6.532	5.702	174.3E6	275.4E6	37.781	36.671
16)	A 4,4'-DDD	6.665	5.856	465298	1405401	0.123m	0.218m#
17)	MA 4,4'-DDT	6.979	6.102	309.8E6	529.1E6	73.292	76.458
18)	B Endrin al...	6.879	6.171	1009715	9847104	0.328m	1.845m#
20)	A Methoxychlor	7.452	6.674	376.0E6	628.6E6	180.319	166.763
21)	B Endrin ke...	7.586	6.896	1666712	2086500	0.393m	0.277m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:39
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/19/2025
Supervised By : mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097316.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.999	8.900	9.100	18.680	20.000	-6.6
Tetrachloro-m-xylene	3.527	3.480	3.580	17.520	20.000	-12.4
alpha-BHC	3.974	3.920	4.020	8.550	10.000	-14.5
beta-BHC	4.489	4.440	4.540	9.390	10.000	-6.1
gamma-BHC (Lindane)	4.301	4.250	4.350	8.730	10.000	-12.7
Endrin	6.530	6.460	6.600	42.130	50.000	-15.7
4,4'-DDT	6.978	6.910	7.050	83.940	100.000	-16.1
Methoxychlor	7.451	7.380	7.520	204.110	250.000	-18.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097316.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.982	7.880	8.080	16.770	20.000	-16.2
Tetrachloro-m-xylene	2.824	2.770	2.870	17.780	20.000	-11.1
alpha-BHC	3.328	3.280	3.380	8.450	10.000	-15.5
beta-BHC	3.956	3.910	4.010	9.160	10.000	-8.4
gamma-BHC (Lindane)	3.660	3.610	3.710	8.590	10.000	-14.1
Endrin	5.701	5.630	5.770	42.060	50.000	-15.9
4,4'-DDT	6.100	6.030	6.170	85.270	100.000	-14.7
Methoxychlor	6.673	6.600	6.740	176.620	250.000	-29.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097316.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 11:20
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:45:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	79301035	116.9E6	17.517	17.780
28)	SA Decachlor...	8.999	7.982	61113780	95442246	18.677	16.769
Target Compounds							
2)	A alpha-BHC	3.974	3.328	59728278	84100216	8.548	8.446
3)	MA gamma-BHC...	4.301	3.660	57729518	80052449	8.730	8.594
6)	B beta-BHC	4.489	3.956	25610468	37189948	9.394	9.155
14)	MA Endrin	6.530	5.701	194.4E6	315.9E6	42.131	42.059
16)	A 4,4'-DDD	6.664	5.849	9359056	15979610	2.482	2.478
17)	MA 4,4'-DDT	6.978	6.100	354.8E6	590.1E6	83.938	85.274
18)	B Endrin al...	6.874	6.171	2149412	14279473	0.697m	2.675 #
20)	A Methoxychlor	7.451	6.673	425.7E6	665.8E6	204.115	176.621
21)	B Endrin ke...	7.585	6.898	6376897	13364872	1.502	1.775

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097316.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 11:20
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/23/2025
Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:45:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097340.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:55

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	18.730	20.000	-6.4
Tetrachloro-m-xylene	3.528	3.480	3.580	18.350	20.000	-8.3
alpha-BHC	3.974	3.920	4.020	8.690	10.000	-13.1
beta-BHC	4.490	4.440	4.540	9.410	10.000	-5.9
gamma-BHC (Lindane)	4.302	4.250	4.350	8.850	10.000	-11.5
Endrin	6.531	6.460	6.600	42.820	50.000	-14.4
4,4'-DDT	6.979	6.910	7.050	81.390	100.000	-18.6
Methoxychlor	7.452	7.380	7.520	197.170	250.000	-21.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097340.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:55

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.880	8.080	17.340	20.000	-13.3
Tetrachloro-m-xylene	2.824	2.770	2.870	18.060	20.000	-9.7
alpha-BHC	3.329	3.280	3.380	8.600	10.000	-14.0
beta-BHC	3.957	3.910	4.010	9.230	10.000	-7.7
gamma-BHC (Lindane)	3.661	3.610	3.710	8.720	10.000	-12.8
Endrin	5.702	5.630	5.770	42.230	50.000	-15.5
4,4'-DDT	6.102	6.030	6.170	82.920	100.000	-17.1
Methoxychlor	6.674	6.600	6.740	175.530	250.000	-29.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097340.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:55
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:49:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	83095318	118.7E6	18.355	18.056
28)	SA Decachlor...	9.000	7.984	61292884	98676908	18.731	17.337
Target Compounds							
2)	A alpha-BHC	3.974	3.329	60697471	85662634	8.687	8.603
3)	MA gamma-BHC...	4.302	3.661	58558042	81230119	8.855	8.721
6)	B beta-BHC	4.490	3.957	25643491	37511854	9.406	9.235
14)	MA Endrin	6.531	5.702	197.5E6	317.1E6	42.817	42.225
16)	A 4,4'-DDD	6.665	5.849	16585764	27649893	4.399	4.287
17)	MA 4,4'-DDT	6.979	6.102	344.1E6	573.8E6	81.392	82.919
18)	B Endrin al...	6.874	6.173	2701989	13823352	0.876m	2.590 #
20)	A Methoxychlor	7.452	6.674	411.2E6	661.6E6	197.174	175.527
21)	B Endrin ke...	7.586	6.899	10340978	22323757	2.436	2.965

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097340.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:55
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097358.D Date Analyzed: 09/23/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:29

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	19.890	20.000	-0.6
Tetrachloro-m-xylene	3.527	3.480	3.580	19.560	20.000	-2.2
alpha-BHC	3.974	3.920	4.020	9.120	10.000	-8.8
beta-BHC	4.489	4.440	4.540	9.560	10.000	-4.4
gamma-BHC (Lindane)	4.302	4.250	4.350	9.380	10.000	-6.2
Endrin	6.531	6.460	6.600	44.210	50.000	-11.6
4,4'-DDT	6.979	6.910	7.050	81.540	100.000	-18.5
Methoxychlor	7.452	7.380	7.520	198.490	250.000	-20.6

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097358.D Date Analyzed: 09/23/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:29

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.880	8.080	18.140	20.000	-9.3
Tetrachloro-m-xylene	2.824	2.770	2.870	18.590	20.000	-7.1
alpha-BHC	3.328	3.280	3.380	8.860	10.000	-11.4
beta-BHC	3.956	3.910	4.010	9.450	10.000	-5.5
gamma-BHC (Lindane)	3.660	3.610	3.710	8.990	10.000	-10.1
Endrin	5.702	5.630	5.770	41.890	50.000	-16.2
4,4'-DDT	6.101	6.030	6.170	80.180	100.000	-19.8
Methoxychlor	6.673	6.600	6.740	172.630	250.000	-30.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 09:29
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	88546646	122.3E6	19.559	18.594
28)	SA Decachlor...	9.000	7.983	65069354	103.2E6	19.885	18.140
Target Compounds							
2)	A alpha-BHC	3.974	3.328	63757976	88195831	9.125	8.857
3)	MA gamma-BHC...	4.302	3.660	62033294	83769301	9.380	8.993
6)	B beta-BHC	4.489	3.956	26052455	38395150	9.556	9.452
12)	B 4,4'-DDE	6.153	5.297	1821811	855419	0.382m	0.111m#
14)	MA Endrin	6.531	5.702	204.0E6	314.6E6	44.212	41.888
16)	A 4,4'-DDD	6.664	5.849	25716618	40045732	6.820	6.210
17)	MA 4,4'-DDT	6.979	6.101	344.7E6	554.9E6	81.542	80.183
18)	B Endrin al...	6.874	6.173	3317697	12840139	1.076m	2.406 #
20)	A Methoxychlor	7.452	6.673	413.9E6	650.7E6	198.487	172.634
21)	B Endrin ke...	7.586	6.899	14520295	27993583	3.421	3.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:29
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Analytical Sequence

Client: Atlantic Recovery Services, Inc.

SDG No.: Q3092

Project: PGW Phila. Gas Works

Instrument ID: ECD_L

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 09/18/2025 09/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/18/2025	16:25	PL097268.D	9.00	3.53
PEM	PEM	09/18/2025	16:39	PL097269.D	9.00	3.53
RESCHK	RESCHK	09/18/2025	16:52	PL097270.D	9.00	3.53
PSTDICCC100	PSTDICCC100	09/18/2025	17:06	PL097271.D	9.00	3.53
PSTDICCC075	PSTDICCC075	09/18/2025	17:19	PL097272.D	9.00	3.53
PSTDICCC050	PSTDICCC050	09/18/2025	17:33	PL097273.D	9.00	3.53
PSTDICCC025	PSTDICCC025	09/18/2025	17:47	PL097274.D	9.00	3.53
PSTDICCC005	PSTDICCC005	09/18/2025	18:00	PL097275.D	9.00	3.53
PCHLORICC500	PCHLORICC500	09/18/2025	18:41	PL097278.D	9.00	3.53
PTOXICC500	PTOXICC500	09/18/2025	19:49	PL097283.D	9.00	3.53
PEM	PEM	09/22/2025	11:20	PL097316.D	9.00	3.53
IBLK	IBLK	09/22/2025	17:58	PL097327.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/22/2025	18:11	PL097328.D	9.00	3.53
PB169782BL	PB169782BL	09/22/2025	18:52	PL097331.D	9.00	3.53
291376	Q3092-02	09/22/2025	19:33	PL097334.D	9.00	3.53
291376MS	Q3092-02MS	09/22/2025	19:47	PL097335.D	9.00	3.53
291376MSD	Q3092-02MSD	09/22/2025	20:00	PL097336.D	9.00	3.53
PB169677TB	PB169677TB	09/22/2025	20:14	PL097337.D	9.00	3.53
IBLK	IBLK	09/22/2025	20:41	PL097339.D	9.00	3.53
PEM	PEM	09/22/2025	20:55	PL097340.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/22/2025	21:08	PL097341.D	9.00	3.53
279182	Q3092-04	09/22/2025	21:49	PL097342.D	9.00	3.53
IBLK	IBLK	09/23/2025	00:32	PL097354.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	00:46	PL097355.D	9.00	3.53
IBLK	IBLK	09/23/2025	09:16	PL097357.D	9.00	3.53
PEM	PEM	09/23/2025	09:29	PL097358.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	10:03	PL097359.D	9.00	3.53
PB169782BS	PB169782BS	09/23/2025	11:47	PL097362.D	9.00	3.53
IBLK	IBLK	09/23/2025	12:09	PL097363.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	12:52	PL097364.D	9.00	3.53

Analytical Sequence

Client: Atlantic Recovery Services, Inc.	SDG No.: Q3092
Project: PGW Phila. Gas Works	Instrument ID: ECD_L
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 09/18/2025 09/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/18/2025	16:25	PL097268.D	7.98	2.82
PEM	PEM	09/18/2025	16:39	PL097269.D	7.98	2.82
RESCHK	RESCHK	09/18/2025	16:52	PL097270.D	7.98	2.82
PSTDICC100	PSTDICC100	09/18/2025	17:06	PL097271.D	7.98	2.82
PSTDICC075	PSTDICC075	09/18/2025	17:19	PL097272.D	7.99	2.82
PSTDICC050	PSTDICC050	09/18/2025	17:33	PL097273.D	7.98	2.82
PSTDICC025	PSTDICC025	09/18/2025	17:47	PL097274.D	7.98	2.82
PSTDICC005	PSTDICC005	09/18/2025	18:00	PL097275.D	7.98	2.82
PCHLORICC500	PCHLORICC500	09/18/2025	18:41	PL097278.D	7.98	2.82
PTOXICC500	PTOXICC500	09/18/2025	19:49	PL097283.D	7.99	2.82
PEM	PEM	09/22/2025	11:20	PL097316.D	7.98	2.82
IBLK	IBLK	09/22/2025	17:58	PL097327.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/22/2025	18:11	PL097328.D	7.98	2.82
PB169782BL	PB169782BL	09/22/2025	18:52	PL097331.D	7.98	2.83
291376	Q3092-02	09/22/2025	19:33	PL097334.D	7.98	2.82
291376MS	Q3092-02MS	09/22/2025	19:47	PL097335.D	7.98	2.83
291376MSD	Q3092-02MSD	09/22/2025	20:00	PL097336.D	7.98	2.82
PB169677TB	PB169677TB	09/22/2025	20:14	PL097337.D	7.98	2.82
IBLK	IBLK	09/22/2025	20:41	PL097339.D	7.98	2.82
PEM	PEM	09/22/2025	20:55	PL097340.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/22/2025	21:08	PL097341.D	7.98	2.82
279182	Q3092-04	09/22/2025	21:49	PL097342.D	7.98	2.83
IBLK	IBLK	09/23/2025	00:32	PL097354.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	00:46	PL097355.D	7.98	2.82
IBLK	IBLK	09/23/2025	09:16	PL097357.D	7.98	2.82
PEM	PEM	09/23/2025	09:29	PL097358.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	10:03	PL097359.D	7.98	2.82
PB169782BS	PB169782BS	09/23/2025	11:47	PL097362.D	7.98	2.82
IBLK	IBLK	09/23/2025	12:09	PL097363.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	12:52	PL097364.D	7.98	2.82

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

291376MS

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Lab Sample ID: Q3092-02MS

Date(s) Analyzed: 09/22/2025 09/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.45	7.40	7.50	4.20	4.9
	2	6.67	6.62	6.72	4.00	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	4.60	2.2
	2	3.66	3.61	3.71	4.70	
Heptachlor	1	4.89	4.84	4.94	4.50	0
	2	4.01	3.96	4.06	4.50	
Heptachlor epoxide	1	5.65	5.60	5.70	4.70	0
	2	4.79	4.74	4.84	4.70	
Endrin	1	6.53	6.48	6.58	4.40	2.2
	2	5.70	5.65	5.75	4.50	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

291376MSD

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Lab Sample ID: Q3092-02MSD

Date(s) Analyzed: 09/22/2025 09/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.45	7.40	7.50	4.20	4.9
	2	6.67	6.62	6.72	4.00	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	4.60	0
	2	3.66	3.61	3.71	4.60	
Heptachlor	1	4.89	4.84	4.94	4.50	2.2
	2	4.01	3.96	4.06	4.40	
Heptachlor epoxide	1	5.65	5.60	5.70	4.70	2.2
	2	4.79	4.74	4.84	4.60	
Endrin	1	6.53	6.48	6.58	4.50	0
	2	5.70	5.65	5.75	4.50	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169782BS

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Lab Sample ID: PB169782BS

Date(s) Analyzed: 09/23/2025 09/23/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.53	6.48	6.58	0.45	0.3
	2	5.70	5.65	5.75	0.45	
Methoxychlor	1	7.45	7.40	7.50	0.42	3
	2	6.67	6.62	6.72	0.41	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	0.51	1.8
	2	3.66	3.61	3.71	0.52	
Heptachlor	1	4.89	4.84	4.94	0.48	0.4
	2	4.01	3.96	4.06	0.49	
Heptachlor epoxide	1	5.65	5.60	5.70	0.52	1.4
	2	4.79	4.74	4.84	0.51	



QC SAMPLE DATA

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	
Project:	PGW Phila. Gas Works	Date Received:	
Client Sample ID:	PB169782BL	SDG No.:	Q3092
Lab Sample ID:	PB169782BL	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097331.D	1	09/22/25 10:05	09/22/25 18:52	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.5		57 - 171	88%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.9		61 - 148	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097331.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 18:52
 Operator : AR\AJ
 Sample : PB169782BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169782BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.825	75396086	111.2E6	16.654	16.905
2) SA Decachlor...	9.001	7.983	57319860	88056925	17.517	15.471

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 18:52
Operator : AR\AJ
Sample : PB169782BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169782BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL097331.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 75396086
Conc: 16.65 ng/ml


Instrument :
ECD_L
ClientSampleId :
PB169782BL



Time Response_ Signal: PL097331.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.825 min
Delta R.T.: 0.001 min
Response: 111158080
Conc: 16.90 ng/ml



Time Response_ Signal: PL097331.D\ECD1A.ch

#28 Decachlorobiphenyl


R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 57319860
Conc: 17.52 ng/ml



Time Response_ Signal: PL097331.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 88056925
Conc: 15.47 ng/ml



Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/18/25
Project:	PGW Phila. Gas Works	Date Received:	09/18/25
Client Sample ID:	PIBLK-PL097268.D	SDG No.:	Q3092
Lab Sample ID:	I.BLK-PL097268.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097268.D	1		09/18/25	PL092025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.3		57 - 171	87%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.0		61 - 148	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:25
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	70468568	105.0E6	15.566	15.973
28)	SA Decachlor...	9.001	7.984	56635933	97364753	17.308	17.107

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:25
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 70468568
Conc: 15.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 105031523
Conc: 15.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 56635933
Conc: 17.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 97364753
Conc: 17.11 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/22/25
Project:	PGW Phila. Gas Works	Date Received:	09/22/25
Client Sample ID:	PIBLK-PL097327.D	SDG No.:	Q3092
Lab Sample ID:	I.BLK-PL097327.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097327.D	1		09/22/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.2		57 - 171	101%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.6		61 - 148	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 17:58
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	84311640	122.3E6	18.624	18.599
28)	SA Decachlor...	9.000	7.983	66073819	98546486	20.192	17.314

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 17:58
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 84311640
Conc: 18.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 122301082
Conc: 18.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: 0.000 min
Response: 66073819
Conc: 20.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 98546486
Conc: 17.31 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.			Date Collected:	09/22/25	
Project:	PGW Phila. Gas Works			Date Received:	09/22/25	
Client Sample ID:	PIBLK-PL097339.D			SDG No.:	Q3092	
Lab Sample ID:	I.BLK-PL097339.D			Matrix:	TCLP	
Analytical Method:	8081B			% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	10000	uL
Soil Aliquot Vol:			uL	Test:	TCLP Pesticide	
Extraction Type:				Injection Volume :		
GPC Factor :	1.0		PH :			
Prep Method :	3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097339.D	1		09/22/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.0		57 - 171	105%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		61 - 148	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097339.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:41
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:49:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	82711967	124.4E6	18.270	18.913
28)	SA Decachlor...	8.999	7.983	68660624	109.3E6	20.983	19.202

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097339.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:41
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 82711967
Conc: 18.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 124360641
Conc: 18.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 68660624
Conc: 20.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 109291468
Conc: 19.20 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.			Date Collected:	09/23/25	
Project:	PGW Phila. Gas Works			Date Received:	09/23/25	
Client Sample ID:	PIBLK-PL097354.D			SDG No.:	Q3092	
Lab Sample ID:	I.BLK-PL097354.D			Matrix:	TCLP	
Analytical Method:	8081B			% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	10000	uL
Soil Aliquot Vol:			uL	Test:	TCLP Pesticide	
Extraction Type:				Injection Volume :		
GPC Factor :	1.0		PH :			
Prep Method :	3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097354.D	1		09/23/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		57 - 171	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.6		61 - 148	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097354.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 00:32
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:51:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	81964806	122.4E6	18.105	18.621
28)	SA Decachlor...	9.000	7.983	67978796	99090184	20.775	17.410

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 00:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:51:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL097354.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 81964806
Conc: 18.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




Time

Response_ Signal: PL097354.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.001 min
Response: 122444037
Conc: 18.62 ng/ml




Time

Response_ Signal: PL097354.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: -0.001 min
Response: 67978796
Conc: 20.77 ng/ml




Time

Response_ Signal: PL097354.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: -0.001 min
Response: 99090184
Conc: 17.41 ng/ml



Time

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/23/25
Project:	PGW Phila. Gas Works	Date Received:	09/23/25
Client Sample ID:	PIBLK-PL097357.D	SDG No.:	Q3092
Lab Sample ID:	I.BLK-PL097357.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097357.D	1		09/23/25	pl092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.2		57 - 171	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.7		61 - 148	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 09:16
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:25:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.823	89086643	127.1E6	19.678	19.331
28)	SA Decachlor...	8.999	7.982	72637734	114.4E6	22.198	20.091

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:16
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:25:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL097357.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 89086643
Conc: 19.68 ng/ml


Instrument :
ECD_L
ClientSampleId :
I.BLK



Time Response_ Signal: PL097357.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.823 min
Delta R.T.: 0.000 min
Response: 127111040
Conc: 19.33 ng/ml



Time Response_ Signal: PL097357.D\ECD1A.ch

#28 Decachlorobiphenyl


R.T.: 8.999 min
Delta R.T.: -0.001 min
Response: 72637734
Conc: 22.20 ng/ml



Time Response_ Signal: PL097357.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.982 min
Delta R.T.: -0.002 min
Response: 114350562
Conc: 20.09 ng/ml



Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/23/25
Project:	PGW Phila. Gas Works	Date Received:	09/23/25
Client Sample ID:	PIBLK-PL097363.D	SDG No.:	Q3092
Lab Sample ID:	I.BLK-PL097363.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097363.D	1		09/23/25	pl092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.4		57 - 171	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.6		61 - 148	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:09
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.822	85041873	129.1E6	18.785	19.632
28)	SA Decachlor...	9.001	7.983	70128233	116.0E6	21.431	20.373

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:09
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 85041873
Conc: 18.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.822 min
Delta R.T.: -0.002 min
Response: 129088691
Conc: 19.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 70128233
Conc: 21.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 115954509
Conc: 20.37 ng/ml

Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	
Project:	PGW Phila. Gas Works	Date Received:	
Client Sample ID:	PB169782BS	SDG No.:	Q3092
Lab Sample ID:	PB169782BS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097362.D	1	09/22/25 10:05	09/23/25 11:47	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.52		0.0037	0.050	ug/L
76-44-8	Heptachlor	0.49		0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.52		0.0096	0.050	ug/L
72-20-8	Endrin	0.45		0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.42		0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.1		57 - 171	91%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.4		61 - 148	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:47
 Operator : AR\AJ
 Sample : PB169782BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169782BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.823	77023023	114.6E6	17.014	17.436
2)	SA Decachlor...	8.999	7.982	59276355	98829039	18.115	17.364
Target Compounds							
2)	A alpha-BHC	3.973	3.328	361.5E6	522.2E6	51.737	52.440
3)	MA gamma-BHC...	4.301	3.660	336.8E6	483.0E6	50.928	51.858
4)	MA Heptachlor	4.892	4.007	322.6E6	451.9E6	48.367	48.561
5)	MB Aldrin	5.231	4.289	323.8E6	451.1E6	51.117	51.972
6)	B beta-BHC	4.489	3.956	137.2E6	198.4E6	50.337	48.846
7)	B delta-BHC	4.733	4.188	318.3E6	467.2E6	52.449m	50.823
8)	B Heptachlo...	5.650	4.791	298.1E6	405.6E6	51.871	51.137
9)	A Endosulfan I	6.032	5.162	265.4E6	380.9E6	51.229	49.569
10)	B gamma-Chl...	5.904	5.043	290.8E6	444.0E6	51.441	50.837
11)	B alpha-Chl...	5.984	5.107	288.7E6	426.0E6	50.823	51.327
12)	B 4,4'-DDE	6.154	5.296	252.8E6	396.5E6	53.047	51.593
13)	MA Dieldrin	6.304	5.426	281.0E6	421.9E6	51.252	51.597
14)	MA Endrin	6.530	5.701	209.5E6	340.2E6	45.418	45.300
15)	B Endosulfa...	6.744	5.993	238.4E6	356.2E6	48.474	50.104
16)	A 4,4'-DDD	6.664	5.848	216.2E6	354.9E6	57.338	55.031
17)	MA 4,4'-DDT	6.978	6.100	185.7E6	293.1E6	43.927	42.353
18)	B Endrin al...	6.872	6.171	173.5E6	270.2E6	56.275	50.633
19)	B Endosulfa...	7.106	6.394	208.8E6	337.9E6	51.150	49.324
20)	A Methoxychlor	7.451	6.672	87648949	153.8E6	42.030	40.795
21)	B Endrin ke...	7.585	6.898	232.3E6	408.2E6	54.730	54.206
22)	Mirex	8.063	7.087	166.3E6	284.0E6	46.829	47.498



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:47
Operator : AR\AJ
Sample : PB169782BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169782BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/12/25
Project:	PGW Phila. Gas Works	Date Received:	09/12/25
Client Sample ID:	291376MS	SDG No.:	Q3092
Lab Sample ID:	Q3092-02MS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097335.D	1	09/22/25 10:05	09/22/25 19:47	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	4.70		0.037	0.50	ug/L
76-44-8	Heptachlor	4.50		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	4.70		0.096	0.50	ug/L
72-20-8	Endrin	4.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	4.20		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.9		57 - 171	84%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.4		61 - 148	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097335.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 19:47
 Operator : AR\AJ
 Sample : Q3092-02MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 291376MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.825	74231591	114.3E6	16.397m	17.383
28)	SA Decachlor...	8.999	7.984	55196689	87385123	16.868	15.353
Target Compounds							
2)	A alpha-BHC	3.974	3.330	324.7E6	466.2E6	46.465	46.821
3)	MA gamma-BHC...	4.302	3.661	303.7E6	434.0E6	45.923	46.590
4)	MA Heptachlor	4.893	4.009	301.1E6	416.4E6	45.135	44.749
5)	MB Aldrin	5.232	4.291	286.5E6	399.5E6	45.241	46.027
6)	B beta-BHC	4.489	3.957	124.8E6	176.2E6	45.770	43.372
7)	B delta-BHC	4.734	4.190	290.5E6	421.8E6	47.869m	45.886
8)	B Heptachlo...	5.651	4.793	271.4E6	370.7E6	47.220	46.740
9)	A Endosulfan I	6.033	5.163	238.2E6	351.2E6	45.969	45.701
10)	B gamma-Chl...	5.905	5.045	261.3E6	404.4E6	46.214	46.305
11)	B alpha-Chl...	5.985	5.108	257.9E6	389.1E6	45.399	46.882
12)	B 4,4'-DDE	6.155	5.297	222.3E6	360.2E6	46.650	46.868
13)	MA Dieldrin	6.305	5.428	251.4E6	384.0E6	45.859	46.957
14)	MA Endrin	6.531	5.702	203.3E6	337.7E6	44.076	44.970
15)	B Endosulfa...	6.745	5.994	210.5E6	324.0E6	42.807	45.574
16)	A 4,4'-DDD	6.664	5.849	182.2E6	305.5E6	48.310	47.378
17)	MA 4,4'-DDT	6.978	6.102	188.1E6	292.1E6	44.503	42.203
18)	B Endrin al...	6.873	6.172	153.8E6	243.7E6	49.881	45.660
19)	B Endosulfa...	7.107	6.396	188.0E6	302.2E6	46.057	44.125
20)	A Methoxychlor	7.451	6.674	88194366	152.3E6	42.292	40.398
21)	B Endrin ke...	7.586	6.900	200.8E6	335.5E6	47.304	44.558
22)	Mirex	8.065	7.089	149.9E6	249.0E6	42.217	41.646

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097335.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 19:47
Operator : AR\AJ
Sample : Q3092-02MS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

291376MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Report of Analysis

Client:	Atlantic Recovery Services, Inc.	Date Collected:	09/12/25
Project:	PGW Phila. Gas Works	Date Received:	09/12/25
Client Sample ID:	291376MSD	SDG No.:	Q3092
Lab Sample ID:	Q3092-02MSD	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097336.D	1	09/22/25 10:05	09/22/25 20:00	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	4.60		0.037	0.50	ug/L
76-44-8	Heptachlor	4.50		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	4.70		0.096	0.50	ug/L
72-20-8	Endrin	4.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	4.20		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.9		57 - 171	85%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.1		61 - 148	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:00
 Operator : AR\AJ
 Sample : Q3092-02MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 291376MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.824	74636866	112.8E6	16.486m	17.149
28)	SA Decachlor...	8.999	7.984	55476108	85996726	16.954	15.109
Target Compounds							
2)	A alpha-BHC	3.975	3.329	324.2E6	461.3E6	46.399	46.324
3)	MA gamma-BHC...	4.302	3.661	303.3E6	430.2E6	45.865	46.185
4)	MA Heptachlor	4.893	4.009	301.0E6	412.6E6	45.125	44.340
5)	MB Aldrin	5.232	4.291	286.3E6	396.1E6	45.196	45.636
6)	B beta-BHC	4.489	3.957	124.8E6	174.1E6	45.780	42.855
7)	B delta-BHC	4.733	4.190	289.1E6	418.4E6	47.643m	45.514
8)	B Heptachlo...	5.651	4.792	270.4E6	368.5E6	47.053	46.460
9)	A Endosulfan I	6.033	5.163	238.3E6	348.9E6	45.994	45.402
10)	B gamma-Chl...	5.905	5.044	262.2E6	401.5E6	46.375	45.970
11)	B alpha-Chl...	5.985	5.108	257.7E6	386.7E6	45.377	46.588
12)	B 4,4'-DDE	6.156	5.297	222.7E6	357.0E6	46.746	46.451
13)	MA Dieldrin	6.305	5.428	251.0E6	380.7E6	45.781	46.563
14)	MA Endrin	6.531	5.702	207.4E6	335.7E6	44.958	44.692
15)	B Endosulfa...	6.745	5.994	212.6E6	323.9E6	43.219	45.563
16)	A 4,4'-DDD	6.664	5.849	181.7E6	305.3E6	48.197	47.338
17)	MA 4,4'-DDT	6.979	6.102	187.6E6	291.5E6	44.377	42.118
18)	B Endrin al...	6.873	6.172	153.4E6	243.6E6	49.749	45.646
19)	B Endosulfa...	7.107	6.395	187.2E6	302.9E6	45.861	44.222
20)	A Methoxychlor	7.451	6.674	87326449	151.9E6	41.876	40.301
21)	B Endrin ke...	7.586	6.900	197.6E6	335.1E6	46.560	44.502
22)	Mirex	8.064	7.089	149.6E6	247.8E6	42.150	41.443

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:00
Operator : AR\AJ
Sample : Q3092-02MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

291376MSD

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Manual Integration Report

Sequence:	PL092025	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL097269.D	4,4"-DDD	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	4,4"-DDD #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin aldehyde	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin aldehyde #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin ketone	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin ketone #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	alpha-Chlordane #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	delta-BHC	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endosulfan II	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endrin aldehyde #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endrin ketone #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	gamma-Chlordane #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Heptachlor epoxide	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:	PL092025	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDICV050	PL097286.D	4,4"-DDD	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	4,4"-DDE	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	4,4"-DDE #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	alpha-Chlordane	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	delta-BHC	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Dieldrin #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endosulfan I	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endosulfan II	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endrin	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endrin ketone #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	gamma-Chlordane	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Heptachlor epoxide	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Heptachlor epoxide #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:

PL092025

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PCHLORICV500	PL097287.D	Chlordane-2	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-3	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-4	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-4 #2	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-5	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PTOXICV500	PL097288.D	Toxaphene-1 #2	Abdul	9/22/2025 9:04:35 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:	pl092225	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL097316.D	Endrin aldehyde	Abdul	9/23/2025 8:41:59 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097317.D	delta-BHC	Abdul	9/23/2025 8:42:03 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097328.D	delta-BHC	Abdul	9/23/2025 8:42:21 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02	PL097334.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:18 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MS	PL097335.D	delta-BHC	Abdul	9/23/2025 8:43:22 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MS	PL097335.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:22 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MSD	PL097336.D	delta-BHC	Abdul	9/23/2025 8:43:24 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MSD	PL097336.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:24 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PEM	PL097340.D	Endrin aldehyde	Abdul	9/23/2025 8:43:27 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097341.D	4,4"-DDE #2	Abdul	9/23/2025 8:43:30 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097341.D	delta-BHC	Abdul	9/23/2025 8:43:30 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-04	PL097342.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:34 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097355.D	delta-BHC	Abdul	9/23/2025 8:44:17 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

pl092225

Instrument

ECD_I

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

pl092325

Instrument

ECD_I

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097267.D	18 Sep 2025 16:11	AR\AJ	Ok
2	I.BLK	PL097268.D	18 Sep 2025 16:25	AR\AJ	Ok
3	PEM	PL097269.D	18 Sep 2025 16:39	AR\AJ	Ok,M
4	RESCHK	PL097270.D	18 Sep 2025 16:52	AR\AJ	Ok
5	PSTDICC100	PL097271.D	18 Sep 2025 17:06	AR\AJ	Ok
6	PSTDICC075	PL097272.D	18 Sep 2025 17:19	AR\AJ	Ok
7	PSTDICC050	PL097273.D	18 Sep 2025 17:33	AR\AJ	Ok
8	PSTDICC025	PL097274.D	18 Sep 2025 17:47	AR\AJ	Ok
9	PSTDICC005	PL097275.D	18 Sep 2025 18:00	AR\AJ	Ok,M
10	PCHLORICC1000	PL097276.D	18 Sep 2025 18:14	AR\AJ	Ok
11	PCHLORICC750	PL097277.D	18 Sep 2025 18:28	AR\AJ	Ok
12	PCHLORICC500	PL097278.D	18 Sep 2025 18:41	AR\AJ	Ok
13	PCHLORICC250	PL097279.D	18 Sep 2025 18:55	AR\AJ	Ok
14	PCHLORICC050	PL097280.D	18 Sep 2025 19:08	AR\AJ	Ok
15	PTOXICC1000	PL097281.D	18 Sep 2025 19:22	AR\AJ	Ok
16	PTOXICC750	PL097282.D	18 Sep 2025 19:35	AR\AJ	Ok,M
17	PTOXICC500	PL097283.D	18 Sep 2025 19:49	AR\AJ	Ok
18	PTOXICC250	PL097284.D	18 Sep 2025 20:03	AR\AJ	Ok
19	PTOXICC100	PL097285.D	18 Sep 2025 20:16	AR\AJ	Ok,M
20	PSTDICV050	PL097286.D	19 Sep 2025 08:39	AR\AJ	Ok,M
21	PCHLORICV500	PL097287.D	19 Sep 2025 08:52	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL097288.D	19 Sep 2025 09:25	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097314.D	22 Sep 2025 10:12	AR\AJ	Ok
2	I.BLK	PL097315.D	22 Sep 2025 11:06	AR\AJ	Ok
3	PEM	PL097316.D	22 Sep 2025 11:20	AR\AJ	Ok,M
4	PSTDCCC050	PL097317.D	22 Sep 2025 11:33	AR\AJ	Ok,M
5	PB169749BS	PL097318.D	22 Sep 2025 12:31	AR\AJ	Ok,M
6	PB169772BL	PL097319.D	22 Sep 2025 13:30	AR\AJ	Ok
7	PB169772BS	PL097320.D	22 Sep 2025 13:44	AR\AJ	Ok,M
8	Q3152-01	PL097321.D	22 Sep 2025 14:05	AR\AJ	Ok,M
9	Q3135-01	PL097322.D	22 Sep 2025 14:19	AR\AJ	Ok,M
10	Q3151-02	PL097323.D	22 Sep 2025 14:32	AR\AJ	Ok
11	PB169766BL	PL097324.D	22 Sep 2025 14:46	AR\AJ	Ok
12	Q3153-04	PL097325.D	22 Sep 2025 17:30	AR\AJ	Ok
13	Q3153-08	PL097326.D	22 Sep 2025 17:44	AR\AJ	Ok
14	I.BLK	PL097327.D	22 Sep 2025 17:58	AR\AJ	Ok
15	PSTDCCC050	PL097328.D	22 Sep 2025 18:11	AR\AJ	Ok,M
16	PB169766BS	PL097329.D	22 Sep 2025 18:25	AR\AJ	Not Ok
17	PB169766BSD	PL097330.D	22 Sep 2025 18:39	AR\AJ	Not Ok
18	PB169782BL	PL097331.D	22 Sep 2025 18:52	AR\AJ	Ok
19	PB169782BS	PL097332.D	22 Sep 2025 19:06	AR\AJ	Not Ok
20	Q3065-04	PL097333.D	22 Sep 2025 19:20	AR\AJ	Ok
21	Q3092-02	PL097334.D	22 Sep 2025 19:33	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3092-02MS	PL097335.D	22 Sep 2025 19:47	AR\AJ	Ok,M
23	Q3092-02MSD	PL097336.D	22 Sep 2025 20:00	AR\AJ	Ok,M
24	PB169677TB	PL097337.D	22 Sep 2025 20:14	AR\AJ	Ok
25	PB169754TB	PL097338.D	22 Sep 2025 20:28	AR\AJ	Ok
26	I.BLK	PL097339.D	22 Sep 2025 20:41	AR\AJ	Ok
27	PEM	PL097340.D	22 Sep 2025 20:55	AR\AJ	Ok,M
28	PSTDCCC050	PL097341.D	22 Sep 2025 21:08	AR\AJ	Ok,M
29	Q3092-04	PL097342.D	22 Sep 2025 21:49	AR\AJ	Ok,M
30	Q3157-02	PL097343.D	22 Sep 2025 22:03	AR\AJ	Ok
31	Q3145-01	PL097344.D	22 Sep 2025 22:16	AR\AJ	Ok,M
32	Q3145-03	PL097345.D	22 Sep 2025 22:30	AR\AJ	Ok,M
33	Q3153-01	PL097346.D	22 Sep 2025 22:44	AR\AJ	Ok,M
34	Q3153-01MS	PL097347.D	22 Sep 2025 22:57	AR\AJ	Ok,M
35	Q3153-01MSD	PL097348.D	22 Sep 2025 23:11	AR\AJ	Ok,M
36	Q3153-05	PL097349.D	22 Sep 2025 23:24	AR\AJ	Ok,M
37	Q3155-01	PL097350.D	22 Sep 2025 23:38	AR\AJ	Ok,M
38	Q3155-02	PL097351.D	22 Sep 2025 23:52	AR\AJ	Ok,M
39	Q3155-03	PL097352.D	23 Sep 2025 00:05	AR\AJ	Ok,M
40	Q3157-01	PL097353.D	23 Sep 2025 00:19	AR\AJ	Ok,M
41	I.BLK	PL097354.D	23 Sep 2025 00:32	AR\AJ	Ok
42	PSTDCCC050	PL097355.D	23 Sep 2025 00:46	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092325

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	PL092325	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097356.D	23 Sep 2025 09:02	AR\AJ	Ok
2	I.BLK	PL097357.D	23 Sep 2025 09:16	AR\AJ	Ok
3	PEM	PL097358.D	23 Sep 2025 09:29	AR\AJ	Ok,NR
4	PSTDCCC050	PL097359.D	23 Sep 2025 10:03	AR\AJ	Ok,NR
5	PB169766BS	PL097360.D	23 Sep 2025 11:14	AR\AJ	Ok,NR
6	PB169766BSD	PL097361.D	23 Sep 2025 11:34	AR\AJ	Ok,NR
7	PB169782BS	PL097362.D	23 Sep 2025 11:47	AR\AJ	Ok,NR
8	I.BLK	PL097363.D	23 Sep 2025 12:09	AR\AJ	Ok
9	PSTDCCC050	PL097364.D	23 Sep 2025 12:52	AR\AJ	Ok,NR

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method p092025 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097267.D	18 Sep 2025 16:11		AR\AJ	Ok
2	I.BLK	I.BLK	PL097268.D	18 Sep 2025 16:25		AR\AJ	Ok
3	PEM	PEM	PL097269.D	18 Sep 2025 16:39		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL097270.D	18 Sep 2025 16:52		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL097271.D	18 Sep 2025 17:06		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL097272.D	18 Sep 2025 17:19		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL097273.D	18 Sep 2025 17:33		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL097274.D	18 Sep 2025 17:47		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL097275.D	18 Sep 2025 18:00		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL097276.D	18 Sep 2025 18:14		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL097277.D	18 Sep 2025 18:28		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL097278.D	18 Sep 2025 18:41		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL097279.D	18 Sep 2025 18:55		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL097280.D	18 Sep 2025 19:08		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL097281.D	18 Sep 2025 19:22		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL097282.D	18 Sep 2025 19:35		AR\AJ	Ok,M
17	PTOXICC500	PTOXICC500	PL097283.D	18 Sep 2025 19:49		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL097284.D	18 Sep 2025 20:03		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL097285.D	18 Sep 2025 20:16		AR\AJ	Ok,M
20	PSTDICV050	ICVPL092025	PL097286.D	19 Sep 2025 08:39		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL092025	PL097287.D	19 Sep 2025 08:52		AR\AJ	Ok,M
22	PTOXICV500	ICVPL092025	PL097288.D	19 Sep 2025 09:25		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method p1092025 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097314.D	22 Sep 2025 10:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL097315.D	22 Sep 2025 11:06		AR\AJ	Ok
3	PEM	PEM	PL097316.D	22 Sep 2025 11:20		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL097317.D	22 Sep 2025 11:33		AR\AJ	Ok,M
5	PB169749BS	PB169749BS	PL097318.D	22 Sep 2025 12:31		AR\AJ	Ok,M
6	PB169772BL	PB169772BL	PL097319.D	22 Sep 2025 13:30		AR\AJ	Ok
7	PB169772BS	PB169772BS	PL097320.D	22 Sep 2025 13:44		AR\AJ	Ok,M
8	Q3152-01	EO-02-091925	PL097321.D	22 Sep 2025 14:05		AR\AJ	Ok,M
9	Q3135-01	MH-121	PL097322.D	22 Sep 2025 14:19	DCB high in 2nd column	AR\AJ	Ok,M
10	Q3151-02	1402	PL097323.D	22 Sep 2025 14:32	DCB high in 2nd column	AR\AJ	Ok
11	PB169766BL	PB169766BL	PL097324.D	22 Sep 2025 14:46		AR\AJ	Ok
12	Q3153-04	OR-500-COMP-39	PL097325.D	22 Sep 2025 17:30		AR\AJ	Ok
13	Q3153-08	OR-500-COMP-40	PL097326.D	22 Sep 2025 17:44		AR\AJ	Ok
14	I.BLK	I.BLK	PL097327.D	22 Sep 2025 17:58		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL097328.D	22 Sep 2025 18:11		AR\AJ	Ok,M
16	PB169766BS	PB169766BS	PL097329.D	22 Sep 2025 18:25	Almost all compound recovery fail	AR\AJ	Not Ok
17	PB169766BSD	PB169766BSD	PL097330.D	22 Sep 2025 18:39	Almost all compound recovery fail	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PB169782BL	PB169782BL	PL097331.D	22 Sep 2025 18:52		AR\AJ	Ok
19	PB169782BS	PB169782BS	PL097332.D	22 Sep 2025 19:06	Almost all compound recovery fail	AR\AJ	Not Ok
20	Q3065-04	VNJ-204	PL097333.D	22 Sep 2025 19:20		AR\AJ	Ok
21	Q3092-02	291376	PL097334.D	22 Sep 2025 19:33		AR\AJ	Ok,M
22	Q3092-02MS	291376MS	PL097335.D	22 Sep 2025 19:47		AR\AJ	Ok,M
23	Q3092-02MSD	291376MSD	PL097336.D	22 Sep 2025 20:00		AR\AJ	Ok,M
24	PB169677TB	PB169677TB	PL097337.D	22 Sep 2025 20:14		AR\AJ	Ok
25	PB169754TB	PB169754TB	PL097338.D	22 Sep 2025 20:28		AR\AJ	Ok
26	I.BLK	I.BLK	PL097339.D	22 Sep 2025 20:41		AR\AJ	Ok
27	PEM	PEM	PL097340.D	22 Sep 2025 20:55		AR\AJ	Ok,M
28	PSTDCCC050	PSTDCCC050	PL097341.D	22 Sep 2025 21:08		AR\AJ	Ok,M
29	Q3092-04	279182	PL097342.D	22 Sep 2025 21:49		AR\AJ	Ok,M
30	Q3157-02	20-DEAD-1	PL097343.D	22 Sep 2025 22:03		AR\AJ	Ok
31	Q3145-01	SP-A	PL097344.D	22 Sep 2025 22:16		AR\AJ	Ok,M
32	Q3145-03	SP-B	PL097345.D	22 Sep 2025 22:30		AR\AJ	Ok,M
33	Q3153-01	OR-500-COMP-39	PL097346.D	22 Sep 2025 22:44		AR\AJ	Ok,M
34	Q3153-01MS	OR-500-COMP-39MS	PL097347.D	22 Sep 2025 22:57		AR\AJ	Ok,M
35	Q3153-01MSD	OR-500-COMP-39MSD	PL097348.D	22 Sep 2025 23:11		AR\AJ	Ok,M
36	Q3153-05	OR-500-COMP-40	PL097349.D	22 Sep 2025 23:24		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q3155-01	COMP-1	PL097350.D	22 Sep 2025 23:38		AR\AJ	Ok,M
38	Q3155-02	COMP-2	PL097351.D	22 Sep 2025 23:52		AR\AJ	Ok,M
39	Q3155-03	COMP-3	PL097352.D	23 Sep 2025 00:05		AR\AJ	Ok,M
40	Q3157-01	20-DEAD-1	PL097353.D	23 Sep 2025 00:19		AR\AJ	Ok,M
41	I.BLK	I.BLK	PL097354.D	23 Sep 2025 00:32		AR\AJ	Ok
42	PSTDCCC050	PSTDCCC050	PL097355.D	23 Sep 2025 00:46		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092325

Review By	Review On
Supervise By	Supervise On
SubDirectory	PL092325
HP Acquire Method	HP Processing Method
	p1092025 8081
STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097356.D	23 Sep 2025 09:02		AR\AJ	Ok
2	I.BLK	I.BLK	PL097357.D	23 Sep 2025 09:16		AR\AJ	Ok
3	PEM	PEM	PL097358.D	23 Sep 2025 09:29		AR\AJ	Ok,NR
4	PSTDCCC050	PSTDCCC050	PL097359.D	23 Sep 2025 10:03		AR\AJ	Ok,NR
5	PB169766BS	PB169766BS	PL097360.D	23 Sep 2025 11:14		AR\AJ	Ok,NR
6	PB169766BSD	PB169766BSD	PL097361.D	23 Sep 2025 11:34		AR\AJ	Ok,NR
7	PB169782BS	PB169782BS	PL097362.D	23 Sep 2025 11:47		AR\AJ	Ok,NR
8	I.BLK	I.BLK	PL097363.D	23 Sep 2025 12:09		AR\AJ	Ok
9	PSTDCCC050	PSTDCCC050	PL097364.D	23 Sep 2025 12:52		AR\AJ	Ok,NR

M : Manual Integration

SOP ID :	M1311-TCLP-16	
SDG No :	N/A	Start Prep Date : 09/12/2025 Time : 17:00
Weigh By :	JP	End Prep Date : 09/13/2025 Time : 10:40
Balance ID :	WC SC-7	Combination Ratio : 20
pH Meter ID :	WC PH METER-1	ZHE Cleaning Batch : N/A ⁷⁸
Extraction By :	JP	Initial Room Temperature: 24 °C
Filter By :	JP	Final Room Temperature: 22 °C
Pipette ID :	WC	TCLP Technician Signature : ^{JP}
Tumbler ID :	T-1 / T-2	Supervisor By : ¹²
TCLP Filter ID :	40819719	

Standarded Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP114525
HCL-TCLP,1N	N/A	WP114527
HNO3-TCLP,1N	N/A	WP114528
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP86978

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 / T-2 checked,30 rpm. Q3092-04 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/15/25 10:00	^{JP} 100 Room	^{SKS.} RJ/EXT
	Preparation Group	Analysis Group ^{met my}

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169677TB	LEB677	11	N/A	2000	N/A	N/A	N/A	4.94	1.5	T-2
Q3082-02	WC-29	01	100.03	2000	N/A	N/A	N/A	4.0	1.0	T-1
Q3082-04	WC-30	02	100.02	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3082-06	WC-31	03	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q3082-08	WC-32	04	100.04	2000	N/A	N/A	N/A	6.2	1.0	T-1
Q3082-10	WC-33	05	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q3085-01	PAULDING-WC1	06	100.02	2000	N/A	N/A	N/A	7.6	1.5	T-1
Q3091-07	12/9C-COMP	07	100.03	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3091-14	12/9B-COMP	08	100.02	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3092-02	291376	09	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-1
Q3092-04	279182	10	100.02	2000	N/A	N/A	N/A	5.0	1.5	T-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB169677TB	LEB677	N/A	N/A	N/A	N/A	N/A	N/A
Q3082-02	WC-29	N/A	N/A	N/A	N/A	100	N/A
Q3082-04	WC-30	N/A	N/A	N/A	N/A	100	N/A
Q3082-06	WC-31	N/A	N/A	N/A	N/A	100	N/A
Q3082-08	WC-32	N/A	N/A	N/A	N/A	100	N/A
Q3082-10	WC-33	N/A	N/A	N/A	N/A	100	N/A
Q3085-01	PAULDING-WC1	N/A	N/A	N/A	N/A	100	N/A
Q3091-07	12/9C-COMP	N/A	N/A	N/A	N/A	100	N/A
Q3091-14	12/9B-COMP	N/A	N/A	N/A	N/A	100	N/A
Q3092-02	291376	N/A	N/A	N/A	N/A	100	N/A
Q3092-04	279182	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2
Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB169677TB	LEB677	N/A	N/A	N/A	N/A	#1	4.94
Q3082-02	WC-29	5.02	96.5	7.0	2.5	#1	4.94
Q3082-04	WC-30	5.03	96.5	7.2	2.0	#1	4.94
Q3082-06	WC-31	5.03	96.5	7.2	2.5	#1	4.94
Q3082-08	WC-32	5.02	96.5	8.6	3.5	#1	4.94
Q3082-10	WC-33	5.01	96.5	8.4	3.0	#1	4.94
Q3085-01	PAULDING-WC1	5.02	96.5	9.1	3.5	#1	4.94
Q3091-07	12/9C-COMP	5.03	96.5	5.6	1.5	#1	4.94
Q3091-14	12/9B-COMP	5.02	96.5	5.8	1.5	#1	4.94
Q3092-02	291376	5.01	96.5	6.6	2.0	#1	4.94
Q3092-04	279182	5.02	96.5	7.0	2.5	#1	4.94

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tclp q3081 WorkList ID : 191840 Department : TCLP Extraction Date : 09-12-2025 09:16:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3082-02	WC-29	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/11/2025	1311
Q3082-04	WC-30	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/11/2025	1311
Q3082-06	WC-31	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/11/2025	1311
Q3082-08	WC-32	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/11/2025	1311
Q3082-10	WC-33	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/11/2025	1311
Q3085-01	PAULDING-WC1	Solid	TCLP Extraction	Cool 4 deg C	CORE02	D31	08/29/2025	1311
Q3091-07	12/9C-COMP	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/12/2025	1311
Q3091-14	12/9B-COMP	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/12/2025	1311
Q3092-02	291376	Solid	TCLP Extraction	Cool 4 deg C	ATLA10	D41	09/12/2025	1311
Q3092-04	279182	Solid	TCLP Extraction	Cool 4 deg C	ATLA10	D41	09/12/2025	1311

Date/Time 09.12.25 14:10
Raw Sample Received by: JG WOC
Raw Sample Relinquished by: JG WOC

Date/Time 09.12.25 14:10
Raw Sample Received by: JG WOC
Raw Sample Relinquished by: JG WOC

SOP ID: M3510C,3580A-Extraction Pesticide-17

Clean Up SOP #: Florisil

Matrix : Water

Weight By: N/A **Extraction By:** RJ

Balance check: N/A **Filter By:** RJ

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3953 **Hood ID:** 4,5,6,7

Extraction Method: ☒ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 09/22/2025

Extraction Start Time : 10:05

Extraction End Date : 09/22/2025

Extraction End Time : 14:30

Concentration By: EH

Supervisor By : ritesh

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3973
Baked Na2SO4	N/A	EP2639
Hexane	N/A	E3971
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A **Envap ID:** NEVAP-02


KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
9/22/25	RJ (Ext-106)	RJ (Ext-106)
14:35	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-17

Concentration Date: 09/22/2025

Sample ID	Client Sample ID	Test	g / ml	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB169677TB	PB169677TB	TCLP Pesticide	100	6	ritesh	Evelyn	10			SEP-1
PB169754TB	PB169754TB	TCLP Pesticide	100	6	ritesh	Evelyn	10			2
PB169782BL	PBLK782	TCLP Pesticide	1000	6	ritesh	Evelyn	10			3
PB169782BS	PLCS782	TCLP Pesticide	1000	6	ritesh	Evelyn	10			4
Q3065-04	VNJ-204	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		5
Q3092-02	291376	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		6
Q3092-02MS	291376MS	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		7
Q3092-02MS D	291376MSD	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		8
Q3092-04	279182	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		9
Q3153-04	OR-500-COMP-39	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		10
Q3153-08	OR-500-COMP-40	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		11
Q3157-02	20-DEAD-1	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		12



RJ
9/22

* Extracts relinquished on the same date as received.

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169677TB	LEB677	11	N/A	2000	N/A	N/A	N/A	4.94	1.5	T-2
Q3082-02	WC-29	01	100.03	2000	N/A	N/A	N/A	4.0	1.0	T-1
Q3082-04	WC-30	02	100.02	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3082-06	WC-31	03	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q3082-08	WC-32	04	100.04	2000	N/A	N/A	N/A	6.2	1.0	T-1
Q3082-10	WC-33	05	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q3085-01	PAULDING-WC1	06	100.02	2000	N/A	N/A	N/A	7.6	1.5	T-1
Q3091-07	12/9C-COMP	07	100.03	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3091-14	12/9B-COMP	08	100.02	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3092-02	291376	09	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-1
Q3092-04	279182	10	100.02	2000	N/A	N/A	N/A	5.0	1.5	T-1

09/11/15/125
101.00

TCLP EXTRACTION LOGPAGE

PB169754

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169754TB	LEB754	17	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-2
Q3065-04	VNJ-204	01	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3133-02	VNJ-238	02	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-04	VNJ-222	03	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3133-06	361	04	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-08	VNJ-210	05	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3136-01	34922	06	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3136-02	34923	07	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3136-03	34924	08	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3136-04	34925	09	100.01	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3136-05	34926	10	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3138-08	MOO-25-0239-0241	11	100.03	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q3150-01	MER Rd Con Ed	12	100.02	2000	N/A	N/A	N/A	7.2	1.0	T-2
Q3150-03	Mid Site Grid 5-7	13	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q3153-04	OR-500-COMP-39	14	100.02	2000	N/A	N/A	N/A	5.8	1.0	T-2
Q3153-08	OR-500-COMP-40	15	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q3157-02	20-DEAD-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

09/19/25
701.00



SHIPPING DOCUMENTS

CLIENT INFORMATION

COMPANY: **PGW / ATLANTIC**
ADDRESS: **3100 E. Venango**
CITY: **Philadelphia** STATE: **PA** ZIP:
ATTENTION: **Kenny Pyle**
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **Atlantic Recovery Services**
PROJECT NO.: LOCATION:
PROJECT MANAGER: **Kenny Pyle**
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. **TCLP VOA**
2. **TCLP ZHE Extra**
3. **TCLP BNA**
4. **TCLP Herbicide**
5. **TCLP ICP Metals**
6. **TCLP Mercury**
7. **TCLP Pesticide**
8.
9.

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	291376	Soil		X	9-12-25	1111	2	X									0.2ppm	
2.	291376		X			1122	4		X	X	X	X	X	X				
3.	279182			X		1131	2	X									0.0ppm	
4.	279182		X			1143	4		X	X	X	X	X	X				
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. JT	DATE/TIME: 1200 9-12-25	RECEIVED BY: 1.	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.7 °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments:
RELINQUISHED BY SAMPLER: 3. JT	DATE/TIME: 1340 9-12-25	RECEIVED BY: 3.	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO



Environmental Laboratory

www.chemtech.net | EMAIL: PM@chemtech.net

Project Name: Atlantic Recovery Services

Service Order #:

Sampler Name: IT

Work Order #:

Client Project Coordinator & Phone:

Labor WBS #:

Page #: 1 of 1

Facility/Site:

Date: 9.12.25

Site Address:

Arrive Time: 1100

PGW Venango Plant
Philadelphia PA.

Depart Time: 1200

Waste Stream (circle one): drum / oil-dr / soil pile / in-situ / linear construction / frac-tank

Sample Matrices (circle all that apply): Water / Soil / NAPL / Concrete / Wipe

Collection Depths: N/A

Temp (range): _____ °C PID Readings (range): 0.0-0.2 PPM N/A



Sample Description: wet well silt

Odor: Y / N Color: Y Q

Field Observations: 2 Roll-offs

Grid/Area Composite Map:

QA Control # A3041134



PGW Venango Location

Sampler Signature: IT

Supervisor Review/Date: _____

Client Signature: _____

Date/Time Arrived at Lab: _____

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312