

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091625\
 Data File : BF143767.D
 Acq On : 16 Sep 2025 11:23
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Sep 16 16:05:45 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 16 15:40:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	9211	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	37291	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	19751	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	32987	20.000	ng	0.00
76) Chrysene-d12	14.063	240	18916	20.000	ng	-0.01
86) Perylene-d12	15.545	264	15882	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	2535	4.394	ng	-0.01
7) Phenol-d6	6.498	99	3369	5.000	ng	-0.03
23) Nitrobenzene-d5	7.445	82	3111	4.936	ng	-0.02
42) 2,4,6-Tribromophenol	10.716	330	877	4.402	ng	-0.02
45) 2-Fluorobiphenyl	9.245	172	7359	5.129	ng	-0.01
79) Terphenyl-d14	13.004	244	7156	5.892	ng	-0.01
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.287	70	1054	2.692	ng	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091625\
Data File : BF143767.D
Acq On : 16 Sep 2025 11:23
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Sep 16 16:05:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 16 15:40:08 2025
Response via : Initial Calibration

