

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091025

Review By	Rahul	Review On	9/10/2025 9:03:37 AM
Supervise By	Jagrut	Supervise On	9/10/2025 11:02:08 AM
SubDirectory	BF091025	HP Acquire Method	BNA_F
		HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143675.D	09 Sep 2025 08:44		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143676.D	09 Sep 2025 09:12		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143677.D	09 Sep 2025 09:41		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF143678.D	09 Sep 2025 10:10		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF143679.D	09 Sep 2025 10:39		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF143680.D	09 Sep 2025 11:08		RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF143681.D	09 Sep 2025 11:37		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF143682.D	09 Sep 2025 12:07		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF143683.D	09 Sep 2025 13:54		RC/JU	Ok,M
10	SSTDICV040	ICVBF091025	BF143684.D	09 Sep 2025 15:03		RC/JU	Ok,M
11	PB169490BL	PB169490BL	BF143685.D	09 Sep 2025 15:32		RC/JU	Ok
12	Q2893-02	LOQ-SOIL-02-QT3-202	BF143686.D	09 Sep 2025 16:01	LOQ-SOIL-10PPM (For Benzidine only)	RC/JU	Ok
13	Q2893-08	LOQ-WATER-02-QT3-2	BF143687.D	09 Sep 2025 16:30	LOQ-WATER-10PPM (For Benzidine only)	RC/JU	Ok
14	Q2893-01	LOD-MDL-SOIL-01-QT	BF143688.D	09 Sep 2025 16:59	LOD-MDL-SOIL-8PPM (For Benzidine only)	RC/JU	Ok
15	Q2893-07	LOD-MDL-WATER-01-Q	BF143689.D	09 Sep 2025 17:28	LOD-MDL-WATER-8 PPM (For Benzidine only)	RC/JU	Ok,M
16	PB169499BL	PB169499BL	BF143690.D	09 Sep 2025 17:58		RC/JU	Ok
17	SSTDCCC040	SSTDCCC040EC	BF143691.D	09 Sep 2025 18:27		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091025

Review By	Rahul	Review On	9/10/2025 9:03:37 AM		
Supervise By	Jagrut	Supervise On	9/10/2025 11:02:08 AM		
SubDirectory	BF091025	HP Acquire Method	BNA_F	HP Processing Method	bf091025
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13174,10ul/1000ul sample				
ICV/I.BLK	SP6796				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	DFTPP	DFTPP	BF143692.D	10 Sep 2025 09:13		RC/JU	Ok
19	SSTDCCC040	SSTDCCC040	BF143693.D	10 Sep 2025 09:41		RC/JU	Ok,M
20	PB169528BL	PB169528BL	BF143694.D	10 Sep 2025 10:10		RC/JU	Ok
21	Q2893-03	MDL-SOIL-03-QT3-202	BF143695.D	10 Sep 2025 10:49	MDL-SOIL 8PPM (For Benzidine only)	RC/JU	Ok,M
22	Q2893-09	MDL-WATER-03-QT3-2	BF143696.D	10 Sep 2025 11:18	MDL-WATER 8PPM (For Benzidine only)	RC/JU	Ok,M
23	PB169526BL	PB169526BL	BF143697.D	10 Sep 2025 11:49		RC/JU	Ok
24	SP6869	SP6869	BF143698.D	10 Sep 2025 12:41		RC/JU	Ok
25	SSTDCCC040	SSTDCCC040EC	BF143699.D	10 Sep 2025 13:47		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091525

Review By	Rahul	Review On	9/16/2025 8:59:16 AM
Supervise By	Jagrut	Supervise On	9/16/2025 11:33:47 AM
SubDirectory	BF091525	HP Acquire Method	BNA_F
		HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143747.D	15 Sep 2025 12:20		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143748.D	15 Sep 2025 12:49		RC/JU	Ok
3	PB169679BL	PB169679BL	BF143749.D	15 Sep 2025 13:20		RC/JU	Ok
4	PB169679BS	PB169679BS	BF143750.D	15 Sep 2025 13:49		RC/JU	Ok,M
5	PB169640TB	PB169640TB	BF143751.D	15 Sep 2025 14:18		RC/JU	Ok
6	PB169696BL	PB169696BL	BF143752.D	15 Sep 2025 15:22	Surrogate Fail	RC/JU	Not Ok
7	PB169696BS	PB169696BS	BF143753.D	15 Sep 2025 15:51	Surrogate Fail & Recovery fail high for many compound	RC/JU	Not Ok
8	PB169677TB	PB169677TB	BF143754.D	15 Sep 2025 16:20		RC/JU	Ok
9	Q3054-08	MH-384	BF143755.D	15 Sep 2025 16:59		RC/JU	Ok
10	Q3082-02	WC-29	BF143756.D	15 Sep 2025 17:29		RC/JU	Ok
11	Q3082-02MS	WC-29MS	BF143757.D	15 Sep 2025 17:58		RC/JU	Ok,M
12	Q3082-02MSD	WC-29MSD	BF143758.D	15 Sep 2025 18:28		RC/JU	Ok,M
13	Q3082-04	WC-30	BF143759.D	15 Sep 2025 18:57		RC/JU	Ok
14	Q3082-06	WC-31	BF143760.D	15 Sep 2025 19:27		RC/JU	Ok
15	Q3082-08	WC-32	BF143761.D	15 Sep 2025 19:56		RC/JU	Ok
16	Q3082-10	WC-33	BF143762.D	15 Sep 2025 20:26		RC/JU	Ok
17	Q3092-02	291376	BF143763.D	15 Sep 2025 20:55		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091525

Review By	Rahul	Review On	9/16/2025 8:59:16 AM		
Supervise By	Jagrut	Supervise On	9/16/2025 11:33:47 AM		
SubDirectory	BF091525	HP Acquire Method	BNA_F	HP Processing Method	bf091025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13174,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3092-04	279182	BF143764.D	15 Sep 2025 21:24	Internal Standard Fail	RC/JU	ReRun
----	----------	--------	------------	-------------------	------------------------	-------	-------

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091625

Review By	Rahul	Review On	9/17/2025 9:11:13 AM
Supervise By	Jagrut	Supervise On	9/17/2025 10:43:20 AM
SubDirectory	BF091625	HP Acquire Method	BNA_F
		HP Processing Method	bf091625

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13175,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143765.D	16 Sep 2025 10:25		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143766.D	16 Sep 2025 10:54	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF143767.D	16 Sep 2025 11:23		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF143768.D	16 Sep 2025 11:52		RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BF143769.D	16 Sep 2025 12:21		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF143770.D	16 Sep 2025 12:50		RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF143771.D	16 Sep 2025 13:19	Compound #32,54 kept on QR and compound #65 kept on LR	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF143772.D	16 Sep 2025 13:48		RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BF143773.D	16 Sep 2025 14:17		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BF143774.D	16 Sep 2025 14:45	Compound #77 removed from 80PPM	RC/JU	Ok,M
11	SSTDICV040	ICVBF091625	BF143775.D	16 Sep 2025 16:23		RC/JU	Ok
12	PB169696BL	PB169696BL	BF143776.D	16 Sep 2025 16:53		RC/JU	Ok
13	PB169696BS	PB169696BS	BF143777.D	16 Sep 2025 17:22		RC/JU	Ok,M
14	PB169672BL	PB169672BL	BF143778.D	16 Sep 2025 17:51		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP091225

Review By	Rahul	Review On	9/12/2025 10:23:37 AM
Supervise By	Jagrut	Supervise On	9/12/2025 11:49:47 AM
SubDirectory	BP091225	HP Acquire Method	BNA_P
		HP Processing Method	BP091225

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6836 S13174,10ul/1000ul sample SP6796

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025724.D	11 Sep 2025 08:09		CG/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP025725.D	11 Sep 2025 08:56		CG/JU	Ok,M
3	SSTDICC005	SSTDICC005	BP025726.D	11 Sep 2025 09:37		CG/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP025727.D	11 Sep 2025 10:18		CG/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP025728.D	11 Sep 2025 10:59		CG/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP025729.D	11 Sep 2025 12:21	Compound#54 & 56 are Kept on LR	CG/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP025730.D	11 Sep 2025 13:02	Not Used	CG/JU	Not Ok
8	SSTDICC060	SSTDICC060	BP025731.D	11 Sep 2025 13:43		CG/JU	Ok
9	SSTDICC080	SSTDICC080	BP025732.D	11 Sep 2025 14:24		CG/JU	Ok
10	SSTDICC050	SSTDICC050	BP025733.D	11 Sep 2025 15:20		CG/JU	Ok,M
11	SSTDICV040	ICVBP091225	BP025734.D	11 Sep 2025 16:54		CG/JU	Ok,M
12	PB169633BL	PB169633BL	BP025735.D	11 Sep 2025 17:35		CG/JU	Ok
13	DFTPP	DFTPP	BP025736.D	12 Sep 2025 09:03		CG/JU	Ok
14	SSTDCCC040	SSTDCCC040	BP025737.D	12 Sep 2025 10:24		CG/JU	Ok
15	PB169490BL	PB169490BL	BP025738.D	12 Sep 2025 11:05		CG/JU	Ok
16	Q2893-02	LOQ-SOIL-02-QT3-202	BP025739.D	12 Sep 2025 11:46	LOQ-SOIL-10PPM	CG/JU	Ok,M
17	Q2893-08	LOQ-WATER-02-QT3-2	BP025740.D	12 Sep 2025 12:27	LOQ-WATER-10 PPM	CG/JU	Ok,M

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP091225

Review By	Rahul	Review On	9/12/2025 10:23:37 AM		
Supervise By	Jagrut	Supervise On	9/12/2025 11:49:47 AM		
SubDirectory	BP091225	HP Acquire Method	BNA_P	HP Processing Method	BP091225
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13174,10ul/1000ul sample				
ICV/I.BLK	SP6796				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2893-01	LOD-MDL-SOIL-01-QT	BP025741.D	12 Sep 2025 13:08	LOD-MDL-SOIL-8PPM	CG/JU	Ok,M
19	Q2893-07	LOD-MDL-WATER-01-QT	BP025742.D	12 Sep 2025 13:49	LOD-MDL-WATER-8PPM	CG/JU	Ok,M
20	PB169499BL	PB169499BL	BP025743.D	12 Sep 2025 14:30		CG/JU	Ok
21	SSTDCCC040	SSTDCCC040EC	BP025744.D	12 Sep 2025 15:35		CG/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP091625

Review By	Rahul	Review On	9/17/2025 11:07:55 AM
Supervise By	Jagrut	Supervise On	9/17/2025 11:08:04 AM
SubDirectory	BP091625	HP Acquire Method	BNA_P
		HP Processing Method	BP091225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13175,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025755.D	16 Sep 2025 12:50		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025756.D	16 Sep 2025 13:31		CG/JU	Ok,M
3	PB169701BL	PB169701BL	BP025757.D	16 Sep 2025 14:13		CG/JU	Ok
4	PB169701BS	PB169701BS	BP025758.D	16 Sep 2025 14:54		CG/JU	Ok
5	Q3091-08	12/9B-COMP	BP025759.D	16 Sep 2025 15:50		CG/JU	Ok
6	Q3091-08MS	12/9B-COMPMS	BP025760.D	16 Sep 2025 16:31		CG/JU	Ok
7	Q3091-08MSD	12/9B-COMPMSD	BP025761.D	16 Sep 2025 17:12		CG/JU	Ok
8	Q3092-04	279182	BP025762.D	16 Sep 2025 17:53		CG/JU	Ok
9	Q3091-01	12/9C-COMP	BP025763.D	16 Sep 2025 18:34		CG/JU	Ok
10	Q3105-01	FREEZE-PIT-SAMPLE	BP025764.D	16 Sep 2025 19:15		CG/JU	Ok
11	Q3104-01	SU-04-091525	BP025765.D	16 Sep 2025 19:56		CG/JU	Ok
12	Q3106-01	NB-07-09152025	BP025766.D	16 Sep 2025 20:37		CG/JU	Ok,M
13	Q3106-03	NB-08-09152025	BP025767.D	16 Sep 2025 21:18		CG/JU	Ok

M : Manual Integration