

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3092

Client ID : SSTDCCC040

Date Analyzed: 09/15/2025

Lab File ID: BF143748.D

Time Analyzed: 12:49

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	29333	6.899	108964	8.18	61007	9.93
UPPER LIMIT	58666	7.399	217928	8.681	122014	10.434
LOWER LIMIT	14666.5	6.399	54482	7.681	30503.5	9.434
EPA SAMPLE NO.						
01 PB169677TB	17700	6.89	66317	8.18	37178	9.93
02 WC-29MS	15629	6.89	59287	8.18	33245	9.93
03 WC-29MSD	17438	6.89	66564	8.18	36375	9.93
04 291376	16953	6.89	62637	8.18	34545	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.