

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3092

Client ID : SSTDICCC040

Date Analyzed: 09/16/2025

Lab File ID: BF143771.D

Time Analyzed: 13:19

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5123	6.893	18088	8.18	8844	9.93
UPPER LIMIT	10246	7.393	36176	8.675	17688	10.434
LOWER LIMIT	2561.5	6.393	9044	7.675	4422	9.434
EPA SAMPLE NO.						
01 PB169696BL	5265	6.89	17131	8.17	9057	9.93
02 PB169696BS	4147	6.89	15087	8.18	7826	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.