

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3092

Client ID : SSTDCCC040

Date Analyzed: 09/16/2025

Lab File ID: BP025756.D

Time Analyzed: 13:31

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	131105	7.754	526142	10.53	351438	14.37
UPPER LIMIT	262210	8.254	1052280	11.025	702876	14.871
LOWER LIMIT	65552.5	7.254	263071	10.025	175719	13.871
EPA SAMPLE NO.						
01 279182	149569	7.76	565486	10.53	361652	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.