

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3092

Client ID: SSTDCCC040

Date Analyzed: 09/16/2025

Lab File ID: BP025756.D

Time Analyzed: 13:31

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	718838	17.165	849419	21.63	1092590	24.994
UPPER LIMIT	1437680	17.665	1698840	22.13	2185180	25.494
LOWER LIMIT	359419	16.665	424710	21.13	546295	24.494
EPA SAMPLE NO.						
279182	738027	17.17	763749	21.62	802889	24.98

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.