

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3092</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Atlantic Recovery Services, Inc.</u>	Datafile :	<u>VN087823.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0915WBS01	Vinyl chloride	20	19.2	ug/L	96			65	117	
	1,1-Dichloroethene	20	19.2	ug/L	96			74	110	
	2-Butanone	100	91.2	ug/L	91			65	122	
	Carbon Tetrachloride	20	19.4	ug/L	97			77	113	
	Chloroform	20	19.3	ug/L	97			79	113	
	Benzene	20	18.9	ug/L	95			82	109	
	1,2-Dichloroethane	20	19.4	ug/L	97			80	115	
	Trichloroethene	20	18.5	ug/L	93			77	113	
	Tetrachloroethene	20	17.6	ug/L	88			67	123	
	Chlorobenzene	20	18.2	ug/L	91			82	109	