

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATLA10
Lab Code: ACE SDG NO.: Q3092
Lab File ID: VN087820.D Date Analyzed: 09/15/2025
Instrument ID: MSVOA_N Time Analyzed: 08:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	257288	8.21	490599	9.08	467132	11.84
UPPER LIMIT	514576	8.706	981198	9.582	934264	12.341
LOWER LIMIT	128644	7.706	245300	8.582	233566	11.341
EPA SAMPLE NO.						
291376	169484	8.21	394773	9.08	406404	11.85
279182	170877	8.21	400335	9.08	402622	11.85
VN0915WBL01	194621	8.21	447138	9.08	421003	11.84
VN0915WBS01	212267	8.21	415751	9.08	395901	11.85
VN0915WBSD01	231988	8.21	436544	9.08	400333	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.