

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATLA10
 Lab Code: ACE SDG NO.: Q3092
 Lab File ID: VN087820.D Date Analyzed: 09/15/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	259703	13.765				
UPPER LIMIT	519406	14.265				
LOWER LIMIT	129852	13.265				
EPA SAMPLE NO.						
291376	181401	13.77				
279182	182693	13.77				
VN0915WBL01	194722	13.77				
VN0915WBS01	209334	13.77				
VN0915WBSD01	213781	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.