

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ATLA10

Lab Code: ACE

SDG NO.: Q3092

Lab File ID: VN087819.D

BFB Injection Date: 09/15/2025

Instrument ID: MSVOA_N

BFB Injection Time: 08:18

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.1
75	30.0 - 60.0% of mass 95	51.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	74.1
175	5.0 - 9.0% of mass 174	6.3 (8.5) 1
176	95.0 - 101.0% of mass 174	70.8 (95.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087820.D	09/15/2025	08:39
VN0915WBL01	VN0915WBL01	VN087822.D	09/15/2025	09:35
VN0915WBS01	VN0915WBS01	VN087823.D	09/15/2025	09:56
VN0915WBSD01	VN0915WBSD01	VN087824.D	09/15/2025	10:29
291376	Q3092-01	VN087833.D	09/15/2025	13:38
279182	Q3092-03	VN087835.D	09/15/2025	14:20