

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3093

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137185

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	09/15/2025
Sample ID: CCV1 pH	pH	2.02	2.00	101	90-110	09/15/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	09/15/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3093

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137264

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	09/22/2025
Sample ID: CCV2 COD	mg/L	50.173	50	100	95-105	09/22/2025