

Initial and Continuing Calibration Verification

Client: Scientific Design
Project: Carrier Post-Treatment

SDG No.: Q3094
RunNo.: LB137188

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	10.2	10	102	90-110	07/03/2025
Sample ID: CCV1 TOC	mg/L	10.7	10	107	90-110	09/16/2025
Sample ID: CCV2 TOC	mg/L	10.5	10	105	90-110	09/16/2025
Sample ID: CCV3 TOC	mg/L	10.7	10	107	90-110	09/17/2025
Sample ID: CCV4 TOC	mg/L	10.5	10	105	90-110	09/17/2025

Initial and Continuing Calibration Verification

Client: Scientific Design

SDG No.: Q3094

Project: Carrier Post-Treatment

RunNo.: LB137188

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------