

Initial and Continuing Calibration Blank Summary

Client: Scientific Design
Project: Carrier Post-Treatment

SDG No.: Q3094
RunNo.: LB137188

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 0.5000	0.5000	U	0.40	1	07/03/2025
Sample ID: CCB1 TOC	mg/L	< 0.5000	0.5000	U	0.40	1	09/16/2025
Sample ID: CCB2 TOC	mg/L	0.46	0.5000	J	0.40	1	09/16/2025
Sample ID: CCB3 TOC	mg/L	< 0.5000	0.5000	U	0.40	1	09/17/2025
Sample ID: CCB4 TOC	mg/L	< 0.5000	0.5000	U	0.40	1	09/17/2025

Initial and Continuing Calibration Blank Summary

Client: Scientific Design

SDG No.: Q3094

Project: Carrier Post-Treatment

RunNo.: LB137188

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------