

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091925\
 Data File : BN037830.D
 Acq On : 20 Sep 2025 00:39
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 01:04:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	-0.01
2	1,4-Dioxane	0.400	0.379	5.3	89	0.00
3	n-Nitrosodimethylamine	0.400	0.399	0.3	92	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	89	0.00
5 S	Phenol-d6	0.400	0.365	8.8	90	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	89	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.400	0.386	3.5	96	-0.01
9	Naphthalene	0.400	0.381	4.8	94	-0.01
10	Hexachlorobutadiene	0.400	0.389	2.8	95	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.355	11.3	92	-0.02
12	2-Methylnaphthalene	0.400	0.370	7.5	92	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.404	-1.0	97	-0.01
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	91	-0.01
16	Acenaphthylene	0.400	0.370	7.5	88	-0.01
17	Acenaphthene	0.400	0.382	4.5	91	-0.01
18	Fluorene	0.400	0.352	12.0	85	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	83	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.297	25.8#	69	-0.01
21	4-Bromophenyl-phenylether	0.400	0.383	4.3	82	-0.01
22	Hexachlorobenzene	0.400	0.424	-6.0	90	-0.01
23	Atrazine	0.400	0.361	9.8	82	-0.01
24	Pentachlorophenol	0.400	0.408	-2.0	99	-0.01
25	Phenanthrene	0.400	0.383	4.3	82	-0.01
26	Anthracene	0.400	0.369	7.8	80	-0.01
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	86	-0.01
28	Fluoranthene	0.400	0.402	-0.5	88	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	90	-0.02
30	Pyrene	0.400	0.388	3.0	88	-0.01
31 S	Terphenyl-d14	0.400	0.365	8.8	83	-0.01
32	Benzo(a)anthracene	0.400	0.370	7.5	86	-0.02
33	Chrysene	0.400	0.399	0.3	90	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.414	-3.5	102	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	93	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.408	-2.0	103	-0.03
37	Benzo(b)fluoranthene	0.400	0.402	-0.5	95	-0.01
38	Benzo(k)fluoranthene	0.400	0.392	2.0	96	-0.02
39 C	Benzo(a)pyrene	0.400	0.387	3.3	95	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	102	-0.03
41	Benzo(g,h,i)perylene	0.400	0.396	1.0	96	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4