

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091925\
 Data File : BN037847.D
 Acq On : 20 Sep 2025 11:39
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 22 00:38:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	-0.01
2	1,4-Dioxane	0.400	0.406	-1.5	123	0.00
3	n-Nitrosodimethylamine	0.400	0.381	4.8	113	0.00
4 S	2-Fluorophenol	0.400	0.381	4.8	121	0.00
5 S	Phenol-d6	0.400	0.373	6.8	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	118	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	122	-0.02
8 S	Nitrobenzene-d5	0.400	0.394	1.5	124	-0.02
9	Naphthalene	0.400	0.384	4.0	120	-0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	116	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.355	11.3	115	-0.02
12	2-Methylnaphthalene	0.400	0.369	7.8	116	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	117	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.472	-18.0	143	-0.01
15 S	2-Fluorobiphenyl	0.400	0.384	4.0	114	-0.01
16	Acenaphthylene	0.400	0.365	8.8	110	-0.01
17	Acenaphthene	0.400	0.377	5.8	112	-0.02
18	Fluorene	0.400	0.364	9.0	111	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.327	18.3	106	-0.01
21	4-Bromophenyl-phenylether	0.400	0.363	9.3	109	-0.02
22	Hexachlorobenzene	0.400	0.407	-1.7	121	-0.01
23	Atrazine	0.400	0.357	10.8	113	-0.01
24	Pentachlorophenol	0.400	0.423	-5.7	144	-0.01
25	Phenanthrene	0.400	0.384	4.0	115	-0.02
26	Anthracene	0.400	0.375	6.3	113	-0.01
27 SURR	Fluoranthene-d10	0.400	0.365	8.8	114	-0.01
28	Fluoranthene	0.400	0.383	4.3	117	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	128	-0.02
30	Pyrene	0.400	0.367	8.3	118	-0.01
31 S	Terphenyl-d14	0.400	0.348	13.0	113	-0.01
32	Benzo(a)anthracene	0.400	0.378	5.5	125	0.00
33	Chrysene	0.400	0.381	4.8	122	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.402	-0.5	142	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	132	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.413	-3.2	149	-0.03
37	Benzo(b)fluoranthene	0.400	0.397	0.8	133	-0.02
38	Benzo(k)fluoranthene	0.400	0.401	-0.3	140	-0.02
39 C	Benzo(a)pyrene	0.400	0.384	4.0	134	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.392	2.0	144	-0.04
41	Benzo(g,h,i)perylene	0.400	0.390	2.5	134	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4