

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3095

Analytical Method: SW8270-Modified

Client: AECOM Technical Services, Inc.

DataFile: BN037743.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3095-06MS		Client Sample ID: MW203D2-20250909MS									
1,4-Dioxane	0.4	0.85	0.97	ug/L	30	*			70	130	