

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3095

Client ID: SSTDCCC0.4

Date Analyzed: 09/20/2025

Lab File ID: BN037832.D

Time Analyzed: 02:35

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	11907	17.223	9777	21.402	10066	23.736
UPPER LIMIT	23814	17.723	19554	21.902	20132	24.236
LOWER LIMIT	5953.5	16.723	4888.5	20.902	5033	23.236
EPA SAMPLE NO.						
01 RE104D3-20250910	10379	17.22	8825	21.40	32 *	23.74

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.