

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037794.D
Instrument ID: BNA_N

Contract: AECO15
SDG NO.: Q3095
DFTPP Injection Date: 09/19/2025
DFTPP Injection Time: 02:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.4) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	86.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.9 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037795.D	09/19/2025	02:47
PB169692BL	PB169692BL	BN037796.D	09/19/2025	03:24
RE122D2-20250910	Q3095-29	BN037797.D	09/19/2025	04:00
RE104D1-20250910	Q3095-30	BN037798.D	09/19/2025	04:37
TT180D-20250911	Q3095-31	BN037799.D	09/19/2025	05:13
RE122D3-20250911	Q3095-32	BN037800.D	09/19/2025	05:49
SSTDCCC0.4EC	SSTDCCC0.4	BN037811.D	09/19/2025	13:03