

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3095</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/18/2025 11:30</u>
Lab File ID:	<u>VX047631.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.628		21.24	20
Chloromethane	0.680	0.746	0.1	9.71	20
Vinyl Chloride	0.666	0.733		10.06	20
Bromomethane	0.422	0.455		7.82	20
Chloroethane	0.445	0.461		3.6	20
Trichlorofluoromethane	0.989	1.010		2.12	20
1,1,2-Trichlorotrifluoroethane	0.578	0.587		1.56	20
1,1-Dichloroethene	0.594	0.605		1.85	20
Acetone	0.346	0.404		16.76	20
Carbon Disulfide	1.626	1.789		10.02	20
Methyl tert-butyl Ether	2.169	2.130		-1.8	20
Methyl Acetate	0.770	0.761		-1.17	20
Methylene Chloride	0.732	0.699		-4.51	20
trans-1,2-Dichloroethene	0.640	0.655		2.34	20
1,1-Dichloroethane	1.263	1.244	0.1	-1.5	20
Cyclohexane	1.052	1.079		2.57	20
2-Butanone	0.432	0.471		9.03	20
Carbon Tetrachloride	0.532	0.514		-3.38	20
cis-1,2-Dichloroethene	0.783	0.777		-0.77	20
Bromochloromethane	0.551	0.563		2.18	20
Chloroform	1.276	1.249		-2.12	20
1,1,1-Trichloroethane	1.082	1.070		-1.11	20
Methylcyclohexane	0.556	0.565		1.62	20
Benzene	1.488	1.458		-2.02	20
1,2-Dichloroethane	0.566	0.532		-6.01	20
Trichloroethene	0.360	0.361		0.28	20
1,2-Dichloropropane	0.381	0.370		-2.89	20
Bromodichloromethane	0.572	0.548		-4.2	20
4-Methyl-2-Pentanone	0.477	0.493		3.35	20
Toluene	0.917	0.902		-1.64	20
t-1,3-Dichloropropene	0.584	0.574		-1.71	20
cis-1,3-Dichloropropene	0.624	0.620		-0.64	20
1,1,2-Trichloroethane	0.354	0.342		-3.39	20
2-Hexanone	0.343	0.365		6.41	20
Dibromochloromethane	0.407	0.403		-0.98	20
1,2-Dibromoethane	0.360	0.364		1.11	20
Tetrachloroethene	0.326	0.324		-0.61	20
Chlorobenzene	1.136	1.105	0.3	-2.73	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3095</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/18/2025 11:30</u>
Lab File ID:	<u>VX047631.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.938		-1.12	20
m/p-Xylenes	0.729	0.727		-0.27	20
o-Xylene	0.706	0.706		0	20
Styrene	1.236	1.217		-1.54	20
Bromoform	0.293	0.288	0.1	-1.71	20
Isopropylbenzene	3.801	3.740		-1.61	20
1,1,2,2-Tetrachloroethane	1.150	1.146	0.3	-0.35	20
1,3-Dichlorobenzene	1.739	1.679		-3.45	20
1,4-Dichlorobenzene	1.785	1.687		-5.49	20
1,2-Dichlorobenzene	1.657	1.620		-2.23	20
1,2-Dibromo-3-Chloropropane	0.233	0.239		2.58	20
1,2,4-Trichlorobenzene	1.077	1.063		-1.3	20
1,2,3-Trichlorobenzene	1.002	0.981		-2.1	20
1,2-Dichloroethane-d4	0.796	0.745		-6.41	20
Dibromofluoromethane	0.335	0.325		-2.98	20
Toluene-d8	1.160	1.135		-2.15	20
4-Bromofluorobenzene	0.440	0.422		-4.09	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.