

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3095</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/19/2025 09:22</u>
Lab File ID:	<u>VX047660.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.650		25.48	20
Chloromethane	0.680	0.796	0.1	17.06	20
Vinyl Chloride	0.666	0.765		14.86	20
Bromomethane	0.422	0.509		20.62	20
Chloroethane	0.445	0.487		9.44	20
Trichlorofluoromethane	0.989	1.055		6.67	20
1,1,2-Trichlorotrifluoroethane	0.578	0.606		4.84	20
1,1-Dichloroethene	0.594	0.636		7.07	20
Acetone	0.346	0.370		6.94	20
Carbon Disulfide	1.626	1.859		14.33	20
Methyl tert-butyl Ether	2.169	2.275		4.89	20
Methyl Acetate	0.770	0.823		6.88	20
Methylene Chloride	0.732	0.763		4.24	20
trans-1,2-Dichloroethene	0.640	0.690		7.81	20
1,1-Dichloroethane	1.263	1.331	0.1	5.38	20
Cyclohexane	1.052	1.072		1.9	20
2-Butanone	0.432	0.467		8.1	20
Carbon Tetrachloride	0.532	0.515		-3.19	20
cis-1,2-Dichloroethene	0.783	0.828		5.75	20
Bromochloromethane	0.551	0.612		11.07	20
Chloroform	1.276	1.333		4.47	20
1,1,1-Trichloroethane	1.082	1.113		2.87	20
Methylcyclohexane	0.556	0.565		1.62	20
Benzene	1.488	1.532		2.96	20
1,2-Dichloroethane	0.566	0.567		0.18	20
Trichloroethene	0.360	0.371		3.06	20
1,2-Dichloropropane	0.381	0.386		1.31	20
Bromodichloromethane	0.572	0.580		1.4	20
4-Methyl-2-Pentanone	0.477	0.504		5.66	20
Toluene	0.917	0.930		1.42	20
t-1,3-Dichloropropene	0.584	0.600		2.74	20
cis-1,3-Dichloropropene	0.624	0.646		3.53	20
1,1,2-Trichloroethane	0.354	0.359		1.41	20
2-Hexanone	0.343	0.360		4.96	20
Dibromochloromethane	0.407	0.423		3.93	20
1,2-Dibromoethane	0.360	0.383		6.39	20
Tetrachloroethene	0.326	0.319		-2.15	20
Chlorobenzene	1.136	1.134	0.3	-0.18	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3095</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/19/2025</u> <u>09:22</u>
Lab File ID:	<u>VX047660.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.969		0.46	20
m/p-Xylenes	0.729	0.735		0.82	20
o-Xylene	0.706	0.714		1.13	20
Styrene	1.236	1.272		2.91	20
Bromoform	0.293	0.298	0.1	1.71	20
Isopropylbenzene	3.801	3.802		0.03	20
1,1,2,2-Tetrachloroethane	1.150	1.194	0.3	3.83	20
1,3-Dichlorobenzene	1.739	1.744		0.29	20
1,4-Dichlorobenzene	1.785	1.754		-1.74	20
1,2-Dichlorobenzene	1.657	1.660		0.18	20
1,2-Dibromo-3-Chloropropane	0.233	0.242		3.86	20
1,2,4-Trichlorobenzene	1.077	1.091		1.3	20
1,2,3-Trichlorobenzene	1.002	1.028		2.6	20
1,2-Dichloroethane-d4	0.796	0.779		-2.14	20
Dibromofluoromethane	0.335	0.329		-1.79	20
Toluene-d8	1.160	1.122		-3.28	20
4-Bromofluorobenzene	0.440	0.425		-3.41	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.