

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047587.D
 Acq On : 16 Sep 2025 10:37
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Sep 17 06:12:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:04:35 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 09/17/2025
 Supervised By : Mahesh Dadoda 09/18/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	189559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	336287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	298629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	49073	16.264	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	32.520%#		
35) Dibromofluoromethane	5.379	113	35788	15.868	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	31.740%#		
50) Toluene-d8	8.634	98	126875	16.258	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	32.520%#		
62) 4-Bromofluorobenzene	11.067	95	48454	16.360	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	32.720%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	44126	22.480	ug/l	100
3) Chloromethane	1.300	50	56023	21.716	ug/l	96
4) Vinyl Chloride	1.386	62	54256	21.485	ug/l	96
5) Bromomethane	1.617	94	34193	21.353	ug/l	100
6) Chloroethane	1.697	64	35551	21.063	ug/l	97
7) Trichlorofluoromethane	1.898	101	78667	20.980	ug/l	100
8) Diethyl Ether	2.136	74	31335	20.429	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.337	101	45733	20.859	ug/l	98
10) Methyl Iodide	2.459	142	69935	20.452	ug/l	99
11) Tert butyl alcohol	2.946	59	32456	96.582	ug/l	98
12) 1,1-Dichloroethene	2.325	96	46863	20.808	ug/l	96
13) Acrolein	2.239	56	26159	84.473	ug/l	100
14) Allyl chloride	2.666	41	94710	20.379	ug/l	97
15) Acrylonitrile	3.062	53	130143	104.554	ug/l	99
16) Acetone	2.373	43	151538	115.507	ug/l	100
17) Carbon Disulfide	2.520	76	133296	21.620	ug/l	98
18) Methyl Acetate	2.703	43	54120	18.536	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	168426	20.486	ug/l	98
20) Methylene Chloride	2.788	84	57153	20.581	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	50413	20.791	ug/l	94
22) Diisopropyl ether	3.757	45	188598	20.940	ug/l	96
23) Vinyl Acetate	3.715	43	765675	106.218	ug/l	100
24) 1,1-Dichloroethane	3.605	63	98509	20.578	ug/l	99
25) 2-Butanone	4.538	43	179487	109.679	ug/l	94
26) 2,2-Dichloropropane	4.471	77	81413	20.459	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	60865	20.497	ug/l	98
28) Bromochloromethane	4.891	49	34085	16.309	ug/l	100
29) Tetrahydrofuran	4.995	42	102935	104.560	ug/l	97
30) Chloroform	5.080	83	103586	21.413	ug/l	99
31) Cyclohexane	5.464	56	82770	20.759	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	86500	21.081	ug/l	99
36) 1,1-Dichloropropene	5.684	75	68395	20.754	ug/l	99
37) Ethyl Acetate	4.702	43	74499	21.113	ug/l	96
38) Carbon Tetrachloride	5.665	117	74636	20.844	ug/l	95
39) Methylcyclohexane	7.366	83	79499	21.251	ug/l	98
40) Benzene	6.025	78	210711	21.051	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047587.D
 Acq On : 16 Sep 2025 10:37
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/17/2025
 Supervised By : Mahesh Dadoda 09/18/2025

Quant Time: Sep 17 06:12:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:04:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	38525	21.148	ug/l	99
42) 1,2-Dichloroethane	6.074	62	80646	21.175	ug/l	99
43) Isopropyl Acetate	6.324	43	119623	20.670	ug/l	100
44) Trichloroethene	7.110	130	51369	21.228	ug/l	97
45) 1,2-Dichloropropane	7.415	63	53046	20.697	ug/l	97
46) Dibromomethane	7.568	93	38910	20.862	ug/l	99
47) Bromodichloromethane	7.805	83	80831	21.002	ug/l	98
48) Methyl methacrylate	7.677	41	60182	20.877	ug/l	98
49) 1,4-Dioxane	7.647	88	12047	413.530	ug/l	95
51) 4-Methyl-2-Pentanone	8.561	43	338370	105.456	ug/l	100
52) Toluene	8.708	92	129193	20.950	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	80972	20.625	ug/l	100
54) cis-1,3-Dichloropropene	8.354	75	86648	20.650	ug/l	99
55) 1,1,2-Trichloroethane	9.140	97	49857	20.967	ug/l	98
56) Ethyl methacrylate	9.104	69	79011	21.158	ug/l	99
57) 1,3-Dichloropropane	9.293	76	85511	20.934	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.226	63	141395	86.617	ug/l	100
59) 2-Hexanone	9.415	43	250007	108.488	ug/l	99
60) Dibromochloromethane	9.506	129	57994	21.167	ug/l	99
61) 1,2-Dibromoethane	9.598	107	51614	21.310	ug/l	99
64) Tetrachloroethene	9.262	164	40840	20.979	ug/l	95
65) Chlorobenzene	10.061	112	140642	20.731	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.146	131	48022	20.513	ug/l	99
67) Ethyl Benzene	10.177	91	244737	20.907	ug/l	98
68) m/p-Xylenes	10.287	106	182778	41.951	ug/l	100
69) o-Xylene	10.628	106	87651	20.799	ug/l	98
70) Styrene	10.640	104	154077	20.866	ug/l	99
71) Bromoform	10.780	173	35922	20.550	ug/l #	98
73) Isopropylbenzene	10.945	105	225730	20.970	ug/l	99
74) N-amyl acetate	10.829	43	104823	21.354	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.195	83	68132	20.920	ug/l	99
76) 1,2,3-Trichloropropane	11.225	75	60797m	20.680	ug/l	
77) Bromobenzene	11.183	156	56161	21.183	ug/l	99
78) n-propylbenzene	11.286	91	268399	21.092	ug/l	100
79) 2-Chlorotoluene	11.347	91	166237	21.329	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	187513	21.202	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.000	75	21553	19.173	ug/l	96
82) 4-Chlorotoluene	11.439	91	193122	20.980	ug/l	100
83) tert-Butylbenzene	11.695	119	189484	20.929	ug/l	100
84) 1,2,4-Trimethylbenzene	11.738	105	191477	21.355	ug/l	98
85) sec-Butylbenzene	11.872	105	223434	20.777	ug/l	100
86) p-Isopropyltoluene	11.994	119	190931	20.972	ug/l	98
87) 1,3-Dichlorobenzene	11.951	146	103584	21.035	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	105819	20.937	ug/l	99
89) n-Butylbenzene	12.317	91	176373	20.937	ug/l	99
90) Hexachloroethane	12.518	117	32526	20.349	ug/l	98
91) 1,2-Dichlorobenzene	12.317	146	99493	21.207	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.926	75	13729	20.821	ug/l	96
93) 1,2,4-Trichlorobenzene	13.567	180	62722	20.571	ug/l	96
94) Hexachlorobutadiene	13.707	225	20456	19.751	ug/l	98
95) Naphthalene	13.756	128	193134	20.801	ug/l	99
96) 1,2,3-Trichlorobenzene	13.938	180	58409	20.593	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047587.D
Acq On : 16 Sep 2025 10:37
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025

Quant Time: Sep 17 06:12:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:04:35 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047587.D
 Acq On : 16 Sep 2025 10:37
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Sep 17 06:12:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:04:35 2025
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By : John Carlone 09/17/2025
 Supervised By : Mahesh Dadoda 09/18/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.537 min Scan# 71
Delta R.T. -0.007 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 18955
Ion Ratio Lower Upper
168 100
99 61.9 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#2
Dichlorodifluoromethane
Concen: 22.480 ug/l
RT: 1.179 min Scan# 15
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion: 85 Resp: 44126
Ion Ratio Lower Upper
85 100
87 31.5 15.8 47.3









#3
Chloromethane
Concen: 21.716 ug/l
RT: 1.300 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 5602
Ion Ratio Lower Upper
50 100
52 34.9 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#4
Vinyl Chloride
Concen: 21.485 ug/l
RT: 1.386 min Scan# 49
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 62 Resp: 54256
Ion Ratio Lower Upper
62 100
64 33.4 24.9 37.3





#5
Bromomethane
Concen: 21.353 ug/l
RT: 1.617 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 3419
Ion Ratio Lower Upper
94 100
96 95.3 76.6 115.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#6
Chloroethane
Concen: 21.063 ug/l
RT: 1.697 min Scan# 100
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 64 Resp: 35551
Ion Ratio Lower Upper
64 100
66 31.0 26.1 39.1





#7
 Trichlorofluoromethane
 Concen: 20.980 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020



Tgt Ion:101 Resp: 7866
 Ion Ratio Lower Upper
 101 100
 103 64.4 51.4 77.2




Manual Integrations
 APPROVED


Reviewed By :John Carlone 09/17/2025
 Supervised By :Mahesh Dadoda 09/18/2025



#8
 Diethyl Ether
 Concen: 20.429 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37

Tgt Ion: 74 Resp: 31335
 Ion Ratio Lower Upper
 74 100
 45 107.0 52.3 156.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.859 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:101 Resp: 4573
Ion Ratio Lower Upper
101 100
85 43.6 35.5 53.3
151 72.1 59.4 89.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#10
Methyl Iodide
Concen: 20.452 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion:142 Resp: 69935
Ion Ratio Lower Upper
142 100
127 46.9 37.2 55.8
141 14.3 11.4 17.0









#11
Tert butyl alcohol
Concen: 96.582 ug/l
RT: 2.946 min Scan# 304
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 59 Resp: 32450
Ion Ratio Lower Upper
59 100
57 9.5 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#12
1,1-Dichloroethene
Concen: 20.808 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion: 96 Resp: 46863
Ion Ratio Lower Upper
96 100
61 170.7 131.0 196.6
98 60.8 47.7 71.5





#13

Acrolein

Concen: 84.473 ug/l

RT: 2.239 min Scan# 189

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 2615

Ion Ratio Lower Upper

56 100





55 71.5 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#14

Allyl chloride

Concen: 20.379 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion: 41 Resp: 94710


Ion Ratio Lower Upper

41 100

39 65.5 54.7 82.1


76 30.3 25.7 38.5





#15
Acrylonitrile
Concen: 104.554 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion: 53 Resp: 13014
Ion Ratio Lower Upper
53 100
52 82.3 66.4 99.6
51 36.2 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#16
Acetone
Concen: 115.507 ug/l
RT: 2.373 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Tgt Ion: 43 Resp: 151538
Ion Ratio Lower Upper
43 100
58 29.3 23.4 35.0





#17
Carbon Disulfide
Concen: 21.620 ug/l
RT: 2.520 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 76 Resp: 133290
Ion Ratio Lower Upper
76 100
78 9.1 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#18
Methyl Acetate
Concen: 18.536 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 43 Resp: 54120
Ion Ratio Lower Upper
43 100
74 22.7 17.4 26.2





#19
Methyl tert-butyl Ether
Concen: 20.486 ug/l
RT: 3.111 min Scan# 332
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 73 Resp: 168420
Ion Ratio Lower Upper
73 100
57 23.0 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#20
Methylene Chloride
Concen: 20.581 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 84 Resp: 57153
Ion Ratio Lower Upper
84 100
49 131.7 108.6 162.8
51 37.9 32.6 48.8
86 64.7 52.0 78.0





#21
trans-1,2-Dichloroethene
Concen: 20.791 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 96 Resp: 5041
Ion Ratio Lower Upper
96 100
61 151.3 116.2 174.2
98 67.2 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#22
Diisopropyl ether
Concen: 20.940 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 45 Resp: 188598
Ion Ratio Lower Upper
45 100
43 106.1 81.0 121.4
87 25.5 20.3 30.5
59 11.2 9.0 13.4





#23
 Vinyl Acetate
 Concen: 106.218 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020




Tgt Ion: 43 Resp: 76567
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.8 11.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 09/17/2025
 Supervised By :Mahesh Dadoda 09/18/2025




#24
 1,1-Dichloroethane
 Concen: 20.578 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37




Tgt Ion: 63 Resp: 98509
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.8 11.2
 100 4.1 2.3 6.9





#25
2-Butanone
Concen: 109.679 ug/l
RT: 4.538 min Scan# 566
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion: 43 Resp: 17948
Ion Ratio Lower Upper
43 100
72 22.3 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#26
2,2-Dichloropropane
Concen: 20.459 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion: 77 Resp: 81413
Ion Ratio Lower Upper
77 100
97 21.2 10.9 32.7









#27
 cis-1,2-Dichloroethene
 Concen: 20.497 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Tgt Ion: 96 Resp: 6086
 Ion Ratio Lower Upper
 96 100
 61 154.8 0.0 303.2
 98 63.7 0.0 127.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 09/17/2025
 Supervised By :Mahesh Dadoda 09/18/2025



#28
 Bromochloromethane
 Concen: 16.309 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37


Tgt Ion: 49 Resp: 34085
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.2
 130 72.7 57.8 86.8





#29
Tetrahydrofuran
Concen: 104.560 ug/l
RT: 4.995 min Scan# 641
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 42 Resp: 102935
Ion Ratio Lower Upper
42 100
72 41.5 34.2 51.2
71 37.8 31.9 47.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#30
Chloroform
Concen: 21.413 ug/l
RT: 5.080 min Scan# 655
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 83 Resp: 103586
Ion Ratio Lower Upper
83 100
85 64.2 51.8 77.6





#31
Cyclohexane
Concen: 20.759 ug/l
RT: 5.464 min Scan# 717
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 56 Resp: 82770
Ion Ratio Lower Upper
56 100
69 27.9 23.8 35.8
84 82.2 64.6 97.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#32
1,1,1-Trichloroethane
Concen: 21.081 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 97 Resp: 86500
Ion Ratio Lower Upper
97 100
99 63.5 50.6 76.0
61 45.5 35.8 53.8





#33
 1,2-Dichloroethane-d4
 Concen: 16.264 ug/l
 RT: 5.940 min Scan# 796
 Delta R.T. -0.000 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020




Tgt Ion: 65 Resp: 4907
 Ion Ratio Lower Upper
 65 100
 67 52.2 0.0 103.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 09/17/2025
 Supervised By :Mahesh Dadoda 09/18/2025




#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.751 min Scan# 929
 Delta R.T. 0.006 min
 Lab File: VX047587.D
 Acq: 16 Sep 2025 10:37




Tgt Ion:114 Resp: 336287
 Ion Ratio Lower Upper
 114 100
 63 22.3 0.0 44.0
 88 16.8 0.0 34.2





#35
Dibromofluoromethane
Concen: 15.868 ug/l
RT: 5.379 min Scan# 703
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion: 113 Resp: 35788
Ion Ratio Lower Upper
113 100
111 105.7 80.9 121.3
192 18.8 14.2 21.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#36
1,1-Dichloropropene
Concen: 20.754 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion: 75 Resp: 68395
Ion Ratio Lower Upper
75 100
110 33.9 16.7 50.1
77 31.2 24.3 36.5





#37

Ethyl Acetate

Concen: 21.113 ug/l

RT: 4.702 min Scan# 593

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	43	Resp:	7449
Ion Ratio	Lower	Upper	
43	100		
61	10.8	9.8	14.8
70	8.5	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#38

Carbon Tetrachloride

Concen: 20.844 ug/l




RT: 5.665 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Tgt Ion:	117	Resp:	74636
Ion Ratio	Lower	Upper	
117	100		
119	96.9	73.4	110.2
121	31.6	23.8	35.6





#39

Methylcyclohexane

Concen: 21.251 ug/l

RT: 7.366 min Scan# 1030

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 83	Resp: 7949
Ion Ratio	Lower Upper
83	100
55	81.2 66.4 99.6
98	49.6 38.1 57.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#40

Benzene

Concen: 21.051 ug/l




RT: 6.025 min Scan# 810


Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Tgt Ion: 78	Resp: 210711
Ion Ratio	Lower Upper
78	100
77	24.4 18.8 28.2





#41
Methacrylonitrile
Concen: 21.148 ug/l
RT: 4.910 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 41 Resp: 3852
Ion Ratio Lower Upper
41 100
39 53.8 43.3 64.9
67 72.5 58.6 87.8
52 32.3 26.0 39.0

Manual Integrations

APPROVED




Reviewed By :John Carlone 09/17/2025


Supervised By :Mahesh Dadoda 09/18/2025



#42
1,2-Dichloroethane
Concen: 21.175 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion: 62 Resp: 80646
Ion Ratio Lower Upper
62 100
98 8.7 0.0 17.0









#43
Isopropyl Acetate
Concen: 20.670 ug/l
RT: 6.324 min Scan# 859
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 11962
Ion Ratio Lower Upper
43 100
61 18.9 15.2 22.8
87 12.1 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#44
Trichloroethene
Concen: 21.228 ug/l
RT: 7.110 min Scan# 988
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion:130 Resp: 51369
Ion Ratio Lower Upper
130 100
95 102.3 0.0 210.8





#45

1,2-Dichloropropane

Concen: 20.697 ug/l

RT: 7.415 min Scan# 1038

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 5304

Ion Ratio Lower Upper

63 100





65 29.4 24.7 37.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#46

Dibromomethane

Concen: 20.862 ug/l

RT: 7.568 min Scan# 1063

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion: 93 Resp: 38910


Ion Ratio Lower Upper

93 100

95 82.0 65.0 97.6

174 89.2 71.6 107.4





#47
Bromodichloromethane
Concen: 21.002 ug/l
RT: 7.805 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 83 Resp: 8083
Ion Ratio Lower Upper
83 100
85 63.5 52.5 78.7
127 8.2 6.9 10.3

Manual Integrations




APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#48
Methyl methacrylate
Concen: 20.877 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 41 Resp: 60182
Ion Ratio Lower Upper
41 100
69 74.2 61.4 92.2
39 57.2 45.8 68.6





#49
1,4-Dioxane
Concen: 413.530 ug/l
RT: 7.647 min Scan# 1076
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 88 Resp: 1204
Ion Ratio Lower Upper
88 100
43 44.7 30.2 45.2
58 75.8 59.5 89.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#50
Toluene-d8
Concen: 16.258 ug/l
RT: 8.634 min Scan# 1238
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 98 Resp: 126875
Ion Ratio Lower Upper
98 100
100 67.9 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 105.456 ug/l
RT: 8.561 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion: 43 Resp: 338370
Ion Ratio Lower Upper
43 100
58 37.8 30.3 45.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#52
Toluene
Concen: 20.950 ug/l
RT: 8.708 min Scan# 1250
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion: 92 Resp: 129193
Ion Ratio Lower Upper
92 100
91 173.2 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 20.625 ug/l

RT: 8.964 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 75 Resp: 80972

Ion Ratio Lower Upper

75 100





77 31.9 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#54

cis-1,3-Dichloropropene

Concen: 20.650 ug/l

RT: 8.354 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion: 75 Resp: 86648


Ion Ratio Lower Upper

75 100

77 30.9 25.8 38.6


39 52.0 41.4 62.2





#55
1,1,2-Trichloroethane
Concen: 20.967 ug/l
RT: 9.140 min Scan# 1320
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion: 97 Resp: 4985
Ion Ratio Lower Upper
97 100
83 83.3 69.6 104.4
85 55.2 45.0 67.6
99 62.5 49.6 74.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#56
Ethyl methacrylate
Concen: 21.158 ug/l
RT: 9.104 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion: 69 Resp: 79011
Ion Ratio Lower Upper
69 100
41 72.6 58.6 87.8
39 38.6 31.5 47.3









#57
1,3-Dichloropropane
Concen: 20.934 ug/l
RT: 9.293 min Scan# 1134
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 76 Resp: 8551
Ion Ratio Lower Upper
76 100
78 32.6 26.2 39.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#58
2-Chloroethyl Vinyl ether
Concen: 86.617 ug/l
RT: 8.226 min Scan# 1171
Delta R.T. 0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 63 Resp: 141395
Ion Ratio Lower Upper
63 100
106 27.5 21.9 32.9





#59
2-Hexanone
Concen: 108.488 ug/l
RT: 9.415 min Scan# 1366
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 43 Resp: 25000
Ion Ratio Lower Upper
43 100
58 51.4 26.1 78.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#60
Dibromochloromethane
Concen: 21.167 ug/l
RT: 9.506 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion:129 Resp: 57994
Ion Ratio Lower Upper
129 100
127 77.7 38.4 115.2





#61
1,2-Dibromoethane
Concen: 21.310 ug/l
RT: 9.598 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion:107 Resp: 51614
Ion Ratio Lower Upper
107 100
109 93.4 75.3 112.9

Manual Integrations




APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#62
4-Bromofluorobenzene
Concen: 16.360 ug/l
RT: 11.067 min Scan# 1637
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 95 Resp: 48454
Ion Ratio Lower Upper
95 100
174 70.2 0.0 141.6
176 70.3 0.0 138.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:117 Resp: 298629
Ion Ratio Lower Upper
117 100
82 58.6 47.8 71.6
119 31.3 25.2 37.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#64
Tetrachloroethene
Concen: 20.979 ug/l
RT: 9.262 min Scan# 1341
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion:164 Resp: 40840
Ion Ratio Lower Upper
164 100
166 125.4 100.2 150.2
129 91.8 81.8 122.6
131 93.7 78.6 118.0





#65

Chlorobenzene

Concen: 20.731 ug/l

RT: 10.061 min Scan# 1472

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:112 Resp: 14064

Ion Ratio Lower Upper

112 100





114 31.4 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 20.513 ug/l

RT: 10.146 min Scan# 1486

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion:131 Resp: 48022


Ion Ratio Lower Upper

131 100

133 95.9 47.3 142.0

119 63.9 32.0 96.0









#67
Ethyl Benzene
Concen: 20.907 ug/l
RT: 10.177 min Scan# 1491
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 91 Resp: 24473
Ion Ratio Lower Upper
91 100
106 29.3 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#68
m/p-Xylenes
Concen: 41.951 ug/l
RT: 10.287 min Scan# 1509
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion:106 Resp: 182778
Ion Ratio Lower Upper
106 100
91 210.2 167.8 251.8





#69

o-Xylene

Concen: 20.799 ug/l

RT: 10.628 min Scan# 1565

Delta R.T. 0.006 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion:106 Resp: 8765

Ion Ratio Lower Upper

106 100





91 225.3 111.1 333.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#70

Styrene

Concen: 20.866 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion:104 Resp: 154077


Ion Ratio Lower Upper

104 100

78 53.6 43.1 64.7

103 54.4 43.1 64.7





#71

Bromoform

Concen: 20.550 ug/l

RT: 10.780 min Scan# 1590

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	173	Resp:	3592
Ion Ratio	Lower	Upper	
173	100		
175	48.3	24.9	74.6
254	0.0	0.1	0.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.006 min Scan# 1791


Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37


Tgt Ion:	152	Resp:	141590
Ion Ratio	Lower	Upper	
152	100		
115	73.2	44.9	134.5
150	167.1	0.0	352.0





#73
Isopropylbenzene
Concen: 20.970 ug/l
RT: 10.945 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 105 Resp: 225730
Ion Ratio Lower Upper
105 100
120 25.9 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#74
N-ethyl acetate
Concen: 21.354 ug/l
RT: 10.829 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion: 43 Resp: 104823
Ion Ratio Lower Upper
43 100
70 39.8 32.0 48.0
55 23.4 18.9 28.3
61 23.6 18.8 28.2





#75
1,1,2,2-Tetrachloroethane
Concen: 20.920 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion: 83 Resp: 6813
Ion Ratio Lower Upper
83 100
131 10.0 5.0 14.9
85 64.4 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#76
1,2,3-Trichloropropane
Concen: 20.680 ug/l m
RT: 11.225 min Scan# 1663
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion: 75 Resp: 60797
Ion Ratio Lower Upper
75 100
77 39.6 18.8 56.4





#77

Bromobenzene

Concen: 21.183 ug/l

RT: 11.183 min Scan# 1656

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	156	Resp:	5616
Ion Ratio	Lower	Upper	
156	100		
77	177.4	89.7	269.1
158	95.2	47.9	143.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#78

n-propylbenzene

Concen: 21.092 ug/l




RT: 11.286 min Scan# 1673


Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Tgt Ion:	91	Resp:	268399
Ion Ratio	Lower	Upper	
91	100		
120	22.3	11.1	33.1









#79
2-Chlorotoluene
Concen: 21.329 ug/l
RT: 11.347 min Scan# 1683
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 16623
Ion Ratio Lower Upper
91 100
126 32.1 16.1 48.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#80
1,3,5-Trimethylbenzene
Concen: 21.202 ug/l
RT: 11.433 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37


Tgt Ion:105 Resp: 187513
Ion Ratio Lower Upper
105 100
120 47.4 23.6 70.8





#81
trans-1,4-Dichloro-2-butene
Concen: 19.173 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion: 75 Resp: 2155
Ion Ratio Lower Upper
75 100
53 109.1 83.4 125.2
89 47.6 36.3 54.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#82
4-Chlorotoluene
Concen: 20.980 ug/l
RT: 11.439 min Scan# 1698
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Tgt Ion: 91 Resp: 193122
Ion Ratio Lower Upper
91 100
126 28.5 14.2 42.6





#83
tert-Butylbenzene
Concen: 20.929 ug/l
RT: 11.695 min Scan# 1169
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion:119 Resp: 189484
Ion Ratio Lower Upper
119 100
91 59.2 29.6 88.8
134 21.8 11.0 33.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#84
1,2,4-Trimethylbenzene
Concen: 21.355 ug/l
RT: 11.738 min Scan# 1747
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37



Tgt Ion:105 Resp: 191477
Ion Ratio Lower Upper
105 100
120 42.6 22.1 66.1





#85

sec-Butylbenzene

Concen: 20.777 ug/l

RT: 11.872 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:105 Resp: 223434

Ion Ratio Lower Upper

105 100





134 19.7 9.8 29.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#86

p-Isopropyltoluene

Concen: 20.972 ug/l

RT: 11.994 min Scan# 1789

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37




Tgt Ion:119 Resp: 190931


Ion Ratio Lower Upper

119 100

134 26.1 12.6 37.8


91 26.2 12.6 37.6





#87
1,3-Dichlorobenzene
Concen: 21.035 ug/l
RT: 11.951 min Scan# 1782
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:146 Resp: 103584
Ion Ratio Lower Upper
146 100
111 42.4 21.5 64.5
148 62.7 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025




#88
1,4-Dichlorobenzene
Concen: 20.937 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Tgt Ion:146 Resp: 105819
Ion Ratio Lower Upper
146 100
111 42.6 20.8 62.4
148 64.4 31.7 95.1





#89
n-Butylbenzene
Concen: 20.937 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 91 Resp: 17637
Ion Ratio Lower Upper
91 100
92 54.6 27.6 82.7
134 25.3 12.6 37.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/17/2025
Supervised By :Mahesh Dadoda 09/18/2025



#90
Hexachloroethane
Concen: 20.349 ug/l
RT: 12.518 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion:117 Resp: 32526
Ion Ratio Lower Upper
117 100
201 66.6 34.2 102.5





#91
1,2-Dichlorobenzene
Concen: 21.207 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:146 Resp: 9949
Ion Ratio Lower Upper
146 100
111 43.4 22.2 66.6
148 63.9 31.8 95.4

Manual Integrations

APPROVED




Reviewed By :John Carlone 09/17/2025


Supervised By :Mahesh Dadoda 09/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.821 ug/l
RT: 12.926 min Scan# 1942
Delta R.T. 0.006 min
Lab File: VX047587.D
Acq: 16 Sep 2025 10:37

Tgt Ion: 75 Resp: 13729
Ion Ratio Lower Upper
75 100
155 78.8 37.7 113.1
157 95.9 49.8 149.4





#93

1,2,4-Trichlorobenzene

Concen: 20.571 ug/l

RT: 13.567 min Scan# 2047

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:180	Resp: 6272
Ion Ratio	Lower Upper
180 100	
182 97.8	46.6 139.8
145 33.4	16.4 49.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#94

Hexachlorobutadiene

Concen: 19.751 ug/l




RT: 13.707 min Scan# 2070


Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Tgt Ion:225	Resp: 20456
Ion Ratio	Lower Upper
225 100	
223 65.0	31.7 95.1
227 65.1	32.0 96.0





#95

Naphthalene

Concen: 20.801 ug/l

RT: 13.756 min Scan# 2078

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	128	Resp:	193134
Ion Ratio	Lower	Upper	
128	100		
127	12.3	10.2	15.2
129	10.5	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/17/2025

Supervised By :Mahesh Dadoda 09/18/2025



#96

1,2,3-Trichlorobenzene

Concen: 20.593 ug/l

RT: 13.938 min Scan# 2108

Delta R.T. -0.000 min

Lab File: VX047587.D

Acq: 16 Sep 2025 10:37

Tgt Ion:	180	Resp:	58409
Ion Ratio	Lower	Upper	
180	100		
182	95.5	47.6	142.8
145	35.5	17.9	53.7

