

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Contract: AEC015

Lab Code: ACE

SDG NO.: Q3095

Lab File ID: VX047601.D

Date Analyzed: 09/17/2025

Instrument ID: MSVOA X

Time Analyzed: 08:51

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	138326	12.006				
UPPER LIMIT	276652	12.506				
LOWER LIMIT	69163	11.506				
EPA SAMPLE NO.						
MW305I-20250908	165030	12.01				
MW179D-20250908	142056	12.01				
MW201D-20250908	152611	12.01				
MW179D1-20250908	143935	12.01				
MW203D2-20250909	141573	12.01				
MW203D2-20250909MS	124680	12.01				
MW203D2-20250909MSD	122820	12.01				
MW203D1-20250909	144060	12.01				
MW179D2-20250909	144921	12.01				
MW201D1-20250909	144973	12.01				
RE122D1-20250909	143619	12.01				
MW178I1-20250909	144330	12.01				
MW305D-20250909	122022	12.01				
DUP01-20250909	137877	12.01				
RE126D2-20250909	137540	12.01				
RE126D1-20250910	145228	12.01				
MW202D1-20250910	143137	12.01				
VX0917WBL01	224854	12.01				
VX0917WBS01	151595	12.01				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3095
 Lab File ID: VX047601.D Date Analyzed: 09/17/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:51
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	138326	12.00€				
UPPER LIMIT	276652	12.50€				
LOWER LIMIT	69163	11.50€				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.