

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3095</u>
Lab File ID:	<u>VX047631.D</u>	Date Analyzed:	<u>09/18/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>11:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	137093	12.006				
UPPER LIMIT	274186	12.506				
LOWER LIMIT	68546.5	11.506				
EPA SAMPLE NO.						
RE122D1-20250909DL	143980	12.01				
MW202D-20250909	144225	12.01				
MW178I1-20250909DL	151081	12.01				
RE123D1-20250909	151637	12.01				
RE123D3-20250909	153255	12.01				
MW305D-20250909DL	157275	12.01				
DUP01-20250909DL	145612	12.01				
RE126D2-20250909DL	150558	12.01				
RE126D3-20250909	143114	12.01				
MW178S-20250910	154032	12.01				
RE104D2-20250910	145628	12.01				
DUP02-20250910	122561	12.01				
RE104D3-20250910	144761	12.01				
RE122D2-20250910	122418	12.01				
TB01-20250911	142150	12.01				
TB02-20250911	139535	12.01				
VX0918WBL01	168174	12.01				
VX0918WBS01	143266	12.01				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3095
 Lab File ID: VX047631.D Date Analyzed: 09/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	137093	12.00€				
UPPER LIMIT	274186	12.50€				
LOWER LIMIT	68546.5	11.50€				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.