

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3095</u>
Lab File ID:	<u>VX047660.D</u>	Date Analyzed:	<u>09/19/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:22</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	129778	12.00€				
UPPER LIMIT	259556	12.50€				
LOWER LIMIT	64889	11.50€				
EPA SAMPLE NO.						
RE104D2-20250910DL	142184	12.01				
MW178I-20250910	151965	12.01				
DUP02-20250910DL	151017	12.01				
DUP03-20250910	140190	12.01				
RE117D1-20250910	136445	12.01				
RE122D2-20250910DL	137256	12.01				
RE104D1-20250910	140121	12.01				
TT180D-20250911	139806	12.01				
RE122D3-20250911	150121	12.01				
VX0919WBL01	175228	12.01				
VX0919WBS01	146041	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.